A-A-54408
12 September 1991

COMMERCIAL ITEM DESCRIPTION

BLOOD EXCHANGE-TRANSFUSION KIT, DISPOSABLE

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-B-36731 which is cancelled.

This commercial item description covers a ready to use disposable blood exchange-transfusion kit suitable for use in the treatment of RH blood incompatibility and for transfusion of blood without cross infection.

Salient Characteristics:

General. The blood exchange-transfusion kit shall be furnished in a plastic tray approximately 15-1/2 inches long, 9-1/8 inches wide and 1-1/18 inches in height. The blood exchange-transfusion kit shall contain the following components as a minimum:

Quantity	Component
1	Blood Transfusion Kit, With Nylon Blood Filter, Tubing and Control Clamp.
1	Calcium Gluconate Ampule, 10% USP, 10 ml.
1	Container for Waste Blood with Drawstring Closure, 2000 ml, and with 100 ml calibrations.
1	Exchange-Transfusion Record.
1	Extension Tube, 33 Inches long.
1	Stopcock Assembly (Stopcock with 40 inch Extension Tube).
3	Gauze Sponges, 2 by 2 Inch, 12 Ply.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Quantity	Component
1	Ruler, Plastic, 15 cm.
1	Polypropylene Syringe, 5 ml with 23 Gage by 1 inch needle.
2	Polypropylene Syringes, Luer-Lock, 20 ml.
1	Umbilical Catheter, 5 Fr, 14 Inches Long with 0.22 ml capacity.
1	Umbilical Catheter, 8 Fr, 14 Inches Long with 0.88 ml capacity.

Blood transfusion set. The blood transfusion set shall consist of a piercing device, filter, housing, nylon filter, 60 inches of plastic tubing, 2-1/2 inches of natural rubber tubing on proximal end (next to filter), 4 inches of rubber tubing on distal end (next to needle adapter), needle adapter and a flexible aluminum control clamp.

Pyrogens, safety and biological tests. All rubber and plastic components shall be sterilizable, free from pyrogens in accordance with the USP, safe in accordance with the USP (nontoxic) and shall contain not more than 5 parts per million (ppm) of lead, barium and cadmium. The plastic shall meet the requirements of the United States Pharmacopeia (U.S.P.) for Class II plastics.

Material. Plastic tubing shall be transparent, colorless polyvinyl. The piercing device and needle adapter shall be fabricated from suitable rigid plastic. Needles shall be fabricated of type 304 corrosion-resisting steel in accordance with ASTM Designation F899.

Sterility. Components and inside surfaces of the tray shall be sterile in accordance with the USP or AAMI guidelines. Sterility Assurance Level shall be 0.000001.

Expiration dating. The blood exchange-transfusion kit shall have an expiration date of at least 36 months. At time of shipment to the Government, at least 30 months of the expiration dating period shall remain on the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Instructions. Each kit shall be furnished with a list of the kit contents and procedural instructions for use.

Pressure tests. The kit shall not leak nor shall any damage occur when subjected to an internal hydrostatic pressure of 8 psig for 30 seconds.

Workmanship. The kit shall be free from defects which detract from its appearance or impair its serviceability. Components and tray shall be free from burrs, flash, cracks, corrosion or foreign material.

Unit. Kit (KT). One blood exchange-transfusion kit, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one blood exchange-transfusion kit. Sampling shall be inspection level II with an acceptable quality level (AQL) of 1.0.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one blood exchange-transfusion kit. Sampling shall be inspection level be S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers'own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-837-7491	BLOOD, EXCHANGE-TRANSFUSION KIT, Disposable

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-5003
Location: A-A-54408.doc\S-11
Disk #1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE. This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-54408	2 DOCUMENT DATE (YYMMDD) 12 September 1991

3 DOCUMENT TITLE
BLOOD EXCHANGE-TRANSFUSION KIT, DISPOSABLE

4 NATURE OF CHANGE (*Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed*)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle Initial)	b ORGANIZATION	
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME Defense Personnel Support Center ATTN: DPSC-MSE	b TELEPHONE (Include Area Code) (1) Commercial (215)737-2870	(2) AUTOVON 444-2870
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.