COMMERCIAL ITEM DESCRIPTION

PIN, BONE, STEINMANN, BAYONET AND TROCAR POINT

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-P-36627A which is cancelled.

This Commercial Item Description covers bone pins of the Steinmann design, bayonet and trocar point, for traction reduction of fractures.

Salient characteristics:

General. The bone pins shall be of the following types, sizes, and styles, as specified:

Type	Point Compliance
I	Bayonet
II	Trocar

Size	Length (inches)
1	4
2	5
3	6-1/2
4	8
5	9
6	12

Style	Diameter (inches)
1	0.093
2	0.125
3	0.156

Material. The bone pins shall be fabricated from cold drawn wire, conforming to the composition specified in ASTM F138, Grade 2 (0.03 maximum carbon content). Tensile strength of wire shall be 150,00 to 200,00 pounds per square inch (psi) when tested in accordance with ASTM E8. Grain size of the raw material shall be 5 or finer in accordance with ASTM F138. Microcleanliness of the raw material shall be in accordance with ASTM F138.

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Style, design and dimension. The bone pins shall be a Type I (bayonet point) or Type II (trocar point) and have a triangular shank end. The bone pins shall be of the Steinmann design, suitable for implanting into the body for traction reduction of fractures. The bone pins shall be in accordance with the Defense Personnel Support Center (DPSC) Drawing No. 2202, for Type I bone pins and DPSC Drawing No. 1907 for Type II pins. Bone pins shall be furnished in accordance with the appropriate type, size, style, stock number and item identification.

Hardness. The bone pins shall have a Rockwell Hardness of C33 to C43 in accordance with ASTM E18 or E92 whichever is applicable.

Finish. The bone pins shall have a smooth, mirror finish, that is properly electropolished. Definitions of acceptable surface or characteristics shall be as described in ASTM F86.

Passivation. The bone pins shall be properly passivated after completion of all final finishing operations and shall be accomplished by the methods described in ASTM F86. Under no circumstances shall electropolishing be considered to satisfy this passivation requirement.

Corrosion. The bone pins shall not show evidence of corrosion when tested in accordance with ASTM E1089. As an alternate, the following saline test may be used: The bone pins shall be scrubbed w/soap and warm water, rinsed in hot water, followed by immersion into ethyl alcohol (95%) and dried. Bone pins shall be completely immersed into Nomad Saline Solution (USP) and while immersed, placed into an oven at a temperature of 37 $\pm$ 1 degrees centigrade for a period of 7 days, removed from the oven and examined for corrosion while still immersed.

Workmanship. The bone pins shall be free from pits, cracks, slivers, scale, corrosion, tool marks, dull points and other defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One bone pin of the type, size, and style, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory applicants or procedures) and testing (determination by technical means and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one bone pin. Sampling shall be Inspection level II with an acceptable quality level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one bone pin. Sampling shall be Inspection level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

NOTE: The following National Stock Numbers are covered by this document:

NSN	Item Identification	Specification Designation		
		Type	Size	Style
6515-00-299-8315	PIN, BONE, Bayonet Point, Steinmann, 4 inch.	I	1	1
6515-00-299-8316	PIN, BONE, Bayonet Point, Steinmann, 5 inch.	I	2	2
6515-00-299-8317	PIN, BONE, Bayonet Point, Steinmann, 6-1/2 inch.	I	3	2
6515-00-299-8318	PIN, BONE, Bayonet Point, Steinmann, 8 inch.	I	4	3
6515-00-299-8319	PIN, BONE, Bayonet Point, Steinmann, 9 inch.	I	5	3
6515-00-310-7880	PIN, BONE, Trocar Point, Steinmann, 4 inch.	II	1	1
6515-00-310-7900	PIN, BONE, Trocar Point, Steinmann, 5 inch.	II	2	2
6515-00-310-7920	PIN, BONE, Trocar Point, Steinmann, 6-1/2 inch.	II	3	2
6515-00-310-7940	PIN, BONE, Trocar Point, Steinmann, 8 inch.	II	4	3
6515-00-310-7960	PIN, BONE, Trocar Point, Steinmann, 9 inch.	II	5	3

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

NSN	Item Identification	Specification Designation		
		Type	Size	Style
6515-00-664-7854	PIN, BONE, Trocar Point, Steinmann, 9 inch.	II	5	1
6515-00-664-7855	PIN, BONE, Trocar Point, Steinmann, 4 inch.	II	6	3

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-4980

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.

2. The submitter of this form must complete blocks 4, 5, 6, and 7.

3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54416	2. DOCUMENT DATE (YYMMDD) 14 June 1991

3. DOCUMENT TITLE
PIN, BONE, STEINMANN, BAYONET AND TROCAR POINT

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON
c. ADDRESS (Include Zip Code)	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 98/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.