A-A-54429
29 July 1988

COMMERCIAL ITEM DESCRIPTION

BAG, PLASTIC

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a heat or tape sealable plastic bag suitable for use as a dust cover for sterilized medical supplies.

Salient Characteristics:

The plastic bag shall be a clear polyethylene cover suitable for protecting sterilized medical supplies from dust contamination when sealed by heat or tape. The plastic bag shall be supplied in the following sizes:

Size I - 12 inches wide by 20 inches long (30.5 cm by 50.8 cm)

Size II - 16 inches wide by 30 inches long (40.6 cm by 76.2 cm)

Width and length specified for indicated bag size shall be minimum dimensions. Dimensional tolerances within each lot of material shall be plus or minus 1/4 inch for width and length. The polyethylene film used in the construction of the bag shall be 3 mils plus or minus 1/2 mil thick.

The plastic bag shall be sealed, or closed, on three ends. One sealed, or closed, end shall have a tear line suitable for rapid opening without use of a scissors, knife, or other instrument. A perforated tear line shall not be acceptable as this would compromise sterility. The open end of the plastic bag shall be readily sealable by use of heat or tape.

AMSC N/A FSC 8105
DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

The plastic bag shall be distinctively imprinted on one side with the following, or equivalent, labeling:

a. Grasp either corner and snap smartly to open.
b. Dust cover to be used after sterilization.
c. Heat seal or tape.

Printing shall be legible and complete.

The plastic bag shall be free of holes, tears, or perforations. It shall be free of dirt, grease, or loose foreign matter.

Workmanship. The plastic bag shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing two hundred fifty bags, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one plastic bag as specified in this description.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one plastic bag.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Defects.

Bag has no tear line.

Bag seams are not completely sealed or closed.

Bag not free of holes, tears, or perforations.

Bag printing missing, illegible, or incomplete.

Bag seams not free of wrinkles or bunching.

Bag not free of dirt, grease, or loose foreign matter.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

8105-01-265-3379 (Size I)
8105-00-104-9893 (Size II)

MILITARY INTERESTS

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-M8

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA - OSS
USPHS
FDA-MPQAS

PCC 70069
05966

Project Number: 8105-0322

Location: BAGMC13.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.