A-A-54432
25 June 1991

COMMERCIAL ITEM DESCRIPTION

BOTTLE SET, PHYSICIAN'S SATCHEL

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for MIL-B-37523 (DSA-DM) which is cancelled.

This Commercial Item Description covers a bottle set suitable for use in a physician's satchel.

Salient Characteristics:

The physician's satchel bottle set shall consist of a bottle case containing 12 one-half ounce and 8 one ounce screw capped bottles. The bottle set design and the dimensions of the folded case are shown in figure 1. The bottles shall be made of commercial grade clear glass of soda-lime composition. The bottles shall be cylindrical, of the screwcap type, and shall conform to the following dimensions in inches with tolerances of plus or minus 15 percent:

Capacity	Length of bottle w/cap	Outside diameter of bottle	Inside diameter of mouth	Outside diameter of cap
1/2 ounce	3-5/16	5/8	7/16	43/64
1 ounce	3-3/4	15/16	5/8	63/64

The bottles shall be free of cracks, chips, sharp edges, blisters, or other defects.

The bottle screwcaps shall be made of a hard, durable plastic, resistant to impact. The screwcaps shall be lined inside with cork, pulp, or other resilient composition material. The liner shall be faced with a vinyl or vinylidene type coating

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.




FIGURE 1. Bottle set, physician's satchel.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

impervious to moisture.

The bottle case shall be fabricated of chipboard (paperboard) stiffeners, a minimum of 0.065 inch thick, covered throughout with black vinyl coated cloth with a minimum weight of 10.4 ounces per square yard. The surface shall have an artificial leather finish. All edges of the case shall be bound with black vinyl sheeting and stitched. The case lining and binder shall have a minimum weight of 6.4 ounces per square yard. The case shall be a flat cover folding type, with a flap measuring 2 inches in width. The case shall be provided with two commercial, approximately 7/16 inch diameter, snap fasteners. They shall be located one on each side of the coverflap, to secure the case. The fastener shall be plated or painted.

The inside lower edges of both bottle retaining areas of the case shall be provided with a valance to serve as a stop for the bottles. The inside cover of the case shall be provided with eight loops of black plastic coated cloth to retain the one ounce bottles, and the inside back of the case shall have twelve loops of the plastic coated cloth to retain the one-half ounce bottles. The bottle loops shall be one inch in width, a minimum weight of 10.4 ounces per square yard, and of the proper length to retain the bottles securely. The loops at each end of the bottle rows shall be cut to cover the portion of the bottle adjacent to the exterior of the case in order to protect the bottles when the case is closed. The bottle case shall be designed to hold the bottles securely and to be easily opened and closed.

Each of the bottles in the set shall, when filled with water and tightly capped, remain leak free in the inverted position for a period of ten minutes.

The bottle set shall be furnished with the bottles clean and dry, and the bottle case free from scratches, cuts, or tears.

Workmanship. The physician's satchel bottle set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Set (SE). One bottle set, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4.0 (percent defective) and an inspection level of I. The unit of product for examination purposes shall be one bottle set.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one bottle set.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:34Z
Check the source to verify that this is the currentversion before use.

revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

6515-00-308-8235

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-4996

Location: BTLCID13.A1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54432	25 June 1991

3. DOCUMENT TITLE

BOTTLE SET, PHYSICIAN'S SATCHEL

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	AC 215-737-2870	AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19101-8419	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — *98/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.