A-A-54436
25 June 1991

COMMERCIAL ITEM DESCRIPTION

PAD, HEAT TREATMENT

(Electric)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for MIL-P-36358B which is cancelled

This Commercial Item Description covers an electric heat treatment pad

Salient Characteristics:

The electric heat treatment pad shall be a conventional type consisting of a thermoplastic material envelope, such as vinyl, which is used to enclose a resistance wire heating element assembly Wetproof construction shall be provided by means of the envelope which shall have dielectrically or heat sealed seams to exclude moisture The waterproof envelope shall be free of perforations, cuts, tears, and ruptures It shall not exhibit tackiness, blisters, or cracks The envelope shall be capable of being cleaned with a damp antiseptic cleaning cloth Overall dimensions of the pad shall be 15 inches plus or minus 1 inch long by 12 inches plus or minus 1 inch wide

The pad heating element shall be mounted on a layer of burlap or polyester reinforced cotton insulating material, and shall be positioned in such a manner as to distribute the resulting heat evenly over both sides of the pad The pad shall operate satisfactorily on 115 volts 60 Hz AC, and shall be rated at 40-60 watts

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Thermostats, of a type designed for minimum radio interference, shall be provided to limit the maximum temperatures reached by the heating element A single thermostat connected in series with the heating element shall not be acceptable The circuitry shall be such that the thermostat shall shut off the current before the temperature of the pad exceeds the required setting The thermostats shall maintain the temperature of the pad within 120°F to 130°F at any exterior point within a rectangular area inscribed at least 2 inches from each side of the pad. The pad may be a 3-heat pad with a maximum temperature of 160°F if the low heat setting maintains the temperature between 120°F to 130°F

The heating pad shall be provided with not less than 6 feet of a suitable flexible two conductor power cord and a molded two prong plug Strain relief shall be provided on all cords to prevent any mechanical stress on the cord from being transmitted to any wiring corrections The pad, including the pad cover material, shall conform to the applicable requirements of Underwriter's Laboratories, Inc (UL) Standard of Safety UL 130, Electric Heating Pads

Each heating pad shall be furnished with two washable, easily removable, terrycloth or polyester covers with four 16 inch tie straps on each cover The pad covers shall not shrink more than six percent of the original size after they have been washed twice

The pad shall have the following warning or equivalent, marked on the face of the pad:

WARNING

BURNS WILL RESULT FROM IMPROPER USE NEVER USE PAD WITHOUT COVER IN PLACE

CAREFULLY EXAMINE BEFORE EACH USE FOLLOW INSTRUCTIONS SUPPLIED WITH THE PAD DO NOT USE IN AN OXYGEN ATMOSPHERE DO NOT USE PINS

Each pad shall have the following printed instructions, or equivalent, packed with the pad, or marked on the face of the pad:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS IMPORTANT DO NOT DESTROY

OBSERVE THESE INSTRUCTIONS CAREFULLY TO REDUCE THE POSSIBILITY OF FIRE OR ACCIDENT.

DO NOT USE IN OXYGEN ATMOSPHERE
TEMPERATURES SUFFICIENTLY HIGH TO CAUSE BURNS MAY OCCUR REGARDLESS OF THE CONTROL SETTING DO NOT USE ON AN INFANT OR INVALID OR ON A SLEEPING OR OTHERWISE UNCONSCIOUS PERSON UNLESS CAREFULLY ATTENDED DO NOT USE ON INSENSITIVE SKIN OR A PERSON WITH POOR CIRCULATION CHECK THE SKIN UNDER THE PAD FREQUENTLY TO AVOID BURNING AND BLISTERING

DO NOT REST ANY PART OF THE BODY ON PAD
BURNS CAN RESULT REGARDLESS OF CONTROL SETTING

CAREFULLY EXAMINE THE PERMANENT INNER COVERING BEFORE EACH USE DISCARD THE PAD IF THE COVERING SHOWS ANY SIGNS OF DETERIORATION (SUCH AS CHECKING, BLISTERING OR CRACKING)

NEVER USE PINS OR OTHER METALLIC MEANS TO FASTEN THIS PAD IN PLACE.

DO NOT CRUSH AVOID SHARP FOLDS

NEVER PULL THIS PAD BY THE SUPPLY CORD AND DO NOT USE THE CORD AS A HANDLE

UNPLUG WHEN NOT IN USE

Note Warning markings as outlined in Section 39 of UL 130 (Marking Details) shall be acceptable

The pad shall be free from dirt, stains, or discoloration

Workmanship The electric heat treatment pad shall be free from defects which detract from its appearance or impair its serviceability

Unit Each (EA) One pad, as specified, constitutes one unit

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions

Responsibility for inspection Unless otherwise specified in the contract or purchase order the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request at any time or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4 0 (percent defective) and an inspection level of II The unit of product for examination purposes shall be one electric heat treatment pad

Sampling for tests Sampling for tests including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2 5 (percent defective) and an inspection level of II The unit of product for test purposes shall be one electric heat treatment pad

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.
If a product is manufactured to metric dimensions and those

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

dimensions exceed the tolerances specified in the inch/pound units a request should be made to the contracting officer to determine if the product is acceptable

The contracting officer has the option of accepting or rejecting the product

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Regulatory requirements

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder In addition the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers with the requirements of all other applicable Federal State, and local statutes, ordinances and regulations

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Ordering data Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order

NOTE The following National Stock Number is covered by this document

6530-01-015-5168

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6530-2159

Location PADCID13 A1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.

2. The submitter of this form must complete blocks 4, 5, 6, and 7.

3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54436	2. DOCUMENT DATE (YYMMDD): 25 June 1991

3. DOCUMENT TITLE: PAD, HEAT TREATMENT (Electric)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19101-8419	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 96790

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.