A-A-54437
25 June 1991

COMMERCIAL ITEM DESCRIPTION

PITCHER, CLINICAL SOLUTIONS

(Aluminum)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification RR-P-350B which is cancelled

This Commercial Item Description covers a 3 quart aluminum clinical solutions pitcher

Salient Characteristics:

The pitcher body shall be seamlessly formed of a single sheet of American Iron and Steel Institute (AISI) type 3003 aluminum alloy The aluminum sheet shall be 0 064 inch thick (14 gage) which shall not be reduced below 0 046 inch thick at any point in the finished pitcher The pitchers shall be capable of being nested into each other without jamming, and shall have the following dimensions:

Height	-	8-1/2 inches plus or minus 1/2 inch
Bottom Diameter	-	5-3/8 inches plus or minus 1/2 inch
Top Diameter	-	6-1/2 inches plus or minus 1/2 inch

The rim of the pitcher shall be formed with a spout and the edge shall be rolled into a closed bead all around The pitcher shall be polished, alkaline cleaned, and anodized

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

The pitcher shall have a handle molded from a high-impact type phenolic plastic capable of withstanding autoclaving at 20 pounds per square inch for 20 minutes without distortion The handle shall be of any suitable shape and shall permit stacking of the pitchers The handle shall be attached to the pitcher in any suitable manner An aluminum handle retainer, 0 091 inch thick (11 gage), may be spotwelded to the pitcher Any screwposts or screws used for handle attachment shall be corrosion-resisting steel or chromium-plated steel All spot welds shall be ground smooth.

The pitcher body and handle retainers shall be free of cracks, tears, holes, or ripples The handle shall be free of cracks, chipped areas, or other defects The pitcher finish shall be free of pitting and other surface imperfections It shall be free of oil grease, dirt and extraneous matter

Workmanship The clinical solutions pitcher shall be free from defects which detract from its appearance or impair its serviceability

Unit Each (EA) One pitcher as specified constitutes one unit

Quality Assurance Provisions

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4 0 (percent defective) and an inspection level of II The unit of product for examination purposes shall be one clinical solutions pitcher.

Sampling for tests Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2 5 (percent defective) and an inspection level of S-2 The unit of product for test purposes shall be one clinical solutions pitcher

Tests Tests shall be conducted to determine compliance with specification requirements Where feasible, the same sample shall be used for the determination of two or more test characteristics

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376 and all other requirements of this document are met

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units a request should be made to the contracting officer to determine if the product is acceptable

The contracting officer has the option of accepting or rejecting the product

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Regulatory requirements

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

Ordering data Intermediate/exterior package quantities, labeling and marking must be specified in the contract and/or order

NOTE: The following National Stock Number is covered by this document

6530-00-299-8265

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS

Custodians:

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES

VA-OSS
USPHS
FDA-MPQAS

PREPARING ACTIVITY:

DoD-MB

Agent

DLA-DM

Project Number: 6530-2230

Location: PITCID13 A1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.

2 The submitter of this form must complete blocks 4, 5, 6, and 7.

3 The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

1. RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-54437	2 DOCUMENT DATE (YYMMDD) 25 June 1991

3 DOCUMENT TITLE
PITCHER, CLINICAL SOLUTIONS (Aluminum)

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19101-8419	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 90/790

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.