A-A-54443
12 July 1991

COMMERCIAL ITEM DESCRIPTION

PLATE, GROUNDING, ELECTROSURGICAL

The General Services Administration has authorized the use of this Commercial Item Description

This Commercial Item Description covers a disposable electrosurgical grounding plate.

Salient Characteristics:

The grounding plate shall be a thin, lightweight, disposable, flexible, conductive adhesive, dispersive electrode with a preattached cord, suitable for use with a monopolar electrosurgical unit. The plate shall be approximately 9.25 inches by 4 inches. The conductive adhesive shall cover an area of approximately 20 square inches, with a nonconductive adhesive border of approximately 1 inch. There shall be a nonconductive strip at the top of the plate near the tab.

The entire adhesive area shall be covered with a peel-off liner that protects the plate until used. The backing shall be fabricated of a thin occlusive foam. The plate shall easily conform to the body. The nonconductive border shall keep out contaminating fluid to maintain electrical isolation. The preattached cord shall be riveted to the tab and overwrapped with foam to insure a good mechanical and electrical connection. The cord shall be attached to the plate tab away from the adhesive area to eliminate tissue pressure points. The plate shall not cause thermal burns, pressure necrosis, adhesive skin stripping, allergic skin reaction or chemical skin irritation during proper use and shall be easy to remove after use. The plate shall meet the applicable requirements of ANSI/AAMI HF18-1986, Standard for Electrosurgical Devices.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Workmanship. The electrosurgical grounding plate shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing forty individual electrosurgical grounding plates, as specified, constitutes one unit. The individual package shall contain forty plates in a peel-open type envelope.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one package of electrosurgical grounding plates.

Tests. The electrosurgical grounding plate shall be tested in accordance with ANSI/AAMI HF18, Standard for Electrosurgical Devices.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTES: 1. The following National Stock Number is covered by this document.

6515-01-128-6845

2. The electrosurgical grounding plate is intended to be compatible with NSN 6515-01-130-1386, Electrosurgical Apparatus.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA - OSS
USPHS
FDA-MPQAS

Project Number: 6515-4711

Location: PLATMC16.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.