A-A-54446
12 July 1991

COMMERCIAL ITEM DESCRIPTION

# APPLICATOR, POVIDONE-IODINE IMPREGNATED

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-A-37382 which is cancelled.

## Salient characteristics.

Shall be a disposable rayon or cotton tipped applicator with a povidone-iodine solution impregnated tip, suitable for applying the solution to hard-to-reach areas including folds and creases, for use in preparing skin areas prior to injections, aspirations, surgery, catheterization, disinfecting wounds, pre-operative oral degerming, and preparing for dressing changes.

All ingredients shall be of a pharmaceutical grade.

The applicator rod shall be 4-1/8 inches long, ± 3/8 inch and 1/8 inch in diameter. Shall be made of a suitable material sufficiently flexible for the purpose.

The tip shall be cotton or rayon.

The impregnating solution shall be Povidone-Iodine Solution, USP, 10 percent.

Type I - Packaged 1 each per sealed packet. Each packet shall contain a minimum of 1.8 ml povidone-iodine solution.

Type II - Packaged 3 each per sealed packet. Each packet shall contain a minimum of 5.4 ml of povidone-iodine solution.

Type III - Packaged 3 each per sealed packet. Each packet shall contain a minimum of 5.5 ml of povidone-iodine solution.

AMSC/NA FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Workmanship. The applicators shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG).

For Size 1. Shall be supplied 50 applicators individually sealed in a packet. One package containing 50 packets, as specified, constitutes one unit.

For Size 2. Shall be supplied 75 applicators sealed 3 in a packet. One package containing 25 packets, as specified, constitutes one unit.

For Size 3. Shall be supplied 150 applicators sealed 3 in a packet. One package containing 50 packets, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory applicants or procedures) and testing (determination by technical means and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be inspection level S-1 and AQL 15 in accordance with MIL-STD-105. The unit of product for test purposes shall be one applicator.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to the degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Expiration dating required, intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

| NSN | ITEM IDENTIFICATION | SIZE | TYPE |
|---|---|---|---|
| 6510-01-008-7917 | APPLICATOR, POVIDONE-IODINE IMPREGNATED, 4-1/8 Inches Long, 150s | 3 | III |
| 6510-01-170-1177 | APPLICATOR, POVIDONE-IODINE IMPREGNATED. | 1 | I |
| 6510-01-280-9941 | APPLICATOR, POVIDONE-IODINE IMPREGNATED. | 2 | II |

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

PGC 42659

Project No. 6510-0771



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

## INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

| I RECOMMEND A CHANGE: | 1. DOCUMENT NUMBER<br>A-A-54446 | 2. DOCUMENT DATE (YYMMDD)<br>12 July 1991 |
|---|---|---|

3. DOCUMENT TITLE
APPLICATOR, POVIDONE-IODINE IMPREGNATED

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

| a. NAME (Last, First, Middle Initial) | b. ORGANIZATION | |
|---|---|---|
| c. ADDRESS (Include Zip Code) | d. TELEPHONE (Include Area Code)<br>(1) Commercial<br>(2) AUTOVON (If applicable) | 7. DATE SUBMITTED (YYMMDD) |

8. PREPARING ACTIVITY

| a. NAME<br>Defense Personnel Support Center | b. TELEPHONE (Include Area Code)<br>(1) Commercial (2) AUTOVON |
|---|---|
| c. ADDRESS (Include Zip Code)<br>ATTN: DPSC-MSE<br>2800 S. 20th Street<br>Philadelphia, PA 19101 | IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT:<br>Defense Quality and Standardization Office<br>5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466<br>Telephone (703) 756-2340 AUTOVON 289-2340 |

DD Form 1426, OCT 89 Previous editions are obsolete 198/290

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.