A-A-30016C
May 21, 1991
SUPERSEDING
A-A-30016B
March 15, 1985

COMMERCIAL ITEM DESCRIPTION

APPLICATOR, DISPOSABLE (Sterile and Nonsterile)

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification GG-A-616E.

This commercial item description describes a wooden or plastic rod, with or without an absorbent tip, as specified, suitable for use in medical and dental procedures.

Salient characteristics: Applicator, disposable, nonsterile or sterile wooden or plastic, as specified:

Type I — Wooden rod, 1/12 inch diameter plus 0.03 inch minus 0.01 inch, (0.2 cm plus 0.07 cm minus 0.02 cm), 6 inches long ±¼ inch (15.24 cm ±0.63 cm) plain, without absorbent tip, nonsterile. Applicators shall be made from tough, flexible, well seasoned, fungus-free wood, smoothly finished having a moisture content of not less than 6 percent and not more than 12 percent

Type II — Wooden rod, 1/12 inch diameter plus 0.03 inch minus 0.01 inch, (0.2 cm plus 0.07 cm minus 0.02 cm), 6 inches long ±¼ inch (15.24 cm ±0.63 cm) with absorbent fiber tip, nonsterile

Type III — Wooden rod, 1/12 inch diameter plus 0.03 inch minus 0.01 inch, (0.2 cm plus 0.07 cm minus 0.02 cm), 6 inches long ±¼ inch (15.24 cm ±0.63 cm), plain, with absorbent fiber tip, sterile.

Type IV — Wood or plastic 1/12 inch diameter plus 0.03 inch minus 0.03 inch minus 0.01 inch, (0.2 cm plus 0.07 cm minus 0.02 cm), 6 inches long ±¼ inch (15.24 cm ±0.63 cm), with absorbent fiber tip, sterile. The applicator shall be symmetrical throughout the length and the ends cut square and smooth. Wooden applicators shall be made of tough, flexible, well seasoned, fungus-free wood, smoothly finished and having a moisture content of not less than 6 percent and not more than 12 percent. Plastic applicators shall be made of a suitable plastic material and

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:35Z
Check the source to verify that this is the current version before use.

shall have sufficient flexibility to preclude breaking when bent to an angel of at least 80 degrees at the longitudinal midpoint of the applicator. One end of the applicator shall be securely tipped with cotton. The cotton shall be purified cotton, in conformance with the requirements of the United States Pharmacopiea (USP) for purified cotton. The cotton shall be approximately 3/16 inch (0.48 cm) in diameter affixed to the applicator by use of nontoxic adhesive, capable of withstanding sterilization without affecting the intended use of the item. The adhesive shall inhibit mold growth. The applicator shall be capable of withstanding chemicals such as ether and acrylic monomers used in dental laboratories.

Type V — Wooden rod, 1/8 inch diameter plus 0.03 inch minus 0.01 inch (0.3 cm plus 0.07 cm minus 0.02 cm), 12 inches long ±¼ inch (30.48 cm ±0.63 cm), plain without absorbent tip, nonsterile. Applicators shall be made from a tough, flexible, well seasoned, fungus-free wood, smoothly finished having a moisture content of not less than 6 percent and not more than 12 percent.

Type VI — Wooden rod, 1/12 inch diameter plus 0.03 inch minus 0.01 inch, (0.2 cm plus 0.07 cm minus 0.02 cm), 6 inches long ±¼ inch (15.24 cm ±0.63 cm), with absorbent fiber tip, sterile. The applicators shall be symmetrical throughout the length and the ends cut square and smooth. Wooden applicators shall be made of tough, flexible, well seasoned, fungus-free wood, smoothly finished and having a moisture content of not less than 6 percent and not more than 12 percent. One end of the applicator shall be securely tipped with cotton. The cotton shall be purified cotton, in conformance with the requirements of the United States Pharmacopiea (USP) for purified cotton. The cotton shall be approximately 3/16 inch (0.48 cm) in diameter affixed to the applicator by use of nontoxic adhesive, capable of withstanding sterilization without affecting the intended use of the item. The adhesive shall inhibit mold growth. The applicator shall be capable of withstanding chemicals such as ether and acrylic monomers used in dental laboratories.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:35Z
Check the source to verify that this is the current version before use.

Workmanship. The applicators shall be free from defects which detract from its appearance or impair its serviceability.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification: The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements

Federal Food, Drug and Cosmetic Act - If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials - The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:35Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling and marking - Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Numbers are covered by this document:

Type I - NSN 6515-00-303-8100

Type II - NSN 6515-00-059-5235

Type III - NSN 6515-00-564-8242

Type IV - NSN 6515-00-905-1473

Type V - NSN 6515-00-303-8200

Type VI - NSN 6515-01-017-2177

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Copies of United States Pharmacopedia (USP) regulations may be obtained from: USPC, Inc., 12601 Twinbrook Parkway, Rockville, MD 20852 USA.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description
(b) Type

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:35Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:35Z
Check the source to verify that this is the current version before use.