


FED. SUP CLASS
6640



MS36014-1 & -3


MS36014-2


MS36014-4

PART NUMBER	NOMINAL DIMENSIONS, INCHES (CM) A LENGTH	B HEIGHT	C WIDTH	DIAMETER OF RODS
MS36014-2	24 (61)	24 (61)	18 (46)	0.5 (1.27)
MS36014-1	48 (122)	24 (61)	18 (46)	0.5 (1.27)
MS36014-3	48 (122)	48 (122)	18 (46)	0.5 (1.27)
MS36014-4	72 (183)	48 (122)	---	0.5 (1.27)

MATERIAL: Frame rods shall be aluminum alloy (2017) in accordance with Federal Specification QQ-A-225/5.

Connectors and feet shall be of aluminum alloy (2017) in accordance with Federal Specification QQ-A-225/5 or zinc-base alloy in accordance with Federal Specification QQ-Z-363, Composition A.

FINISH: Zinc-base alloy connectors and feet shall be nickel plated in accordance with Federal Specification QQ-N-290, Class 1, Grade D.

DESIGN: All rods shall have length tolerance of ±0.06 in (±0.1524 cm) and diameter tolerance of ±0.0015 in (±0.0038 cm).

Connectors and rods shall be capable of effecting a horizontal or vertical positioned frame. Connectors shall be securely fastened to rods by suitable screws of square head or socket head design.

Frame feet shall be of the flange type of suitable size to permit secure fastening by means of three equally spaced screws.

Laboratory apparatus support frame shall consist of the following components:

Components	MS36014-1	MS36014-2	MS36014-3	MS36014-4
Rods 2" (5.1 cm)	8	8	10	6
Rods 12" (31 cm)	--	--	--	6
Rods 18" (46 cm)	2	2	3	--
Rods 24" (61 cm)	8	5	--	2
Rods 36" (91 cm)	--	2	2	--
Rods 48" (122 cm)	--	4	10	7
Rods 72" (183 cm)	--	--	--	5
Connectors	22	26	42	49
Support feet	4	4	6	8

Appropriate wrench shall be supplied for each set.

Ⓒ ENTIRE STANDARD REVISED

APPROVED 23 Dec 1959 REVISED Ⓐ 15 Dec 1967 Ⓑ 8 Oct 1975 Ⓒ 4 Dec 1985

USER ACTIVITIES: ARMY - ME, NAVY - MS, AIR FORCE - 03

REVIEWER ACTIVITIES: ARMY - MD, NAVY - AS

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document

P.A. EA	INTERNATIONAL INTEREST	TITLE FRAME, LABORATORY APPARATUS SUPPORT	MILITARY STANDARD MS 36014
Other Cust MS 99			
Procurement Specifications - NONE	SUPERSEDES:		SHEET 1 OF 2

DD FORM 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE PROJECT NUMBER 6640-1312

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:32Z
Check the source to verify that this is the current version before use.

FED. SUP CLASS

NOTES:

1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

2. Referenced documents shall be of the issue in effect on date of invitations for bid.

3. FSCM: 22527, 48619, 59728, 64484.

APPROVED REVISED

USER ACTIVITIES:

REVIEWER ACTIVITIES:

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. Other Cust	INTERNATIONAL INTEREST	TITLE FRAME, LABORATORY APPARATUS SUPPORT	MILITARY STANDARD MS 36014
Procurement Specifications -	SUPERSEDES:		SHEET 2 OF 2

DD FORM 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:32Z
Check the source to verify that this is the current version before use.