

FED. SUP CLASS
6640





PART NUMBER	DIMENSIONS OF BASE (INCH) ± 1/8		SUPPORT ROD (INCH) ± 1/8 LENGTH, ± 1/16 DIAMETER		
	A O/L WIDTH	B O/L LENGTH	C LENGTH	DIAMETER	THREAD SIZE
MS36020-1	7	13	23	0.5	1/2-13UNC
MS36020-2	9	9.6	24	0.5	1/2-13UNC

MATERIAL: Base shall be high quality porcelain capable of withstanding a thermal shock of 260°C (500°F). Base shall be glazed on top and side surfaces only. Support rod shall be copper coated steel, stainless steel, Type 302 or anodized aluminum, Type 2024 or 6061.

Feet shall be rubber.

DESIGN: Base shall be of sufficient weight to ensure dependable stability. It shall be provided with a drilled hole through midline of the base. The hole shall be of such diameter to accommodate supporting rod.

Threading of support rod shall be of sufficient length to permit a secure attachment of rod to base by means of a furnished threaded collar and nut through the drilled hole in the base or to permit insertion into completed depth of boss. Threads shall be in accordance with Federal Handbook H28 Unified National Coarse Thread Series.

Rubber feet shall be securely fastened to the four corners of the base.

Base and rod shall be procured as a unit and each size supporting rod shall be used with each corresponding size base.

ⓒ ENTIRE STANDARD REVISED

APPROVED 23 Dec 1959 REVISED Ⓐ 15 Dec 1967 Ⓑ 8 Dec 1975 ⓒ 4 Dec 1985

USER ACTIVITIES: ARMY - ME, NAVY - MS, AIR FORCE - 03

REVIEWER ACTIVITIES: ARMY - MD, NAVY - AS, AIR FORCE - 82

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited. This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		STAND, LABORATORY APPARATUS SUPPORT PORCELAIN BASE, RECTANGULAR SHAPED	MS36020
Procurement Specifications NNN-S-750	SUPERSEDES:		SHEET 1 OF 2

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-1312
REVISION DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:33Z
Check the source to verify that this is the current version before use.

FED. SUP CLASS

NOTES:

1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

2. For design feature purposes, this standard takes precedence over procurement document referenced herein.

3. Referenced documents shall be of the issue in effect on date of invitations for bid.

4. FSCM: 11273, 22527, 53629.

USER ACTIVITIES:

REVIEWER ACTIVITIES:

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

APPROVED REVISED

P.A. / Other Cust	INTERNATIONAL INTEREST	TITLE: STAND, LABORATORY APPARATUS SUPPORT PORCELAIN BASE, RECTANGULAR SHAPED	MILITARY STANDARD MS36020
Procurement Specifications	SUPERSEDES:		SHEET 2 OF 2

DD, FORM, 1 MAY 73, 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:33Z
Check the source to verify that this is the current version before use.