MIL-T-26639C
14 March 1969
SUPERSEDING
MIL-T-26639B
27 July 1961
MIL-T-26556
24 October 1958
MIL-T-26210
24 September 1957

MILITARY SPECIFICATION

TESTERS, TORQUE WRENCH

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

* 1.1 Scope. This specification covers two types of testers for checking manually operated torque-indicating handles.

* 1.2 Classification. The tester shall be of the following types and sizes:

Type I Single Station, Single Range

Size 1 0-100 ounce-inch (oz in)

Size 2 0-500 oz in

Size 3 0-100 pound-inch (lb in)

Size 4 0-250 lb in

Size 5 0-2000 lb in

Size 6 0-1000 pound-feet (lb ft)

Size 7 0-2000 lb ft

Type IA Single Station, Single Range, Dual Scale

Size 1 0-50 lb in

0-800 oz in

Size 2 0-200 lb ft

0-2400 lb in

Size 3 0-500 lb ft

0-6000 lb in

FSC 6635

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T10:05Z
Check the source to verify that this is the current version before use.

Type II Single Station, Dual Range or Dual Station, Dual Range

Size 1 0-100 oz in

0-500 oz in

Size 2 0-250 lb in

0-2000 lb in

2. APPLICABLE DOCUMENTS

* 2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein.

SPECIFICATIONS

Federal

GGG-W-641	Wrench, Socket; (And Sockets, Handles, And Attachments For Socket Wrenches; Hand)
GGG-W-686	Wrench, Torque
PPP-B-636	Box, Fiberboard

Military

MIL-P-116	Preservation, Methods Of
MIL-D-1000	Drawings, Engineering And Associated Lists
MIL-H-4034	Handles: Torque Indicating
MIL-E-4970	Environmental Testing, Ground Support Equipment, General Specification For
MIL-T-18422	Indicators, Dial And Accessories
MIL-H-26497	Handle, Torque, Low Range, Screwdriver Type

STANDARDS

Military

MIL-STD-105	Sampling Procedures And Tables For Inspection By Attributes
MIL-STD-129	Marking For Shipment And Storage
MIL-STD-130	Identification Marking Of US Military Property
MIL-STD-143	Specifications And Standards Order Of Precedence For The Selection Of

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T10:05Z
Check the source to verify that this is the current version before use.

MIL-STD-167	Mechanical Vibrations Of Shipboard Equipment
MIL-STD-831	Test Reports, Preparation Of
MIL-STD-1186	Cushioning, Anchoring, Bracing, Blocking, And Waterproofing; With Appropriate Test Methods

(Copies of specifications, standards, drawings, and publications required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

* 2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

NATIONAL BUREAU OF STANDARDS

Handbook H28 Screw-Thread Standards For Federal Services

(Application for copies should be addressed to the Superintendent of Documents, Government Printing Office, Washington DC 20402.)

OFFICIAL CLASSIFICATION COMMITTEE

Uniform Freight Classification Rules

(Application for copies should be addressed to the Official Classification Committee, Room 202, Union Station, 516 West Jackson Boulevard, Chicago IL 60606.)

3. REQUIREMENTS

3.1 Preproduction testing. This specification makes provisions for preproduction testing.

3.2 Component parts. The tester shall consist of the following component parts:

a. Drive (3.7)

b. Adapter (3.8)

c. Torque indicator (3.9)

d. Housing (3.10)

e. Calibration equipment (3.11)

3.3 Selection of specifications and standards. Specifications and standards for necessary commodities and services not specified herein shall be selected in accordance with MIL-STD-143, except as specified in 3.3.1 and 3.3.2.

Source: https://assist.dla.mil -- Downloaded: 202610-20T10:05Z
Check the source to verify that this is the currentversion before use.

* 3.3.1 Commercial parts. Commercial parts may be used where on date of invitations for bids, there are no AN and MS parts. In any case, commercial utility parts such as screws, bolts, nuts, and cotter pins may be used provided:

a. They can be replaced by the standard parts (MS or AN) without alteration.

b. The corresponding standard part numbers are referenced in the parts list and, if practical, on the contractor's drawings.

3.3.2 Standard parts. With the exception in 3.3.1, AN and MS standard parts shall be used. They shall be identified on the drawings by their part numbers.

3.4 Materials. Materials used in the construction of the tester shall be as specified herein. Materials not definitely specified shall be of uniform quality and condition and free of seams, cracks and other defects which may adversely affect the appearance, strength, endurance or wear resistance of the finished product. Materials shall not be hammered, filed or otherwise treated to conceal defects therein.

* 3.5 Design and construction. The testers covered by this specification shall be designed for checking right hand and left hand torque-indicating wrench handles conforming to MIL-H-4034, MIL-H-26497 or GGG-W-686 of sizes compatible with the specified range of the individual tester. They shall operate equally well when mounted in either the vertical or horizontal plane. Operation shall be either in the down direction, or pulled toward the operator, on the handle being checked. Each tester shall be designed to withstand a 20 percent overload on the maximum full scale reading of the tester. The testers shall be designed as single compact units having ranges as specified in 1.2. When specified by the procuring activity (see 6.2) testers provided for Navy use shall be designed and constructed to withstand the shock and vibrations classified as Type I Environment Vibrations of MIL-STD-167, without impairing their accuracy or operation.

* 3.5.1 Eccentric loading. The tester shall be so designed that a 30 degree side-loading test as specified in MIL-H-4034 and GGG-W-686 will not cause any deviation in the torque reading.

3.6 Performance.

* 3.6.1 Accuracy. The tester shall be accurate within ±1 percent of the applied load throughout the entire range of the scale.

3.6.2 Sensitivity. The tester shall have a movement capable of indicating any change in torque equal to 1/2 of a dial graduation over the entire range of the tester.

* 3.6.3 Endurance. The tester shall satisfactorily withstand 50,000 operations at 90 percent of full load at a rate of not less than 10 nor more than 20 cycles per minute. The load shall be applied with a micrometer, torque presetting audible indicating torque wrench conforming to Type III, Class I of GGG-W-686 or Type I of MIL-H-4034.

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T10:05Z
Check the source to verify that this is the current version before use.

3.6.4 Shock. The tester shall be unaffected by repeated and instantaneous release when loaded to full capacity.

* 3.6.5 Operating temperatures. The tester shall be capable of operating within a temperature of 32° Fahrenheit to 125° Fahrenheit.

* 3.6.6 Storage temperatures. The tester shall be capable of withstanding storage in temperatures ranging from -80° Fahrenheit to +160° Fahrenheit.

* 3.6.7 Hysteresis. The hysteresis effect shall be within the accuracy limits specified in 3.6.1. No preloading or overloading shall be required prior to either right or left hand operation or when changing the direction of operation.

* 3.6.8 Sensitivity. The tester's sensitivity shall be as specified in 3.6.2 when tests are performed by the dead weight method at 25 percent, 50 percent, 75 percent and 100 percent capacities of the tester.

* 3.7 Drive. Testers shall be provided with drives as specified in Table I. The two flats of square drive testers shall be parallel within ±2 degrees to the base of the tester when the tester is in an unloaded condition.

TABLE I

Type	Size	Drive	Paragraph
I	1	1/4" male hex	3.7.1
	2	1/4" male hex	3.7.1
	3	1/4" male hex	3.7.1
	4	1/2" male square	3.7.2
	5	1/2" male square	3.7.2
	6	1" male square	3.7.3
	7	1" male square	3.7.3
IA	1	1/4" male hex	3.7.1
	2	1/2" male square	3.7.2
	3	3/4" male square	3.7.3
II	1	1/4" male hex	3.7.4
	2	1/2" male square	3.7.5

* 3.7.1 Type I, sizes 1, 2 and 3; Type IA, size 1. For testers having 1/4 inch hexagon drives, the distance across flats shall not be less than 0.246 inch nor more than 0.250 inch. The drive shall extend not less than 3/4 of an inch and not more than 7/8 of an inch from the body of the tester.

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T10:05Z
Check the source to verify that this is the current version before use.

* 3.7.2 Type I, sizes 4 and 5; Type IA, sizes 2 and 3. These square drives shall conform to the dimensions and tolerances contained in Table I of GGG-W-641, except that the drive shall extend a minimum of 1 inch from the body of the tester and the spring loaded ball retainer shall be located at least 0.260 inch and not more than 0.265 inch from the end of the drive.

* 3.7.3 Type I, sizes 6 and 7. The drive of this tester shall conform to the dimensions and tolerances contained in Table I of GGG-W-641, except that the drive shall extend a minimum of 1.750 inches from the body of the tester and the spring loaded ball retainer shall be located at least 0.460 inch and not more than 0.470 inch from end of tester.

* 3.7.4 Type II, size 1. The size 1 tester shall have one or more drives conforming to 3.7.1.

* 3.7.5 Type II, size 2. The size 2 tester shall have one or more drives conforming to 3.7.2.

* 3.8 Adapters. Testers shall be provided with adapters as set forth in Table II. These adapters shall conform to Type IV, Class 5, Style A of GGG-W-641. Adapter requirements may vary according to the needs of the procuring activity. (See 6.2)

TABLE II

Type	Size	Adapter
I	1,2	1/4" female hex to 1/4" female square
		1/4" female hex to 3/8" female square
	3	1/4" female hex to 3/8" female square
	4	1/2" female square to 1/4" female square
		1/2" female square to 3/8" female square
	5	1/2" female square to 3/8" female double square
		1/2" female square to 1/2" female double square
	6	1" female square to 1" female square
		1" female square to 3/4" female square
	7	1" female square to 1" female square
IA	1	1/4" female hex to 1/4" female square
		1/4" female hex to 3/8" female square
	2	1/2" female square to 1/2" female square
		1/2" female square to 3/8" female square
	3	3/4" female square to 3/4" female square
		3/4" female square to 1/2" female square
II	1	1/4" female hex to 1/4" female double square
		1/4" female hex to 3/8" female double square
	2	1/4" female square to 1/4" female double square
		1/2" female square to 3/8" female double square
		1/2" female square to 1/2" female double square

Source: https://assist.dla.mil -- Downloaded: 202610-20T10:05Z
Check the source to verify that this is the current version before use.

* 3.9 Torque indicator. Each tester shall be provided with a means of indicating the torque applied to the drive through the use of a dial indicator. The dial indicator shall have a dial face, indicating pointer, and protective cover for the dial and pointer. A dual range or scale may be provided as necessary. The dial indicating mechanism shall meet all the requirements of MIL-I-18422.

* 3.9.1 Dial. The dial, or face, shall be fabricated from material that will not crack, fade, nor distort when subjected to the conditions of high and low temperatures encountered during service and storage. For ease of reading, the space between graduations shall not be less than 0.080 inch. The dial shall be graduated as shown in Table III.

* 3.9.2 Indicating pointer. The indicating pointer shall indicate the value of the torque applied to the tester. The indicating end of the pointer shall extend into the graduations and shall not be wider than the narrowest graduation. A means shall be provided for retaining the highest value of the torque applied until manually returned to zero.

* 3.9.3 Zero adjustment. A zero adjustment shall be provided for the indicator and all scales.

* 3.9.4 Location. The torque indicator shall be so located that the operator can easily read the applied torque from a working position.

TABLE III

Type	Size	Graduation Value (Maximum)	Numbered Graduations (0 to Full Scale)
I	1	0.1 oz in	1 oz in
	2	0.5 oz in	5 oz in
	3	0.1 lb in	1 lb in
	4	0.25 lb in	2 lb in
	5	2 lb in	20 lb in
	6	1 lb ft	10 lb ft
	7	2 lb ft	20 lb ft
IA	1	0.1 lb in	1 lb in
		1 oz in	10 oz in
	2	0.5 lb ft	10 lb ft
		5 lb in	100 lb in
	3	1 lb ft	5 lb ft
		10 lb in	100 lb in
II	1	Same as Type IA, sizes 1 and 2 (Dual Scale)	
	2	Same as Type I, sizes 4 and 5	

Source: https://assist.dla.mil -- Downloaded: 202610-20T10:05Z
Check the source to verify that this is the current version before use.

* 3.10 Housing. The mechanism of the tester shall be inclosed in a housing. The housing shall have removable panels to permit ease of inspection, maintenance and adjustment. Holes shall be provided in the housing so that the tester can be bolted to a horizontal or vertical surface.

* 3.10.1 Handle. A handle shall be attached to the housing so that the tester may be hand carried.

* 3.11 Calibration equipment. The tester shall be capable of being calibrated by a method traceable to the National Bureau of Standards (NBS). When specified by the procuring activity (see 6.2) the tester shall be provided with the necessary calibration equipment, except the dead weights for calibrating by the dead weight method.

* 3.11.1 Calibration certification. A calibration certificate shall be furnished with each tester.

* 3.12 Interchangeability. All parts having the same manufacturer's part number shall be functionally and dimensionally interchangeable. The drawing number requirements of MIL-D-1000 shall govern changes in the manufacturer's part numbers.

* 3.13 Weight. The weight of the tester excluding the calibration equipment shall not exceed the following:

Type I Single Station, Single Range

Size 1 - 25 lbs

Size 2 - 25 lbs

Size 3 - 25 lbs

Size 4 - 25 lbs

Size 5 - 25 lbs

Size 6 - 50 lbs

Size 7 - 100 lbs

Type IA Single Station, Single Range, Dual Scale

Size 1 - 25 lbs

Size 2 - 25 lbs

Size 3 - 25 lbs

Type II Single Station, Dual Range or Dual Station, Dual Range

Size 1 - 25 lbs

Size 2 - 25 lbs

Source: https://assist.dla.mil -- Downloaded: 20261O-20T10:05Z
Check the source to verify that this is the current version before use.

3.14 Screw threads. Unless otherwise specified herein, screw threads shall conform to Handbook H28.

* 3.15 Protective treatment. When materials are used in the construction of components that are subject to deterioration when exposed to climatic and environmental conditions likely to occur during service usage, they shall be protected against such deterioration in a manner that will in no way prevent compliance with the performance requirements of this specification. The use of any protective coating that will crack, chip, or scale with age or extremes of temperature encountered in storage will not be used.

3.16 Identification of product. Equipment assemblies and parts shall be marked for identification in accordance with MIL-STD-130.

* 3.16.1 Serial numbers. Each complete equipment procured under the contract shall be permanently marked with the manufacturer's serial number. This shall be a permanent marking; rubber stamped equipment serial numbers shall not be considered acceptable. Serial numbers may be on identification plates or on the backs of chassis; but in all instances the number shall be specifically identified as a serial number. Such serial numbers shall be apparent without the removal of equipment dust covers. The finished assemblies shall present the appearance of high quality laboratory instruments. All markings and nameplates shall be of such quality and permanence as to be usable throughout the expected life of the instrument and shall present a professional appearance over such period. Embossed tape labels with adhesive backing shall not be acceptable nor shall adhesive backed metal foil labels which are subject to lifting at edges.

* 3.17 Load-applying device. When required by the procuring activity (see 6.2), the Type I, sizes 3 thru 7, Type IA, sizes 2 and 3 and Type III, size 2 testers shall be equipped with a power driven load-applying device. The power required to operate the device shall be 80 pounds per square inch air (maximum) or 110 to 120 volts alternating current, or both. During testing, the application of the moving or restraining force on the handle shall be applied at 90 degrees to the axis of the handle in the plane of rotation. The load-applying device shall be conveniently and easily operated and shall not constitute a hazard to the operator or the tester under normal operating conditions. The loading device must accommodate all torque indicating handles conforming to MIL-H-4034, MIL-H-26497, or GGG-W-686 of the sizes compatible with the individual range.

* 3.18 Conversion table. A conversion table shall be provided and attached to single scale testers to convert the dial readings to other units of interest, such as lb ft to lb in, with the basic units and coverage compatible with the tester. Conversion tables shall be marked in a permanent manner on metal plates.

* 3.19 Workmanship. The tester including all parts and accessories shall be fabricated and finished in a thoroughly workmanlike manner. Particular attention shall be given to freedom from blemishes, defects, burrs and sharp edges; accuracy of dimensions and marking of parts and assemblies; thoroughness of soldering, welding, brazing, painting, alignment of parts and tightness of assembly screws and bolts.

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T10:05Z
Check the source to verify that this is the current version before use.

* 3.19.1 Castings. All castings shall be of uniform quality and free of blow holes, porosity, hard spots, shrinkage defects, cracks or other injurious defects. Strength and other essential physical properties of the castings shall be adequate throughout for the purpose intended.

* 3.19.2 Welding, brazing and soldering. Welding, brazing and soldering shall be performed in accordance with the best commercial practice. In no case shall such processes, including peening or plugging be used on castings or forgings for reclaiming any parts of the tester without authorization of the procuring activity.

* 3.19.3 Fastening devices. All screws, pins, bolts and similar parts shall be installed with an adequate means for preventing loss of proper tightness and adjustment. All such parts, when subject to removal or adjustment, shall not be swaged, peened, staked or otherwise permanently deformed.

* 3.19.4 Cleaning. The tester shall be thoroughly cleaned of loose, spattered or excess solder and metal chips. Other foreign material shall be removed during and after final assembly.

4. QUALITY ASSURANCE PROVISIONS

* 4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

* 4.1.1 Inspection records of the examination and tests shall be kept complete and available to the Government as specified in the contract or order.

* 4.2 Classification of tests. Examination and testing of the testers shall be classified as follows:

a. Preproduction test

b. Quality conformance inspection

4.3 Test conditions.

4.3.1 Qualitative analysis of test results. When more than one test of any specific nature is conducted, the results of those tests shall not be evaluated from the average of all tests, but shall be evaluated on the basis of each test of any specific nature. Any test result not within the requirements of this specification shall be sufficient justification for rejection.

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T10:05Z
Check the source to verify that this is the current version before use.

* 4.4 Preproduction testing.

4.4.1 Test sample tested by the contractor. The contractor shall subject one tester of each size procured to the preproduction tests specified in 4.4.4.

* 4.4.2 Preproduction test report. After the contractor conducts the preproduction tests, he shall prepare a preproduction test report in accordance with MIL-STD-831 and furnish a copy of the report to the procuring activity.

* 4.4.3 Preproduction test sample for the procuring activity. Along with the preproduction test report, the contractor shall submit a sample tester to the procuring activity.

* 4.4.4 Tests. Preproduction tests shall consist of all tests described under 4.6.

* 4.5 Quality conformance inspection. Quality conformance inspection shall consist of the following tests as described under 4.6.

a. Accuracy (4.6.1)

b. Overload (4.6.2)

* 4.5.1 Sampling.

* 4.5.1.1 Inspection lot. All testers of the type and range offered for delivery at one time shall be considered a lot for the purpose of inspection.

* 4.5.1.2 Sampling for examination. A random sample of testers shall be selected from each inspection lot in accordance with MIL-STD-105, General Inspection Level II. These samples shall be examined in accordance with 4.5.1.4. The acceptable quality level (AQL) shall be 1.5 percent defective.

* 4.5.1.3 Sampling for tests. A random sample of testers shall be selected from each inspection lot in accordance with MIL-STD-105, Special Inspection Level S-3. These samples shall be subjected to the tests of 4.6.1 and 4.6.2. The AQL shall be 1.5 percent defective.

* 4.5.1.4 Examination. Each of the testers selected in accordance with 4.5.1.2 shall be examined to verify compliance with this specification as to workmanship, finish, marking, and weight. Examination shall be conducted as specified in Table IV. Any tester in the sample containing one or more defects shall be rejected and if the number of defective testers in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected (see Table IV).

* 4.5.1.5 Lot acceptance tests. Each of the sample testers shall be selected and tested in accordance with 4.5.1.3 to verify compliance with the requirements of this specification. Any sample tester which does not meet the requirements for any of these characteristics shall be rejected and if the number of nonconforming testers in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T10:05Z
Check the source to verify that this is the current version before use.

TABLE IV

Classification of Defects in Accordance with MIL-STD-105

Categories	Defects
Critical:	
1	None defined.
Major:	
101	Type, size and drive square not as specified.
102	Materials defective or not as required; evident cracks, deep pits, voids, heavy scale, deep tool marks, porosity, or foreign inclusions.
103	Scale illegible or not as specified; graduations not uniform, not adequately spaced, not permanent, or graduations not identified (numbered) as required; crystal (if applicable) scratched, not transparent.
104	Pointer(s) bent, malformed, or not readily visible.
105	Male tang defective, not provided with a spring loaded ball (or plunger) or ball (or plunger) not staked in place.
106	Torque indicating mechanism cannot be readily reset to zero.
107	Removable panels not secure or cannot be readily removed for inspection and replacement of parts.
108	Tester not smoothly finished; evidence of fins, burrs, rough surfaces or sharp edges.
109	Marking, manufacturer's name or trademark missing, illegible, incorrect, or not permanent.
110	Indicator not marked in oz in, lb in, or lb ft as specified.
111	No mounting holes provided in housing for bolting testers in the horizontal or vertical position.

Source: https://assist.dla.mil -- Downloaded: 202610-20T10:05Z
Check the source to verify that this is the current version before use.

* 4.6 Test methods.

* 4.6.1 Accuracy. Accuracy tests shall be accomplished in right and left hand directions from zero through one fourth, one half, three fourths and full capacities of each scale. Additional test points may be specified as deemed necessary by the procuring activity. Repeat each test five times. When changing the direction of testing, the tester should not be preloaded, overloaded or cycled. If, when changing direction, any preloading, overloading, or cycling is required before the accuracy requirements of 3.6.1 can be met, the tester will be rejected as having failed to comply with the hysteresis requirements of 3.6.7. When inaccuracies are observed, which are greater than those specified in 3.6.1, rechecks shall be made. In the event inaccuracies are repeated, the tester shall be rejected.

* 4.6.2 Overload test. The tester shall be overloaded 20 percent of maximum capacity. This shall be repeated five times. The tester shall then be rechecked for accuracy in accordance with 3.6.1. The overload of 20 percent shall not adversely affect the accuracy of the tester.

* 4.6.3 Cycling test. The tester shall be subjected to a cycling test consisting of 50,000 cycles at 90 percent of full capacity. The cycling test shall be accomplished at a rate not less than 10 nor more than 20 cycles per minute. An operation cycle shall consist of applying a load of 90 percent of full range, releasing the load and returning the torque indicator to zero. The load shall be applied with a micrometer wrench conforming to Type III, Class I of GGG-W-686.

* 4.6.3.1 Low temperature operation and storage. The tester shall be subjected to low temperature in accordance with Procedure I of MIL-E-4970. With the temperature maintained, the tester shall be tested in accordance with 4.6.1 and 4.6.2. After return of tester to room temperature, it shall be visually inspected for compliance with 3.9.1.

* 4.6.3.2 High temperature operation and storage. The tester shall be subjected to higher temperature in accordance with Procedure I of MIL-E-4970. With the temperature maintained, the tester shall be tested in accordance with 4.6.1 and 4.6.2. After return of tester to room temperature, it shall be visually inspected for compliance with 3.9.1.

* 4.6.4 Shock test. The tester shall be loaded to full capacity and the load shall be released instantaneously. This test shall be repeated 25 times. The tester shall then meet the requirements of 3.6.1 and 3.6.2.

* 4.6.5 Load applying device test. The load applying device for the Type I, size 6 tester shall be set up in conjunction with the tester as specified by the manufacturer. A test per 4.6.1 shall be accomplished using a 2,000-pound break away type torque wrench. The load applying device shall be checked for compliance with 3.17.

* 4.6.6 Eccentric loading. The tester shall be loaded with forces that are 30 degrees from normal in the manner specified in MIL-H-4034. Under such loads, the tester shall satisfactorily meet the requirements of 3.6.1 and 3.6.2 herein.

Source: https://assist.dla.mil-. Downloaded: 2026-10-20T10:05Z
Check the source to verify that this is the current version before use.

* 4.6.7 Inclined plane. The tester shall be mounted vertically, horizontally and 45 degrees from the horizontal and operated in each position. Inability of the tester to operate satisfactorily or loss of accuracy shall be cause for rejection.

* 4.6.8 Environmental vibration (Navy testers). When required by procuring activity (see 6.2), the tester shall be subjected to environmental vibration tests in accordance with MIL-STD-167. After completion of this test, the accuracy and sensitivity tests specified in 4.6.1 and 4.6.5 shall be repeated. Inability of the tester to operate satisfactorily or loss of accuracy shall be cause for rejection.

4.7 Inspection of the preservation, packaging, packing and marking for shipment and storage. Sample items or packs and the inspection of the preservation, packaging, packing and marking for shipment and storage shall be in accordance with the requirements of Section 5, or the documents specified therein.

5. PREPARATION FOR DELIVERY

* 5.1 Preservation and packaging. Preservation and packaging shall be level A or C as specified (see 6.2).

* 5.1.1 Level A. Testers shall be preserved and packaged in accordance with MIL-P-116, Method IA and placed in a unit container conforming to Grade W6C, PPP-B-636.

* 5.1.2 Level C. Testers shall be preserved and packaged in a manner which will afford adequate protection against corrosion, deterioration and physical damage during shipment from supply source to the first receiving activity. The supplier may use his commercial practice providing it meets this requirement.

* 5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

* 5.2.1 Level A. Testers preserved and packaged as specified in 5.1.1 shall be packed in weather-resistant shipping containers conforming to PPP-B-636. The grade of each shipping container, determined by the gross weight and size, shall comply to the applicable special requirements table of the specification. Interior cushioning, blocking, bracing and waterproofing shall be in accordance with MIL-STD-1186. Closure and strapping shall be in accordance with the container specification. Containers shall be of uniform shape and size and of minimum cube and tare.

* 5.2.2 Level B. The testers shall be packed the same as level A, except that waterproofing is not required and domestic containers may be used.

* 5.2.3 Level C. Packages that require overpacking for acceptance by the carrier shall be placed in exterior type shipping containers in a manner that will insure safe transportation at lowest rate to the point of delivery. Containers shall comply with Uniform Freight Classification Rules, or regulations of other carriers as applicable to the mode of transportation.

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T10:05Z
Check the source to verify that this is the current version before use.

* 5.3 Marking. In addition to any other markings required by order or contract, the interior and exterior containers shall be marked in accordance with MIL-STD-129.

6. NOTES

* 6.1 Intended use. The tester covered by this specification is intended for use in machine shops, ground installations and inspection facilities to calibrate torque handles and to verify torque handle settings.

* 6.2 Ordering data. Procurement documents should specify the following:

a. Title, number and date of this specification.

b. Selection of applicable levels of preservation, packaging and packing (see 5.1 and 5.2).

c. Type and size of tester required (see 1.2).

d. If calibration equipment is required (see 3.11).

e. If a power driven load applying device is required (see 3.17).

f. Applicable environmental vibration tests (see 4.6.8).

g. Adapters required (see 3.8).

* 6.3 Supersession data. Types and sizes covered in this specification replace sizes of previous editions as follows:

a. Type I, size 5 replaces size 3 of MIL-T-26639B.

b. Type I, size 6 replaces testers previously covered in MIL-T-26210.

c. Type II, size 1 replaces testers previously covered in MIL-T-26556.

d. Type II, size 2 replaces size 1 of MIL-T-26639B.

e. Type II, size 3 replaces size 2 of MIL-T-26639B.

In addition, Type I, sizes 1, 2, 3 and 4 are included in this revision to cover ranges of lesser torque (see 1.2).

* 6.4 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T10:05Z
Check the source to verify that this is the current version before use.

Custodians:
Army - GL
Navy - AS
Air Force - 82

Review Activities:
Army - AV, GL
Navy - AS, SH
Air Force - 82

Preparing Activity:
Air Force - 82

User Activities:
Army - CE, MI
Navy - MC, OS

Project No. 6635-0038

☆ U. S. GOVERNMENT PRINTING OFFICE: 1977-703-020/942

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T10:05Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED S/N 0102-014-1802

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T10:05Z
Check the source to verify that this is the current version before use.