MIL-B-36270B
3 May 1976
SUPERSEDING
MIL-B-36270A
10 March 1969

MILITARY SPECIFICATION

BANDAGE, GAUZE, CALAMINE-GELATIN-GLYCERIN-ZINC OXIDE IMPREGNATED

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a Calamine-Gelatin-Glycerin-Zinc Oxide Impregnated Gauze Bandage (see 6.2).

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of the specification to the extent specified herein:

SPECIFICATIONS

FEDERAL

PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-636	- Boxes, Fiberboard.
PPP-B-676	- Boxes, Setup.

STANDARDS

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

FSC 6510

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

U.S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE,
FOOD AND DRUG ADMINISTRATION

Federal Food, Drug, and Cosmetic Act and Regulations
Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, DC 20204.)

U. S. PHARMACOPEIAL CONVENTION, INC.

The United States Pharmacopeia.

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

National Formulary

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

3. REQUIREMENTS

3.1 Material. Shall be a rolled gauze bandage impregnated with a calamine-gelatin-glycerin-zinc oxide containing paste.

3.1.1 Paste. Each gauze bandage shall be impregnated with not less than 170 grams and not more than 230 grams of the calamine-gelatin-glycerin-zinc oxide containing paste when determined as specified in 4.6.1.

3.1.2 Zinc oxide content. The zinc oxide content shall be not less than 3.5 percent and not more than 10.1 percent (w/w) of the calamine-gelatin-glycerin-zinc oxide impregnated gauze bandage when assayed as specified in 4.6.2.

3.1.3 Microbial limit. The bandage shall contain no vegetative cells and shall contain not more than 50 aerobic spore-forming organisms when tested using the U.S.P. microbial limit test.

3.1.4 Preservatives. The bandage shall contain suitable preservatives.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

3.1.5 Moisture and adhesion test. When tested as specified in 4.6.3, the bandage shall have sufficient moisture for smooth application and shall adhere to the area being wrapped and to itself when allowed to dry.

3.2 Ingredients.

3.2.1 Gauze bandage. The gauze bandage shall be made of suitable cotton materials measuring approximately 4 inches in width by 10 yards in length, weighing a nominal 30 grams to 38 grams. The gauze bandage shall be loosely woven, may have woven edges, and shall be free of excessive fraying and lint.

3.2.2 Other ingredients. All other ingredients entering into the finished gauze bandage shall be of U.S.P. or N.F. quality or, if not included in either of these compendia, the ingredients shall be suitable for use in this preparation.

3.3 Expiration dating period. The expiration dating period shall be 24 months.

3.3.1 Delivery. Not less than 21 months of the expiration dating period shall remain at time of delivery to the Government.

3.4 Workmanship. The material and its containers shall be free from defects which detract from their appearance or may impair their serviceability

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and, (i) as to any expiration dated item, for the expiration dating period (potency period) specified by the contractor for such item, or such longer period as may be required by regulation of any federal agency; or (ii) as to non-dated items, for not less than 3 years after delivery of the items to the Government.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

4.1.2 No company supplying any ingredient(s) to the contractor shall be considered an acceptable facility for the performance of any inspection requirements specified herein.

4.2 Lot. For purposes of this specification, a lot, batch, or control is that single, uniform, and homogeneous quantity of impregnated gauze bandage produced from one formulation, subjected to the same compounding and manufacturing operation, and packed into final containers.

4.3 Sampling. Sampling shall be conducted in accordance with MIL-STD-105.

TABLE I. Sampling.

For visual examination	Inspection level	AQL (percent defective)	Unit of product
Major	S-2	1.0	Unit of issue
Minor	II	2.5	Unit of issue

4.3.1 For raw material and end item testing. Noncompliance with any of the following is cause for rejection.

TABLE II. Raw material and end item testing.

Characteristic	Requirement paragraph	Inspection level 1/	Minimum sample
Paste	3.1.1	S-2	5 } composite
Zinc oxide content	3.1.2	S-2	5 }
Microbial limit	3.1.3	-	-
Moisture and adhesion	3.1.5	-	-

1/ Where possible, the sample shall be used for two or more tests.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

4.4 Examination. The finished gauze bandage shall be examined to determine compliance with all requirements of this specification. Nonconformance will be permitted to the extent indicated in 4.3.

4.4.1 Classification of defects. Examination will be conducted in accordance with the following classification of defects:

TABLE III. Classification of defects.

Categories	Defects 1/
Major	
101	Bandage not free from foreign matter.
102	Bandage not free of excessive lint.
103	Bandage has excessive fraying.
104	Dimensions of bandage not as specified.
105	Bandage not thoroughly impregnated.
Minor	
201	Exterior of immediate container is not free from contents or foreign matter.
202	Label not free from defacing marks and tears.

1/ Inspection is not restricted to the classified possible defects listed above.

4.5 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of section 5. Defects shall be scored as specified in table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality shall be 4.0 defects per hundred units. The examination is applicable only to the intermediate and exterior containers.

TABLE IV. Examination for preparation for delivery.

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging & packing materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:19Z
Check the source to verify that this is the current version before use.

4.6 Tests.

4.6.1 Paste. Determine the combined average weight of ten dry gauzes and ten empty bags. Weigh each finished unit package and subtract the combined average gauze and bag weight. The result shall be the weight of the paste per bandage.

4.6.2 Zinc oxide content. Weigh accurately about 1.5 grams of the calamine-gelatin-glycerin-zinc oxide impregnated gauze bandage into a porcelain crucible and heat gently until the paste is melted. Continue heating, gradually raising the temperature, until the mass is thoroughly charred. Ignite the mass until the residue is uniformly yellow, and cool. Dissolve the residue in 10 ml of dilute sulfuric acid, warming, if necessary, to effect complete solution. Transfer the solution to a beaker, and rinse the crucible with small portions of water until the combined solution and rinsings measure about 50 ml. Add 15 ml of ammonia-ammonium chloride TS and 1 ml eriochrome black TS, and titrate with 0.05M disodium ethylenediaminetetraacetate until the solution is blue. Each ml of .05M disodium ethylenediaminetetraacetate is equivalent to 4.068 mg of zinc oxide.

4.6.3 Moisture and adhesion test. First apply the bandage around the foot between the toes and the ankle with a circular turn and direct it obliquely towards and around the heel. Repeat the application of the bandage until the foot is adequately covered by one horizontal turn of the bandage followed by an oblique half-turn. The first layers around the foot should be snug and cut frequently during its application. Then apply the bandage to the leg with the greatest pressure applied to the ankle and lower third of the leg with progressively diminishing pressure over the upper two thirds. The pressure is least over the toes. Each turn should overlap one-half of the preceding turn. Cover the leg approximately three times, completing within an inch to two below the knee. Allow the bandage to harden.

5. PACKAGING

5.1 Packaging. Packaging shall be level A, or C, as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Unit package. Each bandage shall be neatly and evenly rolled and packaged in a sealed polyethylene bag of an appropriate size. Each bandage shall be further packaged in a snug-fitting box constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors are acceptable. Closure of the box shall be adequate to prevent spilling of contents under normal handling.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

5.1.1.2 Intermediate package. Twelve unit packages shall be packaged in a box constructed in accordance with PPP-B-566 or PPP-B-676. Closure shall be adequate to prevent accidental opening.

5.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.1).

5.2.1 Level A. Six intermediate packages (72 units) shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, class weather-resistant. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

5.2.2 Level B. Six intermediate packages (72 units) shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Unit packages. In addition to any commercial markings, each unit package shall be marked in accordance with the requirements of the Federal Food, Drug and Cosmetic Act and the requirements of MIL-STD-129. In addition, type II shelf life markings as specified in MIL-STD-129 shall be shown. Marking shall include the lot (control) number, and the legends: "STORE AT CONTROLLED ROOM TEMPERATURE -15° - 30° C (59° - 86°F)" and "DO NOT PERMIT TO FREEZE".

5.3.2 Intermediate package. Each intermediate package shall be marked in accordance with MIL-STD-129. Contract or purchase order number, lot (control) number and name of contractor shall be shown. Type II shelf life markings as specified in MIL-STD-129 shall be shown. Marking shall include the legends: "STORE AT CONTROLLED ROOM TEMPERATURE -15° - 30°C (59° - 86°F)" and "DO NOT PERMIT TO FREEZE".

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

5.3.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Type II shelf life markings as specified in MIL-STD-129 shall also be shown. Markings shall include the lot (control) number and the legends: "STORE AT CONTROLLED ROOM TEMPERATURE -15° - 30°C (59° - 86°F)" and "DO NOT PERMIT TO FREEZE".

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) the National Stock Number.
(c) Selection of applicable levels of packaging and packing (see 5.2 and 5.3).

6.2 This specification covers the following item(s):

National Stock No.	Item Identification
6510-00-450-5605	BANDAGE, GAUZE, CALAMIDE-GELATIN-GLYCERIN-ZINC OXIDE IMPREGNATED, Single-Ply, 4 inches by 10 yards.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
DSA-DM

Project No. 6510-0613

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS. The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-B-36270B BANDAGE, GAUZE, CALAMINE-GELATIN-GLYCERIN-ZINC OXIDE IMPREGNATED

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

☆ GPO 1974— 603-880. 6128 2-1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the current version before use.