MIL-F-36271(DM)
27 August 1964

MILITARY SPECIFICATION

FOOTPRINTER, INFANT IDENTIFICATION, DISPOSABLE, 50s

1. SCOPE

1.1 This specification covers one type of disposable individual infant footprinter, suitable for positive, permanent records.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up, Paperboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (for Carton Sealing).

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. Material shall be of the highest quality throughout, and suitable in all respects for the purpose intended.

3.2 Construction and design. The footprinter shall consist of a pad and a record card.

3.2.1 Pad. The felted paper pad shall carry a sufficient amount of chemical solution for two infants' footprints and one mother's fingerprint. The pad shall be 2-1/2 inches wide $\pm$ 1/16 inch by 3-1/2 inches long $\pm$ 1/16 inch. The pad shall be packaged in a container of appropriate size fabricated from two sheets of polyester film-coated aluminum foil laminated to a base of stiff sulfite board. The foil sheets, with the coating on interior sides shall be heat sealed around the edges in such a manner as to provide a satisfactory moisture barrier. In addition, a free tab on the top edge of the outside foil sheet shall provide means of opening the container by pulling.

3.2.2 Record card. A sensitized paper record card whereon an image is developed by chemical action between the solution transferred from the pad by the infant's foot or the mother's finger and the chemicals contained in the card. The card shall be 7 inches wide by 8-1/2 inches long. The card shall provide a positive, permanent record that shall not rub off or smear in normal handling.

3.3 Count. The unit shall consist of 50 footprinters and 50 record cards. Not less than 48 each, cards and pads, shall be furnished. The average number per unit shall be not less than 50.

3.4 Marking. Each card shall be marked as shown below. Each pad shall bear the name or registered trademark of the manufacturer. Marking shall be permanent, distinct, and legible.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the current version before use.

FOOT-PRINT RECORD

_____
HOSPITAL

City _____ State _____

Name of Child _____ Hospital No. _____

Sex _____ Birth at _____ A.M. P.M. _____

Name of Mother _____ Date _____

LEFT FOOT	RIGHT FOOT	Mother's Right Thumb Print

(Name or Registered Trademark of Manufacturer)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

3.5 Workmanship. Workmanship shall be first class throughout. The tables and accessories shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

	Inspection level	AQL (defects per hundred units)
Major defects	II	1.0
Total (major and minor defects combined)	II	4.0

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:19Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105.

Component (unit of product)	Characteristic	Requirement	Test procedure	Inspection level	AQL
End item	Imprint	3.2	4.4.1	S-1	*

* Acceptance number shall be zero (0).

4.3 Examination. The footprinters shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects:

Categories and defects*

Major
101 Footprinter not complete with all required component parts.
102 Style, design, and dimensions of footprinter not as specified.
103 Pad not suitably heat sealed around the edges.
104 Record card not suitably sensitized.

Minor
201 Tab for easy opening not accessible or not easy to operate.
202 Identification marking not present or not complete with specified information.

* Examination shall not be restricted to the classified possible defects listed above.

4.3.2 Dimensional examination. The footprinters shall be examined for defects in dimensions. Any dimension not within the tolerance specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Imprint test. The selected samples of the pads and record cards shall be tested by pressing the thumb on the pad and transferring it to the card. A clear and distinct image should be observed which will not rub off or smear.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. Fifty (50) footprinter pads and fifty (50) record cards, as specified, constitute one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I. Packaging Quantities

Unit package	Intermediate package	Exterior container
Not required	2 units	60 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit preparation.

5.1.3.1.1 Footprinter pads. Fifty (50) pads packaged as specified in 3.2.1, shall be further packaged in a box of appropriate size constructed in accordance with PPP-B-566, except that commercial colors shall be acceptable. Closure shall be adequate to prevent spillage of contents under normal handling.

5.1.3.1.2 Record cards. Fifty (50) cards shall be packaged in a kraft envelope. Closure of envelope shall be adequate to prevent spillage of contents under normal handling.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the current version before use.

5.1.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with PPP-B-636 or PPP-B-676. Commercial colors shall be acceptable when carton conforming to PPP-B-676 is utilized. Contents shall be packaged in such a manner to insure against movement of, or damage to units. Closure shall be adequate to prevent spillage of contents under normal handling.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF class weather-resistant.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix of PPP-B-636. In addition, all seams, including manufacturer's joint, shall be closed by tape conforming to the requirements of PPP-T-76. The tape shall be not less than 2 inches wide. Three strips shall be applied to both top and bottom so that all seams and joints shall be covered their full length. Tape applied over the lengthwise seams shall extend a minimum of 2 inches onto each of the end panels. Tape applied over the end edge seams shall extend a minimum of 2 inches along the side and top, and side and bottom panels.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood, or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use, PPP-B-621, table I; or PPP-B-636, type CF, class domestic.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the currentversion before use.

5.3 Marking.

5.3.1 Carton and envelope. Each carton and envelope within the unit package shall be identified as to contents.

5.3.2 Intermediate packages. Each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. In addition, the letter "L" shall precede the Federal stock number. One space shall be allowed between the letter "L" and the first digit of the stock number. The terms "Stock number" or "FSN" shall not be indicated.

NOTE: The unit of issue and item identification on unit, intermediate, and exterior containers shall be designated as follows:

	Quantity and Unit	Item Identification
Intermediate package	2 PKG (2 units of 50 pads and 50 cards each.)	FOOTPRINTER, INFANT IDENTIFICATION Disposable, 50s
Exterior container	60 PKG	FOOTPRINTER, INFANT IDENTIFICATION, Disposable, 50s

6. NOTES

6.1 *Ordering data.* Procurement documents should specify the following:

(a) Number and date of this specification.

(b) Selection of applicable levels of packing (see 5.1 and 5.2).

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the currentversion before use.

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Custodians:
Army. - MD
Navy. - MS
Air Force. - 03

Preparing activity:
Defense Supply Agency. - DM

Review activities:
Army. - MD
Navy. - MS
Air Force. - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

FORM APPROVED BUDGET BUREAU NO. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use of the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity.

SPECIFICATION

MIL-F-36271(DM), dated 27 August 1964, FOOTPRINTER, INFANT, IDENTIFICATION, DISPOSABLE, 50

ORGANIZATION	CITY AND STATE

CONTRACT NO.	QUANTITY OF ITEMS PROCURED	DOLLAR AMOUNT
		$

MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO IF "YES" IN WHAT WAY?

4. REMARKS (ATTACH ANY PERTINENT DATA WHICH MAY BE OF USE IN IMPROVING THIS SPECIFICATION. IF THERE ARE ADDITIONAL PAPERS, ATTACH TO FORM AND PLACE BOTH IN AN ENVELOPE ADDRESSED TO PREPARING ACTIVITY)

SUBMITTED BY (PRINTED OR TYPED NAME AND ACTIVITY)	DATE

DD FORM 1426

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

FOLD UNDER

DEFENSE SUPPLY AGENCY
DEFENSE MEDICAL SUPPLY CENTER
3RD AVENUE AND 29TH STREET
BROOKLYN, N.Y., 11232

OFFICIAL BUSINESS

POSTAGE FEES PAID
DEFENSE SUPPLY AGENC

DEFENSE SUPPLY AGENCY

DEFENSE MEDICAL SUPPLY CENTER

ATTN: DIRECTORATE OF TECHNICAL OPERATIONS (CODE TM5)

3RD AVENUE AND 29TH STREET

BROOKLYN, N.Y., 11232

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the current version before use.