MIL-F-3[illegible]4A
7 August 196[illegible]
Superseding
MIL-F-3[illegible]4 (DI:)
15 Septe[illegible]ber 1964

MILITARY SPECIFICATION

FLASKS, DENTURE, UPPER AND LOWER, EJECTOR TYPE

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers ejector type flasks suitable for processing upper and lower dentures.

1.2 Classification. Ejector type denture flasks shall be of the following types, as specified (see 6.1).

Type I. Upper.
Type II. Lower.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

QQ-B-626 Brass, Leaded and Nonleaded: Rod, Shaped, Forgings, and Flat Products with Finished Edges (Bar and Strip).
PPP-B-5[illegible]5 Boxes, Wood, Wirebound.
PPP-B-[illegible]01 Boxes, Wood, Cleated-Plywood.
PPP-B-[illegible]21 Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636 Box, Fiberboard.
PPP-B-665 Boxes; Paperboard, Metal Stayed (Including Stay Material).
PPP-B-676 Boxes, Setup.

Military

MIL-L-10547 Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:20Z
Check the source to verify that this is the current version before use.

STANDARDS

Federal

FED-STD-66 Steel: Chemical Composition and Hardenability.

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

DRAWINGS

DEFENSE MEDICAL SUPPLY CENTER
20485 Flask, Denture, Ejector Type, Upper.

DEFENSE PERSONNEL SUPPORT CENTER
20125 Flask, Denture, Ejector Type, Lower.

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.

3.1.1 Brass. Brass for forging the flasks shall conform to copper alloy number 377 of QQ-B-626.

3.1.2 Steel. Steel for the guide pins shall conform to type 303 corrosion resistant steel of FED-STD-66.

3.2 Style, design, and dimensions.

3.2.1 Type I, upper flask. The style, design, and dimensions of type I, upper flask, shall be in accordance with Defense Medical Supply Center (DMSC) Drawing 20485. Type I flask shall be suitable for processing upper dentures.

3.2.2 Type II, lower flask. The style, design, and dimensions of type II, lower flask, shall be in accordance with Defense Personnel Support Center (DPSC) Drawing 20125. Type II flask shall be suitable for processing lower dentures.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:20Z
Check the source to verify that this is the currentversion before use.

3.3 Construction. The denture flasks shall consist of a top, middle, and bottom section. The middle section shall contain two tapered guide pins to maintain registration between the sections. The bottom section shall contain a removable disk.

3.4 Finish. Flasks shall have a smooth finish. All matching surfaces shall be machined to insure a close fit and preclude separations between the three sections of the flask. Steel for the guide pins shall be ground, polished, and annealed.

3.5 Identification marking. Each ejector type flask shall be serialized, and the top, middle, and bottom section of the flask shall be permanently stamped with the serial number as shown on the drawings. In addition, as part of the forging operation, the name or registered trademark of the manufacturer and the word "UPPER" (for type I flask) or "LOWER" (for type II flask) shall be permanently and legibly marked on the top section of each flask.

3.6 Workmanship. Workmanship shall be first class throughout. The flasks shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Quality conformance inspection. The examinations required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in tables I and II (see 4.2.1).

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:20Z
Check the source to verify that this is the current version before use.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in table I. Unit of product for examination purposes shall be one flask.

Table I. Examination Criteria

Type defect	Inspection level	AQL (percent defective)
Major A	II	1.0
Major B	S-2	1.5
Minor A	II	2.5
Minor B	S-2	4.0

4.2.2 Sampling for raw materials. Sufficient amount of each lot of each metal used in the construction of the flask shall be selected for chemical analysis.

4.3 Examination. Flasks shall be examined for defects including, but not limited to, those listed in table II.

Table II. Classification of Defects

Category	Defect
Major A:	
101	Type not as specified.
102	Matching surfaces of three sections of flask not machined properly.
103	Any component of flask missing or not of design as shown in applicable drawing.
104	Cracks, breaks, or other workmanship defects seriously affecting proper fit and function of flasks present.
Major B:	
151	Dimensions affecting fit and alignment of flask not as shown on applicable drawing.
Minor A:	
201	Name or trademark of manufacturer and the word "UPPER" (for type I) or "LOWER" (for type II) not present on top section of flask.
202	Top, middle, and bottom sections of flask not stamped with serial number.
203	Finish (other than 102) not as specified (see 3.4).
204	Burs, scratches, or sharp edges present.
205	Packaging and packing not as specified.
Minor B:	
251	Dimensions (other than 151) not as specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:20Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One flask, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table III shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table III. Packaging Quantities.

Unit package	Intermediate package	Exterior container
1 unit	Not required	24 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior contained of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive materials and packaged in a carton of appropriate size constructed in accordance with PPP-B-665 or PPP-B-676. Commercial colors are acceptable on cartons and closure shall be adequate to prevent spilling of contents under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type I, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:20Z
Check the source to verify that this is the current version before use.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4, of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic and method I closure shall be utilized.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, constructed so as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) Type required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not cover all types, classes, grades or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Military requirements.

Source: https://assist.dla.mil -- Downloaded: 20261O-21T08:20Z
Check the source to verify that this is the current version before use.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DH

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:20Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:20Z
Check the source to verify that this is the current version before use.

	Form Approved Budget Bureau No. 22-R355

INSTRuctor, involved in the use of the specification is provided for obtaining information on the used minimum amount of delay and at the least cost side, staple in corner, and send to preparing activity authorization to waive any portion of the

SPECIF

ORGANI

CITY AI

MATER

1. NASTION IN PROCUREMENT USE?

A.

B.

2. CON

3. IS T

☐

4. PEN are additional papers, attach to form and finct

SUBMIT	DATE

nn

Preparing activity:
Defense Supply Agency - DN

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:20Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL CODE: ATS
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

☆ U. S. GOVERNMENT PRINTING OFFICE: 1968-341-511/A-1361

DPSC FORM 2416
AUG 68

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:20Z
Check the source to verify that this is the current version before use.