MIL-I-36254A(DM)
9 October 1986
SUPERSEDING
MIL-I-36254 (DM)
7 August 1964

MILITARY SPECIFICATION

INSERT CABINET, MEDICAL INSTRUMENT AND SUPPLY SET CHEST, COMPARTMENTED, 7 DRAWER, WITH FOUR LEGS, TWO SPACERS AND TWO TIE DOWN RODS

This Military specification draft, dated, 9 October 1986, prepared by the Defense Personnel Support Center, Philadelphia, PA 19101, has not been approved for promulgation and is subject to modification.
DO NOT USE FOR PROCUREMENT PURPOSES

1. SCOPE

1.1 Scope. This specification covers a metal insert of drawers, intended to be fastened by means of screws in the standard medical field chest (see 6.1).

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

AMSC N/A FSC 6545

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, by using the self- addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

(DD Form 1426 to be included in the final document.)

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Federal

QQ-A-200/8	Aluminum Alloy Bar, Rod, Shapes, and Tube, Extruded, 6061 and 6062.
QQ-A-200/9	Aluminum Alloy Bar, Rod, Shapes, and Tube, Extruded, 6063.
QQ-A-250/8	Aluminum Alloy, 5052, Plate and Sheet.
QQ-S-763	Steel Bars, Shapes, and Forgings, Corrosion-Resisting.
QQ-S-766	Steel Plates, Sheets, and Strip-Corrosion-Resisting.
TT-E-527	Enamel, Alkyd, Lustreless.
TT-T-306	Thinner, Synthetic-Enamel.
WW-T-700/6	Tube, Aluminum Alloy, Round, Square, Rectangular, and other Shapes, Drawn, Seamless, 6061 and 6062.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

Military

MIL-P-116	Preservation, Methods of.
MIL-C-5541	Chemical Films for Aluminum and Aluminum Alloys.
MIL-P-8585	Primer Coating, Zinc Chromate, Low-Moisture-Sensitivity.
MIL-A-8625	Anodic Coatings, for Aluminum and Aluminum Alloys.
MIL-L-10547	Liners, Case, and Sheet, Overwrap, Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Federal

Fed. Test Method Std. No. 151	Metals; Test Methods.
Fed. Std. No. 595	Colors.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.
MIL-STD-147	Palletized Unit Loads.
MIL-STD-794	Parts and Equipment, Procedures for Packaging of.
MS16686	Medical Services Symbol

DRAWING

Defense Personnel Support Center (DPSC) Drawing 10025

Insert Cabinet, Medical Instrument and Supply Set
Chest, Compartmented, 7 Drawer (and Bill of Materials)

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer).

3. REQUIREMENTS

3.1 Materials. Except as specified herein, materials shall conform to the applicable portions of the specifications, standards, and drawing (with Bill of Materials) listed in section 2, and be subjected to the tests and meet the requirements specified therein. In addition, the metals shall conform to the chemical requirements of their applicable specifications when tested as specified in 4.4.1. When materials to be used are not specifically designated, those used shall be entirely suitable for the purpose intended.

3.1.1 Aluminum alloy sheets. Aluminum alloy sheets for drawer backs, false bottom and upper plate frame shall conform to alloy number 5052, temper H-32 of QQ-A-250/8. Aluminum alloy sheets for drawer bodies, fronts, partitions, racks, dust covers, arms, locking plate, locking angle, bracket frames, trays, case and cover, clamps, retainer, spacers, hammer clip, strap, stop, and stop brackets shall conform to alloy number 5052, temper H-34 of QQ-A-250/8.

3.1.2 Aluminum alloy extruded. Extruded aluminum alloy for rod, container, lower locking plate, knob shanks, and knobs shall be alloy number 6061, temper T-6 of QQ-A-200/8. Alloy for the channel for the frame shall be alloy number 6063 temper T-5 of QQ-A-200/9.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

3.1.3 Aluminum alloy tubing. Aluminum alloy tubing for the container and the legs shall be alloy number 6061, temper T-6 of WW-T-700/6.

3.1.4 Steel, corrosion-resisting. Steel for the locking bar, lock bracket, bracket, and clamp shall be class 304 corrosion-resisting steel of Specification QQ-S-766. Steel for the book, clamping bar, and leg book shall be class 304 corrosion-resisting steel of QQ-S-763.

3.1.5 Plastic. The plastic for the basin shall be molded of extra high impact polystyrene plastic.

3.1.6 Primer. Primer shall conform to color Y of MIL-P-8585.

3.1.7 Enamel. Enamel shall conform to TT-E-527. Color of the enamel shall be 34087 (lustreless olive drab) as shown in Federal Standard No. 595 (see 6.3), except that for marking, as specified in 3.2.3, the color shall be 37038 (lustreless black) of the same standard (see 6.4).

3.2. Construction. Each insert cabinet shall consist of 7 compartmented drawers, four legs, two spacers, and two tie down rods. The cabinet and plastic container shall conform in all details to DPSC Drawing No. 10025, except that part numbers 91 and 92 shall be excluded from this specification. The construction shall be sound and durable in all respects. The insert cabinet assembly shall be welded on jigs capable of maintaining required tolerances on the finished insert.

3.2.1 Welding. Welding shall be done by qualified welders using appropriate equipment in a qualified process. Spot welds shall be capable of developing the tensile strength of the parent metal. All exposed spot welds shall be smoothed sufficiently to blend with the adjacent surfaces. All welds shall be finished in a clean and workmanlike manner. All welds shall be clean, nonporous, sound, thorough, and ground and finished smooth to match adjacent surfaces. All welds shall be free from pitmarks and imperfections. All surfaces shall be free from sharp edges. Other welding shall be continuous on all surfaces, using welding rods of composition similar to parent metal, and suitable for the purpose intended.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

3.2.2 Finish.

3.2.2.1 Aluminum alloy. After welding has been completed, but before final assembly, all accessible aluminum surfaces of the insert cabinet, parts and components, except ferrous metal parts, plastic basin, and the working surfaces of the drawer slides shall be finished with enamel as specified herein. All aluminum alloy surfaces shall be thoroughly cleaned free from dirt, oil, grease, welding flux and loose scale with solvents, detergents, or other processes which have no deleterious effect on the surfaces.

After thorough cleaning and degreasing, surfaces to be enameled shall be prepared for topcoating by one of the following three methods:

(1) Chemical treatment in accordance with MIL-C-5541.

(2) Anodic treatment in accordance with MIL-A-8625.

(3) Priming treatment, using primer specified in 3.1.6, thinned and applied by spray in accordance with the directions specified under "Preparation for Delivery" in MIL-P-8585, and air-dried for not less than 30 minutes before topcoating.

Topcoating shall be with one coat of enamel as specified in 3.1.7. The enamel shall be reduced by taking four parts by volume of enamel with one part by volume of thinner in conformance with TT-T-306. The reduced enamel shall be applied by spraying and cured by a heat treatment equivalent to baking at 250°Fahrenheit for not less than 45 minutes. The resultant finish shall be hard, smooth, and highly resistant to abrasion, chipping, cracking, acid, alkali, and solvents.

3.2.2.1.1. Film thickness of paint. Contractor shall endeavor to control film thickness so that the thickness of a single dried coat of paint is 0.3 to 0.7 for primer and 1.0 to 1.5 mil for enamel.

3.2.2.2 Ferrous metal. All ferrous metal parts shall be finished , before final assembly, by thorough cleaning and chemical degreasing, followed by a coat of primer as specified in 3.1.6 and then by a coat of enamel as specified in 3.1.7. The enamel may be omitted where it would interfere with the mating of threaded elements.

3.2.3 Identification markings. The insert cabinet shall be marked with enamel as specified in 3.1.7, as follows:

Centrally located on the front of the third drawer from the top, a 1 inch medical services symbol as shown for type I in MS16686.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

Centrally located on the upper surface of the top bracket frame, there shall be located the following:

In vertical block letters, 3/8 inch high, the legend:

"Insert Cabinet, Medical Instrument and Supply Set Chest, Compartmented, 7 Drawer"

Below the legend, in vertical block numerals 3/8 inch high, the stock number "6545-00-663-1549".

Below the stock number, in vertical letters 1/4 inch high, the name or registered trademark of the manufacturer.

Below the name or registered trademark of the manufacturer, in vertical letters 1/4 inch high, the contract or order number.

3.3 Workmanship. Workmanship shall be first class throughout. Frame and drawers shall be square and free from twist. Drawers shall operate smoothly and easily. Completed insert cabinets shall be free from defects which detract from their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105.

	Inspection level	AQL (percent defective)	Minimum sample size
Component			
Material testing	S-1	*	3

*Acceptance number shall be zero (0) for all sample sizes.

4.2.3 Raw materials. Four ounces of each lot of (each) metal used in the construction of the insert cabinet shall be selected for chemical analysis as specified in 4.4.1.

4.2.4 Components. Sufficient samples of the components shall be selected from each lot of material used in the manufacture of the insert cabinets to perform tests as required by the applicable subsidiary specifications or by tests specified herein. When sample sized for components covered by the subsidiary specifications is not specified, the sample size for each test shall be as indicated in 4.2.2. The unit of product for testing shall be the quantity of material necessary to perform all required tests one time each, taking into consideration the fact that, where possible, the same material may be used to perform two or more tests.

4.2.5 Where possible, the same sample unit shall be used for the determination of two or more tests characteristics.

4.3 Examination. The insert cabinets shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Chemical analysis of the metals. The samples selected in accordance with 4.2.3. shall be analyzed in conformance with methods 111 or 112 of Federal Test Method Standard No. 151. In case of dispute, method 111 shall be used for referee analysis.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

4.5 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C, as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit packaging. Each cabinet insert shall be packaged in accordance with MIL-P-116, method III.

5.1.1.2 Level C. Packaging for each cabinet insert shall conform to the requirements of MIL-STD-794 for level C.

5.2 Packing. Packing shall be level A, B, or C as specified (see 6.2).

5.2.1 Level A. Each cabinet insert shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type or PPP-B-621, class 2. Closure and strapping shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification.

5.2.2 Level B. Each cabinet insert shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type, PPP-B-621, class 1, or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

5.2.3 Level C. Each cabinet insert shall be packed in accordance with the requirements of MIL-STD-794 for level C.

5.2.4 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 40 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in height, 48 inches in width, and 40 inches in length. Pallet shall conform to Type IV or Type V.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

5.2.4.1 Levels A and B. Cabinet inserts packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets as specified in 5.2.5.

5.2.4.2 Level C. Cabinet inserts packed specifed in 5.2.3 shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Levels A, B and C. Each exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4. General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packing inspection. The inspection of these packaging requirements shall be in accordance with 4.5.

6. NOTES

"6.1 The medical field chest specified in 1.1 is covered by NSN 6545-00-914-3510, Chest, Medical Instrument and Supply Set, No. 6. Copies of Defense Medical Purchase Description covering this item may be obtained upon application to the procuring activity or as directed by the contracting officer."

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2)

6.3 Availability of color standards. The color standards of the color 34087 referenced in 3.1.7 are available from:

Commanding General
Aberdeen Proving Ground, Maryland
ATTN: Coating and Chemical Laboratory

6.4 Availability of color chips. The individual 3 by 5 inch chips of the color 37038 referenced in 3.1.7 are available from the General Services Administration Business Service Center, Region 3, Seventh and D Streets, S.W. Washington, D.C., 20406.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

6.5 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review Activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
DoD-MB

Agent:
DLA-DM

Project No. 6545-0059

Location: NCR DISK #6

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERM [illegible] NO. 4986 WASHINGTON D.C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
MIL-I-36254A(DM)	INSERT CABINET, MEDICAL INSTRUMENT AND SUPPLY SET CHEST

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)* _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.