MIL-I-36279A
7 March 1966

Revision of
MIL-I-36279 (DM)
dated 2 September 1964

MILITARY SPECIFICATION

INCUBATOR, INFANT, Portable

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers portable infant incubators mounted on stands and equipped with swivel casters.

1.2 Classification. The portable infant incubators covered by this specification shall be the following type:

Type I. Incubators, Infant, Portable, Explosion-proof, 110 volts, 60 cycles, AC.

Type II. Incubators, Infant, Portable, 110 volt, 60 cycles, AC.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

FF-C-77	Casters, Institutional Duty.
PPP-B-585	Boxes; Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Boxes, Fiberboard.
TT-E-491	Enamel Gloss, Synthetic (For Metal and Wood Hospital Furniture)

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap, Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Federal

FED-STD-595	Colors

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publication. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

American Standards Association, Inc. (ASA)
ASA C73.11-1963 Caps and Receptacles, 125 Volts, 15 Amperes, 2 Pole, 3 Wire, Grounding Type, General Purpose.

Applications for copies should be addressed to
American Standards Association, Inc.
10 East 40th Street, New York New York. 10016

National Fire Protection Association (NFPA)
NFPA No. 56 Code for the use of FLAMMABLE ANESTHETICS, 1965

Applications for copies should be addressed to
National Fire Protection Association
60 Batterymarch Street, Boston, Mass: 62110.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Preproduction sample. Unless otherwise specified (see 6.1), prior to the start of production of the contract quantity, one preproduction sample of the incubator shall be assembled, completely finished, and tested by the contractor. The sample shall be submitted, together with the contractor's inspection and test reports, to the contracting officer for examination and testing to determine compliance with the requirements of this specification. Written approval of the materials, construction, and performance of the sample shall be obtained from the contracting officer, or his authorized representative, prior to the manufacture and assembly of the remaining quantity to be manufactured. Further production of the item by the contractor, prior to this approval, shall be at the contractor's risk. Accepted preproduction sample shall become the property of the procuring activity and shall be included in the quantity called for in the contract or order. In the event that the preproduction sample is not approved. The contractor shall submit a new sample for approval, or modify the existent sample so as to remove the deficiencies cited.

3.2 Fire and casualty hazards.

3.2.1 The contractor shall submit to the contracting agency proof, prior to delivery, that the equipment he will supply under this specification conforms to the fire and casualty hazards provisions of the applicable standards for operation in hazardous locations, Class I, Group C and as prescribed by NFPA No. 56. Such proof shall consist of a laboratory report or listing by the NFPA Laboratories, Inc. The label or listing of the Underwriters' Laboratories shall be acceptable as evidence that the equipment conforms to the above. A copy of each laboratory report shall be furnished to the contracting officer.

3.2.2 In lieu of the label or listing by the Underwriters' Laboratories Inc., the contractor may submit a laboratory report by a nationally recognized testing agency adequately equipped and competent to perform such services, as shall be evidenced by the attachment of its seal, stamp or label to the article, or the issuance of a certificate. The contractor shall submit the laboratory report which shall include test results and methods used to determine compliance to the contracting officer for approval. The Government reserves the right to require a facilities report showing the qualifications of the testing laboratory and to conduct a facilities survey to determine its acceptability.

3.2.3 Compliance with the above requirements regarding fire and casualty hazards, does not absolve the contractor from complete compliance with the requirements of this specification in order to secure the acceptance of this commodity.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

3.3 Materials.

3.3.1 Sheet steel. Sheet steel shall be electro-galvanized and bonderized commercial grade cold rolled steel not less than 0.035 inch (No. 20 United States Standard Revised Gage)thick.

3.3.2 Enamel. The type II incubator shall be coated with enamel conforming to TT-E-491. The type I incubator shall be coated with electrically conductive enamel.

3.3.3 Casters. Caster shall conform to type II, class F, size 3, wheel style 4, mounting S of FF-C-77, except rubber expansion socket adapters shall not be used. The mounting stem shall be not less than 7/16 inch in diameter and 1-3/8 inches in length. The stem shall be equipped with a friction ring located at the top, to permit easy insertion and removal of caster. Friction shall be adequate to prevent caster from falling out. Two of the casters shall have foot-controlled wheel brakes. The type I incubator shall have electrically conductive casters (see 3.10). The requirement for conductivity shall not apply to the type II incubator.

3.3.4 Chain. Shall be nickel-plated brass, tangle free, bead chain. The chain shall withstand 185 pounds straight pull without breaking, when tested as specified in 4.4.5.

3.3.5 Glass. Shall be 3-ply, 1/4 inch thick shatterproof safety glass.

3.4. Design. Incubator shall be designed to supply constant, automatically-controlled heat and increased humidity, for both premature and full term babies. It shall be portable and be provided with means for the administration of oxygen. The unit shall consist of a portable incubator mounted on a sturdy metal stand equipped with four swivel casters.

3.5 Size. The incubator shall have the following minimum clear inside dimensions, exclusive of controls, heating unit, etc.: 23 inches long, 14-3/8 inches wide, and 11 inches deep. The outside dimensions shall be 31-1/2 inches long, 15 inches wide, and 13-3/4 inches high. Tolerance on all dimensions shall be ± 1/2 inch.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

3.6 Construction. The body of the incubator shall be constructed of sheet steel. The entire top shall be a hinged lid with a single sheet or shatterproof safety glass. The entire front of the infant compartment shall also be of safety glass, so that the infant is in full view at all times. The top lid shall be equipped with a safety catch to hold it securely in the open position, or a bead chain of such length as to permit the lid to open a sufficient amount past the vertical position, so that it will not be accidently closed by a breeze or a slight jar. Lid shall close against a suitable preformed gasket. Ends of cabinet shall have recessed handles for easy carrying. The inside of the incubator shall be smooth, with no hooks or projections that may injure the infant and make cleaning or sterilizing difficult. All outer surfaces shall be smooth with no sharp corners. Handles, inlet tubes, and controls shall be so located as to prevent catching of clothing of passersby. The infant area shall be separated from the heater-humidifier compartment by suitable protective baffles, which are movable to permit cleaning of that compartment.

3.7 Heat source.

3.7.1 Type I heat source. Heat shall be supplied in the type I incubator by radiation from a heating element sealed into a seamless, metal-alloy tube, filled with magnesium-oxide insulation, further sealed in a protective metal tube. The entire operating mechanism shall be contained in a precision machined, heavy walled, explosionproof, cast housing. Protective screens and metal shieldings shall prevent the baby and attendant from coming into contact with the heat source.

3.7.2 Type II heat source. Heat shall be supplied in the type II incubator by radiation from one 150 watt, standard, incandescent, medium socket, electric lamp, enclosed in a gas-tight metal tube. This heating tube and complete operating mechanism shall be sealed in one gas-tight housing which shall provide a safe heat source for oxygen. Protective screens and metal shielding shall protect the infant or attendant from coming in contact with the heating element.

3.8 Humidity. Humidity shall be provided by a completely screened and protected nonferrous water tank which shall be readily removable for filling and cleaning. The size of the waterpan shall be such that a minimum relative humidity of 65 percent will prevail so long as the heat is on in the incubator and the waterpan is filled. The unit shall be tested as specified in 4.4.6.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

3.9 Static charge conductivity. The type I incubator shall be tested for conductivity as specified in 4.4.1. The static charges shall be transmitted to the grounded metal. There shall be no accumulation of charge on the enamel coated surface of the incubator. Small charges which are distributed over the surface of the glass without local accumulation, and which are gradually leaked to the grounded metal shall not be cause for rejection.

3.10 Electrical conductivity. The incubator shall meet the requirements for Reduction of Electrostatic Hazard as specified in section 253 of NFPA No. 56, when tested as specified in 4.4.2.

3.11 Temperature. The temperature in the incubator shall be variable from room temperature to not less than 95 degrees Fahrenheit (F.). When tested as specified in 4.4.3, and 4.4.6 the temperature of the infant compartment shall not exceed the given maximum temperatures for the following operating conditions:

Incubator with water pan (dry)	.90°F. max.
With water pan (wet)	.95°F. max.
Without water pan and thermostat out of circuit	.118°F. max.

3.12 Dielectric strength. The type I incubator shall be tested as specified in 4.4.4 and shall show no signs of electrical breakdown.

3.13 Control switch. A control dial, directly connected to an automatic thermostatic switch, shall be located on the front of the incubator. Dial shall be marked with control numbers. Thermostatic switch shall automatically turn the heat off and on, in order to maintain the desired temperature and humidity within the incubator. The switch shall be sensitive to temperature changes of plus or minus 1°F. in the incubator, when the top lid is closed. The front switch panel shall have a red glass jewel light and a green glass jewel light. The green light shall show that the power is on and the incubator properly connected. The red light shall show when the heating element is in operation. The unit shall be tested as specified in 4.4.6. The type II incubator control dial shall be illuminated.

3.14 Oxygen inlet. An oxygen inlet nipple shall be set in the lower corner of the left hand end, and shall protrude not more than 3/8 inch beyond the surface of the cabinet side. On the inside of the incubator, covering the inside end of the inlet nipple, there shall be a diffusing shield that may be rotated to direct the flow of oxygen into the incubator. Diffusing shield shall be easily removable so that a tube and mask may be attached to the inside end of the inlet nipple.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

3.14.1 Oxygen limiting valve. A valve shall be provided to control the oxygen flow into the incubator. The valve shall specify oxygen concentration by numerical identification (40% and 100%.) and by color code; the 40% limiting position clearly colored a bright green and 100 % flow position colored a bright red.

3.15 Ventilation. Incubator shall be fitted with two ventilators; one at the top front of the cabinet and one at the bottom under the heating unit. The top ventilator shall be controlled by a knurled thumb nut which shall permit it to be locked open, closed or any position between. The bottom ventilator shall have a false or pilot ventilation indicator located at the bottom front of the cabinet. This pilot ventilator shall control and visibly indicate opening of the bottom ventilator. There shall be sufficient ventilation even when the ventilators are closed. Control of the vents shall be from the front of the cabinet.

3.16 Thermometer assembly. A thermometer assembly consisting of a thermometer, thermometer holder, and protective armor shall be securely attached to the inside of the incubator and shall rest against the front glass, away from the baby, so as to be readable from the front outside of the incubator. The thermometer shall be made of glass, with a backing of color contrasting to that of the fluid for easier reading. Calibrations shall be etched in both Centigrade and Fahrenheit scales. The Centigrade scale shall be graduated at one degree intervals and the range shall be from 5 to 40 degrees. The Fahrenheit scale shall be graduated at two degree intervals and the range shall be from 40 to 110 degrees. The accuracy of the thermometer shall be within 1/2 degree, for both scales. The holder and armor shall be fabricated of chrome-plated brass. Two electrically conductive soft rubber grommets shall cushion the thermometer in the holder. The thermometer shall rotate within the thermometer assembly and shall be easily disassembled for cleaning.

3.17 Stand. The incubator shall be mounted on a welded steel stand designed to fit and hold it conveniently. The incubator shall rest securely on the top shelf, and four rounded uprights extending above the top shall hold it securely in place without additional fastenings. It shall be possible to lift the incubator readily off the stand for portable use. The top of the incubator shall be located 40 $\pm$ 1 inches above the floor, excluding the handle. A utility shelf shall be located 17 $\pm$ 1/2 inches below the top shelf in the type I incubator and 14 $\pm$ 1/2 inches below the top in the type II incubator. Each leg shall be provided with a caster mounting and a caster. The two casters at the front of the incubator shall have foot-controlled brake locks.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the current version before use.

3.18 Electric cord and plug. The power supply cord shall be 3-wire having a minimum length of 15 feet for the type I incubator and a minimum of 6 feet for the type II incubator. The free end of the type I incubator cord shall be provided with a tag containing the appropriate information that only an explosion proof male plug shall be attached (of a type to mate with the receptacles available at the using installation) and indicating color codes of the individual conductors. The cord on the type II incubator shall be furnished with a plug for use with receptacles specified in ASA C73.11-1963 for Caps and Receptacles, 125 volts, 15 amperes, 2-pole, 3-wire, Grounding Type.

3.19 Finish.

3.19.1 All metal parts of the type I incubator, except hardware, shall be finished with electrically conductive enamel in accordance with paint manufacturer's instructions. The color of the enamel shall be silver lustre. All hardware shall be nickel or chromium plated. The casters shall be zinc or cadmium plated.

3.19.2 All metal parts of the type II incubator shall be finished with two coats of enamel, baked in between each coat, in accordance with manufacturer's instruction. The color of the enamel shall be lustreless green conforming to color 14241 of FED-STD-595. The casters and hardware shall be finished the same as for the type I incubator.

3.20 Instructions. A metal plate with instructions covering the operation and control of ventilators, temperature, humidity, oxygen and transportation, shall be securely affixed on the outer surface of the incubator. The type II incubator decal may be fabricated from vinyl plastic as an alternate to the metal plate.

3.21 Safety approval. Unit shall be one which has proven satisfactory in actual use in hospitals over a period of not less than 2 years.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the currentversion before use.

3.22 Identification marking. Each incubator shall have one or more metal nameplates bearing the following information:

Name or registered trademark of the manufacturer.
Contract or purchase order number.
Electrical ratings; (volts, amperes, watts and frequency).
Manufacturer's type and serial number.
Federal stock number.
Letters U.S.

The marking shall be clear, legible, and permanent. Nameplates secured by adhesives shall not be used.

3.22.1 Marking for explosionproof characteristics. As a visual indication of its explosionproof characteristics, a bright, red-colored nameplate shall be located on the type I incubator, just above the bottom ventilator control. In addition, a bright red line, 9 inches long by 5/8 inches wide, shall be painted directly under the top ventilator.

3.23 Service data. The contractor shall furnish, with each unit, two copies of service data containing complete instructions for installation, operation, maintenance, and a list of component parts. This information shall specifically include instructions and drawings. The service data shall include such detailed information which shall permit ready installation and maintenance of the item. All information contained in the manual shall be suitably indexed to permit ready reference to the required information. Wiring diagrams shall be included for electrical items or components. Wiring diagrams, drawings, and descriptive material shall include adequate information so that the replacement parts can be ordered by the manufacturer's part number, and without confusion as to the item identification. Identical or interchangeable replacement parts shall be identified by a single part number only. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the contracting officer, via the cognizant Government inspector two preliminary copies of his proposed service data, for review and comment (see 6.1).

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the currentversion before use.

The preliminary service data shall be submitted sufficiently early to permit adequate review (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of required revisions and corrections in the final service data without delaying delivery under the contract or purchase order. One set of the service data, as indicated above, shall be placed in a sealed, waterproof envelope, and affixed securely to the outside of the exterior container. The envelope shall be secured to the exterior container by use of metal ties, nails, or similar means, and shall be located at the end of the exterior container, preferably, when space permits, at the end bearing the markings required by MIL-STD-129. In cases where complete service data is too bulky to be conveniently attached to the outside of the exterior container, a note shall be made on the exterior container which shall indicate where the remainder of the first set of service data will be located.

When two or more containers are required for packing a disassembled item (components of a single stock number), set markings shall be applied in accordance with MIL-STD-129. The envelope shall be affixed to the container designated as No. 1 of the set. The second set and, if necessary, the remainder of the first set of service data, shall be attached to the unit inside the exterior container. In addition, the contractor shall mail 15 copies of the approved service data to the contracting officer for distribution.

If the supplier has furnished acceptable manuals for the identical item within two years preceding the date of contract, and proposes to furnish manuals identical with those previously accepted, the requirements for review of drafts and submission of 15 additional copies may be waived upon request.

3.24 Workmanship. Workmanship shall be first class throughout. Incubator and stand shall be free from defects which detract from their appearance or may impair their serviceability. Any portion of incubator which comes in contact with oxygen shall not be contaminated with grease or oil.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the current version before use.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory, investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	1.0
Total (major and minor combined) defects	II	4.0
For dimensional examination	S-3	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105.

Characteristic	Requirement	Test procedure	Inspection level
Static charge conductivity	3.9	4.4.1	S-4
Electrical conductivity	3.10	4.4.2	S-4
Temperature	3.11	4.4.3	Only one unit.
Dielectric strength	3.12	4.4.4	100%
Chain strength	3.3.4	4.4.5	Only one unit.
Operation Tests:			
Relative humidity	3.8	4.4.6	S-4
Accuracy of control thermostat	3.13	4.4.6	S-4

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the currentversion before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Static charge conductivity. The glass and painted metal surfaces of the assembled type I incubator shall be rubbed with cotton sheeting and observed for static charges by the use of electroscope.

4.4.2 Electrical conductivity. The assembled type I incubator, complete with casters shall be tested using the equipment described in section 2522 (C) of NFPA No. 56.

4.4.3 Temperature. The assembled incubator shall be operated at rated voltage and with normal thermostat setting (90°F.), (1) with water pan in place but dry, (2) with water pan in place with water in it, and (3) without water pan and thermostat out of circuit. The infant compartment shall be loosely packed with clothes to a depth of approximately 5 inches. The temperature shall be measured using three No. 28 Awg iron constantan thermocouples placed at the head, body center, and foot portions of the compartment. The recorded temperature shall comply with 3.11.

4.4.4 Dielectric strength. While in a heated condition, the type I incubator shall be subjected to a gradually increasing potential of twice the rated voltage plus 1000 volts, or 1230 volts, which shall be maintained at this value for a period of one minute.

4.4.5 Chain strength. Strength of chain shall be determined by subjecting not less than 7-inch length of chain to the straight pull specified in 3.3.4.

4.4.6 Operations test. Each unit shall be operated in accordance with the manufacturer's instructions. The unit shall be operated at 95°F. (both vents closed) and relative humidity and temperature variation shall be observed and recorded when the unit reaches an equilibrium state.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One incubator, as specified, constitutes one unit of issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

5.1.2 Level A

5.1.2.1 Unit package

5.1.2.1.1 Incubator. Each incubator shall be disassembled as far as practicable and packaged in a carton of appropriate size and design constructed in accordance with PPP-B-636. All parts of the incubator shall be secured to prevent shifting or breakage. In addition, all exterior finished surfaces shall be protected by wrapping with kraft paper and cushioning with excelsior pads or flexible corrugated fiberboard or other suitable materials. Wrapping and cushioning material shall be secured by tying with cotton cord or tape, or by sealing with kraft gummed paper tape. Work shall be accomplished in such a manner as to insure against marring, scratching, or damage to the unit. Closure of carton shall be adequate to prevent spelling of contents under normal handling.

5.1.2.1.2 Stand. Each stand shall be packaged in a carton, of appropriate size and design, constructed in accordance with Specification PPP-B-636. Carton shall be designed with suitable interior fittings, or shall contain sufficient cushioning material to insure against shifting of, or damage to, the contents. Each carton shall be adequately secured to prevent accidental opening.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4, of PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic and method I closure shall be utilized.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only whe applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. The legend "OPEN THIS SIDE" shall be shown where applicable, to facilitate locating service data, inspecting, and removing the unit.

6. NOTES

6.1 Ordering data. Procurements documents should specify the following:

(a) Title and date of this specification.
(b) Type required (see 2.1).
(c) When preproduction sample is not required.
(d) Where to send preproduction sample or service data.
(e) Selection of applicable levels of packing (see 5.1 and 5.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic and method I closure shall be utilized.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When lablels are utilized, waterproofing shall be required only whe applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. The legend "OPEN THIS SIDE" shall be shown where applicable, to facilitate locating service data, inspecting, and removing the unit.

6. NOTES

6.1 Ordering data. Procurements documents should specify the following:

(a) Title and date of this specification.
(b) Type required (see 2.1).
(c) When preproduction sample is not required.
(d) Where to send preproduction sample or service data.
(e) Selection of applicable levels of packing (see 5.1 and 5.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

6.2 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

CATEGORY - N.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the currentversion before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity (as indicated on reverse hereof).

SPECIFICATION

ORGANIZATION *(Of submitter)*

CITY AND STATE

CONTRACT NO.

QUANTITY OF ITEMS PROCURED

DOLLAR AMOUNT
$

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES.

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO IF "YES", IN WHAT WAY?

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity)*

DATE

DD FORM 1426

SND PPSO 13503

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

FOLD UNDER

POSTAGE FEES PAID
DEFENSE SUPPLY AGENCY

DEFENSE SUPPLY AGENCY
DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20th ST
PHILADELPHIA, PA. 19101

OFFICIAL BUSINESS

DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER

ATTN: DIRECTORATE MEDICAL MATERIEL

2800 SOUTH 20TH STREET

PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the current version before use.