MIL-M-36278B
12 March 1969
SUPERSEDING
MIL-M-36278A
19 October 1967, and
MIL-M-36278(DM)
4 September 1964
NAUTRADEVCEN
3143-265A
14 January 1957

MILITARY SPECIFICATION

MOULAGE SET, WAR WOUNDS, SET OF 19

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers a set of wound and other anatomical training moulages, suitable for the demonstration of procedures and techniques encountered in administering first aid.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein.

SPECIFICATIONS

Federal

PPP-B-640	Boxes, Fiberboard, Corrugated, Triple, Wall.

STANDARDS

Federal

FED. STD. 595	Colors.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

FSC 6910

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Materials.

3.1.1 Moulages, straps and reservoir ports. Shall be made from a vinyl plastisol, or equal, and shall be formulated to produce a finished product which is resilient, has good dimensional stability, retains its finish without chipping, and is capable of withstanding repeated use.

3.1.2 Adaptors and tubing. Adaptors shall be injection-molded vinyl and tubing shall be a commercial quality vinyl.

3.1.3 Reservoirs and blood powder bags. The reservoir bodies and blood powder bags shall be made of vinyl sheet. Vinyl shall be 12 mil for the reservoir bodies and 8 mil for the blood powder bags.

3.1.4 Pumps. The pumps shall be made of new rubber. No reprocessed or reclaimed rubber shall be used.

3.1.5 Carrying case. The top, bottom and sides of the carrying case shall be a minimum of 3/8 inch clear beech, birch, or maple lumber, dressed both-sides (D2s), kiln dried, with a moisture content not to exceed 12 percent. Front and back panels of carrying case shall be 3-ply D2s, 1/8 inch, birch or maple plywood. The plywood shall be thoroughly kiln dried. Wood shall be free of cracks and knotholes, and any knots present shall be sound.

3.1.5.1 Corner guards and hinge. Corner guards and hinge for the carrying case shall be made of brass, or brass plated steel. The hinge shall be piano type

3.1.5.2 Safety hasp. Safety hasps shall be of brass or steel, and nickel or chrome plated. Safety hasps shall be 3 1/4 inches long by 1 1/2 inches wide.

3.1.5.3 Handle and bumpers. Handle shall be made of leather and bumpers of rubber.

3.2 Design and construction. The moulage set, including all components, shall be produced in accordance with Government furnished sample and the requirements of this specification. Except as specified herein, all materials, details, colors and dimensions shall match those found in the Government furnished prototype.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

3.2.1 Moulages. Moulages shall be of the bleeding and non-bleeding types. Each moulage shall be a life-size model of the wound or affected area of the anatomy. Material thickness at any point of the moulage shall be 1/16 inch + 1/16 - 1/32 inch, except for wound areas, where it may be less. The blood vessels shall be permanently and securely attached along their entire lengths to the underside of the moulages. Moulages shall show no evidence of damage to the structure or finish, when tested as specified in 4.4.1.

3.2.1.1 Bleeding. Each bleeding moulage shall be capable of bleeding by means of simulated blood pumped through simulated blood vessels, located in the moulages. The following shall be bleeding moulages:

(a) Amputation.
(b) Compound fracture of the femur.
(c) Compound fracture of the humerus.
(d) Compound fracture of the lower leg.
(e) Gunshot wound of the palm.
(f) Laceration of the scalp.
(g) Shrapnel wound of the abdominal wall with protruding intestine.
(h) Shrapnel wound of the lower jaw including partial loss of the jaw.

3.2.1.2 Nonbleeding. The following shall be nonbleeding moulages:

(a) Atomic burn back.
(b) Atomic burn chest.
(c) Atmoic burn face.
(d) Atmoic burn hand (palm and dorsal area).
(e) Face in shock.
(f) Frostbite of the foot.
(g) Phosphorus burn of the hand.
(h) Second- and third-degree burns of the forearm.
(i) Trench foot.

3.2.1.3 Hypodermic needle technique moulage. A moulage shall be furnished for hypodermic needle insertion.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the current version before use.

3.2.1.4 Sucking wound of the chest. Shall be a blood-suffused moulage which, when air is pumped through the sponge in the bullet hole, produces a frothing action, and when the bulb is released, produces a sucking sound.

3.2.2 Pumps. Pumps including adaptors shall be provided for the bleeding moulages and sucking wound of the chest. Pumps shall be squeeze-type rubber bulbs with a minimum capacity of 2 ounces. The pumps shall be capable of manual operation by the wearer of the moulage. The pumps shall force simulated blood through the blood vessels of the bleeding moulages, and shall force the contents of the bulb (air and simulated blood) through the sponge in the sucking wound of the chest. The bleeding bulb adaptor shall have two orifices, one marked "M" and the other marked "R". These adaptors shall have valves built into them which will perform the appropriate function as described in the direction for use. The sucking wound of the chest bulb adapter shall have a single outlet, and be so designed that a plastic tube will connect the wound to the pump.

3.2.3 Reservoirs. Each bleeding moulage shall be provided with a blood reservoir having a minimum capacity of 16 fluid ounces of simulated blood. Reservoir shall have a plastic loop which can be used to secure, by means of a strap or belt, the reservoir to the wearer.

3.2.4 Tubing. Each set shall have the following vinyl tubing:

Quantity	Minimum overall length with connectors (inches)
2	65
2	50
15	40

3.2.4.1 Connectors. Each length of vinyl tubing shall have an aluminum connector permanently attached to each end as shown on the standard sample.

3.2.5 Simulated blood powder. Each set shall contain five heat-sealed vinyl bags, each containing 37 grams plus or minus 10 percent of the following mixture:

	Percent
Food and Drug Red No. 15	27.0
Food and Drug Red No. 2	27.0
Powdered Caramel Coloring	2.7
Methyl Cellulose (1500 Centipoises)	32.5
Sodium Benzoate	10.8

The above formula shall be uniformly and intimately mixed.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the current version before use.

3.2.6 Adjustable straps. Each moulage except those of the hand and foot, shall be provided with an adjustable strap. One end of each strap shall be permanently attached to the moulage. A sufficient number of straps shall be provided to effect a good fit by means of fasteners. The straps shall be adjustable so as to fit a male, standard sizes from 38 to 42. Each strap shall have non-corrosive metal fasteners.

3.2.7 Carrying case. The carrying case for each complete set shall be adequately compartmented for each moulage, for the pumps, reservoirs, tubes, and blood powder. Each moulage compartment shall be provided with vinyl plastic straps and snap fasteners for securing the moulages therein. Carrying case may be smaller in size than the prototype, providing all other conditions specified in this specification are met. The frame shall be of a lock-corner construction, and glued. Wire brads may be used along with the glue. The separators shall be of rabbeted or dado construction, and glued. Wire brads may be used along with the glue. The leather handle shall be secured to the case with appropriate hardware using roundhead bolts, and nuts. Corner guards shall be nailed or screwed on all corners, when the case is closed. There shall be two nails or screws on side surfaces contacted by the corner guard. The piano hinge shall be attached to the carrying case by means of flathead screws. Hinge shall have holes spaced on 2-inch centers. Holes shall be countersunk so that the head of screws, when inserted, will be flush with the hinge. Safety hasps shall be secured with roundhead screws. Front and back panels shall be glued. Wire brads may be used along with the glue. Six rubber bumpers shall be placed at the bottom of the case, three on each half of the case. All wood shall be sanded smooth.

3.2.7.1 Instruction sheets. The contractor shall furnish two of the following instruction sheets in each case. Instructions shall be laminated between two sheets of clear plastic.

INSTRUCTIONS FOR THE CARE AND USE OF MOULAGE SET, WAR WOUNDS, FSN 6910-540-6378

CONTENTS OF KIT

One (1) set of 19 moulages appropriately labeled and compartmented.
Eight (8) sets of rubber pump assemblies (pump, reservoir, 40-inch tube) for use with the bleeding moulages.
One (1) pump for use with sucking wound of chest.
Five (5) packets of powdered, simulated blood, each packet sufficient to make one gallon.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:19Z
Check the source to verify that this is the current version before use.

PREPARATION OF SIMULATED BLOOD

1. Empty the contents of a packet into a large cup.
2. Add cold water and stir into a smooth paste.
3. Gradually add more cold water with stirring until the cup is almost full.
4. Empty the cup into a gallon of cold water, and shake.

FORMULA FOR SIMULATED BLOOD

F & D Red #15	10 grams
F & D Red # 2	10 grams
Powdered Caramel Coloring	1 gram
Methyl Cellulose (1500 Centipoises)	12 grams
Sodium Benzoate	4 grams

NOTE: Simulated blood can be removed from most clothing by good laundering but may leave permanent stains on woolen clothing.

ASSEMBLY AND OPERATION

1. Dip all bloodline connectors in water before using. Fill plastic reservoir through 3/4-inch filling port. Cap and squeeze reservoir until all air is exhausted through bloodline port. Insert bloodline and connect to pump port "R". Connect pump port "M" to moulage.

2. Squeeze bulb several times until blood fills bulb and bloodline to moulage. Gentle, intermittent squeezing of the bulb will simulate arterial bleeding.

3. The sucking wound chest moulage does not have a reservoir. It is connected directly to the single outlet bulb which has been initially filled with sufficient blood to moisten the wound. The bulb is then operated in a position to pump a mixture of blood and air creating a sucking and frothing effect.

HANDLING AND CLEANING

1. Where necessary, roll, DO NOT FOLD, for storing.
2. Wash with a mild soap and warm water; do not scrub or use detergents or abrasive cleaners.
3. Clean reservoirs, bloodlines, and moulage capillary lines IMMEDIATELY AFTER USING. CAUTION: Do not use boiling water under any circumstance.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

NOTE: The contents of this set are designed to perform under ambient conditions of temperature ranging for 20° F. to 130° F., and relative humidity ranging up to 95 percent.

(See other side for replacement items).

The reverse side of the instruction sheet shall have the following:

REPLACEMENT PARTS

FSN 6910-729-6161	SIMULATED BLOOD, POWDER, 37 Gm, 5s
FSN 6910-729-6186	BAG, PLASTIC, SIMULATED BLOOD, 16 oz, 8s
FSN 6910-729-6187	PLASTIC TUBE ASSEMBLY, WAR WOUNDS MOULAGE SET
FSN 6910-853-5803	TUBE, SUCTION AND PRESSURE, WAR WOUNDS MOULAGE SET
FSN 6910-892-1942	MOULAGE, UPPER ARM SKIN, HYPODERMIC INSERTION, Adult

Manufacturer's name or registered trademark
Contract or purchase order number
Date of manufacture.

3.3 Finish.

3.3.1 Moulages. The moulages shall be finished with a matte coating which will best simulate human skin, and which will form a durable bond with the material used in making the moulages. All colors used shall be those which accurately simulate the color of the particular skin, flesh, bone and tissue of the particular area of the human anatomy. Moulages in the Government-loaned sample moulage set shall be used as a guide for matching colors. The finish shall be durable and shall not chip, soften, crack, become sticky or tacky, when tested as specified in 4.4.2.

3.3.2 Carrying case. The color of the carrying case shall be No. X24087 of FED. STD. 595 and shall be applied as semigloss enamel to the outside only.

3.4 Identification marking.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:19Z
Check the source to verify that this is the current version before use.

3.4.1 Moulages. All moulages shall be labeled as to type of wound or affected anatomical part of the body. Label shall be either of durable material firmly attached to the underside of the moulage or strap, or "hot-stamped" on the strap. As an alternate, labeling shall be applied to the underside of moulage or strap in an ink which is indelible when applied to vinyl plastic.

3.4.2 Carrying case. In the center of both side panels, the carrying case shall be marked with the letters U. S. a minimum of 2 inches in height. Centered below the letters U. S. shall appear the Federal stock number and below that, the item identification in block letters not less than 1 inch high. Numerals and letters shall be yellow or white. Compartments shall be stamped with the name of the moulages they contain.

3.5 Government-loaned property. The following property shall be loaned by the Government:

(a) Moulage molds (one set)

(b) War Wound Moulage Set (set of 19) with carrying case.

NOTE: The contractor shall be responsible for the care, maintenance and return of all Government property in the same condition as when originally supplied by the procuring activity. Return of all Government-loaned property shall be at the expense of the Contractor.

3.6 Workmanship. The sets, including all moulages, pumps, adaptors, reservoirs, straps, tubes, and carrying case shall be first class throughout. Particular attention shall be given to detail in finishing and coloring the moulage surfaces. For the moulages the following shall apply:

(a) Surfaces shall be smoothly finished. Any pits and blisters present on outer surfaces, thickenings and carbonized vinyl present in or on the surfaces of the moulage shall be no greater than those present in the Government furnished prototype.

(b) Moulages shall be free of holes and cuts. Shapes shall be clean without raggededges attached thereto. Identification markings shall be uniformly dark and completely legible.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the current version before use.

(c) The top and bottom of the closed case shall be in complete alignment. Hinges and corner guards shall be securely fastened to the case. Corner guard nails and other sharp metal points shall not project above the inner surface."

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibilities for inspection. The supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own or any other inspection facilities and services acceptable to the Government. Inspection records of the examination and tests shall be kept complete and available to the Government as specified in the contract or purchase order. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

	Inspection level	AQL(defects per hundred units)
For visual examination		
Major A defects	II	1.0
Major B defects	S-2	1.0
Minor defects	II	2.5
For dimensional examination	S-1	1.0

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

(c) The top and bottom of the closed case shall be in complete alignment. Hinges and corner guards shall be securely fastened to the case. Corner guard nails and other sharp metal points shall not project above the inner surface."

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibilities for inspection. The supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own or any other inspection facilities and services acceptable to the Government. Inspection records of the examination and tests shall be kept complete and available to the Government as specified in the contract or purchase order. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

	Inspection level	AQL(defects per hundred units)
For visual examination		
Major A defects	II	1.0
Major B defects	S-2	1.0
Minor defects	II	2.5
For dimensional examination	S-1	1.0

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:19Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105.

Component (unit of product)	Characteristic	Requirement	Test procedure	Inspection level
Single moulage	Drop test	3.2.1	4.4.1	S-1*
Single moulage	Resistance to temperature and humidity	3.3.1	4.4.2	S-1*
Package of Simulated blood powder	Solubility	3.2.5	4.4.3	S-1

*Not less than 3 individual moulages shall be chosen at random from each anticipated delivery prior to assembly in sets. Acceptance number shall be zero.

4.3 Examination. The moulage sets shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects:

Categories and defects*

Major A

101 Set not complete with all components, assemblies, and subassemblies.
102 Moulages not free of holes, tears, blisters, burned areas, and voids.
103 Paint finish not uniformly and evenly applied to moulage.
104 Moulage not free of peeled, cracked, and chipped areas.
105 Simulated blood vessels and manifolds not permanently attached along their entire lengths to the bleeding moulages.
106 One end of securing straps not permanently attached to the moulages.
107 Aluminum connectors not free of sharp edges and burs.
108 Aluminum connectors cannot be readily inserted into bulb adapters and manifolds.
109 Aluminum connectors not a firm fit in bulb adapters and manifolds.
110 Aluminum connectors not permanently attached to tubing.
111 Powder bags not free of leaks.
112 Moulages and case compartment not correctly, permanently, or legibly identified.
113 Halves of carrying case not properly aligned.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

Major E
151 Moulage finish does not tolerate warm mild soap solution.
152 Adaptor valves do not function properly.
153 Reservoirs not free of leaks.
154 Moulage securing devices in carrying case do not function properly.

Minor
201 Tubing not of quantity specified.
202 Instruction sheet not complete, not legible, not furnished in duplicate, or not properly laminated between two sheets of clear plastic.
203 Carrying case not marked correctly, legibly, or permanently.
204 Wood not smooth.
205 Case hardware not corrosion resistant.
206 Case finish not a continuous, smooth, and uniform film.
207 Case finish not free of peel, runs, wrinkles, and drops.
208 Moulages not properly secured in appropriate plastic bag.

*Examination shall not be restricted to the classified possible defects listed above.

4.3.2 Dimensional examination. The moulage sets shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Drop test. Each moulage, including pump, full reservoir, and tubing, shall be dropped from a height of 48 inches onto a concrete floor.

4.4.2 Resistance to temperature and humidity.

4.4.2.1 Resistance to cold. Moulage shall be subjected to a maximum temperature of 20° F for a period of not less than 2 hours.

4.4.2.2 Resistance to heat and humidity. Moulage shall be subjected to a temperature of 130° F. ± 5° F., and a minimum of 95 percent relative humidity for a period of not less than 4 hours.

4.4.3 Solubility of simulated blood powder. One gallon of simulated blood shall be prepared as instructed in 3.2.7.1. The resulting solution shall be free of foreign and undissolved matter which would clog the apparatus during use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Preservation and packaging. Preservation and packaging shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit packaging.

5.1.1.1.1 Moulages. Each moulage shall be evenly rolled and inserted in a bag fabricated from polyethylene, having an average wall thickness of not less than 0.003 inch. Closure shall be effected by means of heat sealing.

5.1.1.1.2 Tubing. Tubing shall be formed into a coil of such diameter that it will not kink or be permanently deformed.

5.1.1.1.3 Reservoirs. Each reservoir shall be rolled in such a manner as to prevent creasing or damage to the reservoir.

5.1.1.1.4 Assembly. Component parts shall be placed in their respective compartments in the carrying case. All work shall be accomplished in such a manner as to prevent shifting of, or damage to, the enclosed parts.

5.1.2 Level C. For level C, preservation and packaging shall conform to manufacturer's commercial practice.

5.2 Packing. Packing shall be level A or C as specified (see 6.2).

5.2.1 Level A. Each carrying case shall be wrapped in kraft paper or other suitable material. Wrapping shall be secured. Each wrapped case shall be overpacked in an exterior container conforming to the requirements of PPP-B-640, Class 2, Style E, regular slotted. Container design shall include suitable interior fittings to adequately protect all exterior hardware and carrying handles, and to prevent shifting or damage to the unit. Closure shall be in accordance with 30.3.5 of PPP-B-640 except that the use of steel strapping is not authorized. Sealing shall be in accordance with 30.2 by use of tape as specified in 30.1.6 of PPP-B-640.

5.2.2 Level C. Units packages as specified in 5.1.2 shall be packed in containers which will insure acceptance by common carrier and safe delivery at destination. Shipping containers shall conform to common carrier rules or other regulations as applicable to the mode of transportation and may be the supplier's commercial practice when such meets the requirements of this level.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

5.2.2 Physical protection. All material shall be packaged and packed in containers constructed to provide a compact and nonshifting load. Adequate cushioning, blocking and bracing shall be provided within shipping containers in a manner designed to prevent damage due to shock, vibration or impact during shipment, handling and storage.

5.3 Marking.

5.3.1 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Marking shall include arrows and the stenciled word UP as specified in the handling marking requirements of MIL-STD-129.

6. NOTES

6.1 Intended use. The war wound moulage set will be used to train military personnel in first aid treatment of combat-type casualties and to permit trainees to practice and develop first aid skills.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification

(b) Selection of applicable levels of preservation and packaging and packing (see 5.1 and 5.2).

6.3 This specification does not covers all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION: MIL-M-362788 FSN:

ORGANIZATION

CITY AND STATE | CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO (If "yes", in what way?)

4. REMARKS (Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)

SUBMITTED BY (Printed or typed name and activity - Optional) | DATE

DD FORM 1426 REPLACES EDITION OF [illegible], WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:19Z
Check the source to verify that this is the current version before use.