MILITARY SPECIFICATION

SLEEVE, PLASTIC, HISTOPATHOLOGICAL SPECIMEN

1. SCOPE

1.1 This specification covers a plastic sleeve, suitable for use in histopathological procedures in military medical facilities where permanent storage and intermittent shipment of fixed specimens is contemplated.

1.2 Classification. Plastic sleeves covered by the specification shall be of the following width, as specified (see 6.1):

Width 1. - 3 inches
Width 2. - 4 inches
Width 3. - 6 inches
Width 4. - 8 inches
Width 5. = 10 inches
Width 6. - 12 inches

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, forms a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

L-P-378 Plastic Film (Polyethylene Thic Gage).
PPP-B-585 Boxes, Wood, Wirebound.
PPP-B-601 Boxes, Wood, Cleated-Plywood.
PPP-B-621 Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636 Box, Fiberboard.
PPP-T-76 Tape, Pressure-Sensitive Adhesive Paper, (for Carton Sealing).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:15Z
Check the source to verify that this is the current version before use.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Basic material. The plastic sleeves shall be made from a continuous roll of natural virgin polyethylene sheet 0.005 gage ± 20 percent thick. The polyethylene sheet shall conform to type I, grade A, finish 1 of Specification L-P-378.

3.2 Construction. The polyethylene sheet shall be seamlessly formed into a flat tube and shall meet the specified width. A plus or minus 3/16 inch tolerance shall be permitted for all widths.

3.3 Workmanship. Workmanship shall be first class throughout. Plastic sleeves shall be free from defects which detract from their appearance or may impair their serviceability. The plastic sleeves shall be flexible, clean film and shall be entirely free of pinholes.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:15Z
Check the source to verify that this is the current version before use.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling. Sampling for examination shall be conducted in accordance with Standard MIL-STD-105.

4.3 Examination. The plastic sleeves shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) roll, of the stock number, specified, constitutes one unit of issue.

Table 1. Packaging quantities

Width	Unit Package	Intermediate package	Exterior container
1	1 unit	Not required	2 units
2	1 unit	Not required	2 units
3	1 unit	Not required	1 unit
4	1 unit	Not required	1 unit
5	1 unit	Not required	1 unit
6	1 unit	Not required	1 unit

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:15Z
Check the source to verify that this is the current version before use.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit (plastic sleeve) shall be evenly wound on a 3 inch (inside diameter) paper core. Each unit (roll) shall be further packaged by a complete wrap of kraft paper of suitable weight. Wrap shall be adequately secured to prevent spillage of contents under normal handling.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with Specification PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type I, class 2.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to Specification MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appexdix of Specification PPP-B-636. In addition, all seams, including manufacturer's joint, shall be closed by tape conforming to the requirements of Specification PPP-T-76. The tape shall be not less than 2 inches wide. Three strips shall be applied to both top and bottom so that all seams and joints shall be covered their full length. Tape applied over the lengthwise seams shall extend a minimum of 2 inches onto each of the end panels. Tape applied over the end edge seams shall extend a minimum of 2 inches along the side and top, and side and bottom panels.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:15Z
Check the source to verify that this is the current version before use.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to Specification PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with Specification PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-[illegible]36, type I, class 1.

5.3 Marking.

5.3.1 Unit package. Each unit package shall be marked in accordance with Standard MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with Standard MIL-STD-129. In addition, the letter "L" shall precede the Federal stock number. One space shall be allowed between the letter "L" and the first digit of the stock number. The terms "Stock number" or "FSN" shall not be indicated.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Width required (see 1.2).
(c) Selection of applicable levels of packing (see 5.2)

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Custodians:
Army. - MD
Navy. BuMed
Air Force. -AFMSG

Review activities:
Ar Army. - MD
Navy. -BuMed
Air Force. - AFMSG

Preparing activity:
Defense Supply Agency-DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:15Z
Check the source to verify that this is the current version before use.