MILITARY SPECIFICATION

SHELF ASSEMBLY, MEDICINE CART,
CORROSION-RESISTING STEEL

1. SCOPE

1.1 This specification covers a shelf assembly suitable for use with medicine cart (see 6.1), to increase the capacity from 24 to 40 medications.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

QQ-S-766	Steel Plates, Sheets, and Strip Corrosion Resisting.
PPP-B-585	Boxes; Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, Water-Resistant (for Carton Sealing).

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

STANDARDS

Federal

Fed. Test Method Std. No. 151 Metals, Test Methods.

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and and Storage.

DRAWING

Military Medical Supply Agency

20248 Shelf Assembly, Medicine Cart, Corrosion-Resisting Steel, 20-1/2 by 6-1/2 by 4-1/2 Inches.

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. Material used in the construction of the shelf assembly shall conform to the requirements for class 302 or 304 of QQ-S-766, and shall meet the chemical requirements specified therein when tested as specified in 4.4.1. Commercial tolerances on gage sizes shall be acceptable.

3.2 Construction. Construction of the shelf assembly shall be in accordance with Drawing 20248. The shelf assembly shall be a 2-step arrangement, each step having 8 circular cutouts. The risers in front of each step shall contain pairs of punched-out lugs suitable for holding securely, 1-3/4 - by 2-inch identification cards.

3.3 Finish. All normally visible surfaces of the shelf assembly shall have a No. 4 commercial finish. All other surfaces shall be cleaned and finished in accordance with the best commercial practice.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

3.4 Identification marking. Each shelf assembly shall be permanently marked with the following:

Name or registered trademark of the manufacturer
Stock number
Contract number
US

The marking shall be permanent, distinct, and legible. Marking shall be on a metal nameplate permanently attached or shall be die stamped. Nameplates secured by adhesives shall not be used.

3.5 Workmanship. Workmanship shall be first class throughout. Shelf assemblies shall be free from defects which detract from their appearance or affect their serviceability. All burs shall be removed and all sharp edges shall be broken.

3.5.1 Welding. Welds shall be durable and sound in every detail. Spotwelding shall be accomplished in such a manner as to assure thorough, well-fused welds of maximum strength. Prior to welding, all components shall be properly aligned in their proper positions to assure a straight, rigid, and level assembly. All visible welds shall be ground smooth so as to blend in with the adjacent surfaces. Discoloration of spotwelds shall be eliminated by grinding and polishing.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, stating the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

4.2.2 Raw materials. Four ounces of each lot of metal used in the construction of the shelf assembly shall be selected for chemical analysis.

4.3 Examination. The shelf assembly shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Chemical composition. Tests for chemical analysis shall be conducted in accordance with method 111.1 or 112.1 of Standard 151.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) shelf assembly, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit and exterior container, as applicable, for the required level of protection specified in procurement document.

Table I. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not required	4 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be packaged in a carton of appropriate size constructed in accordance with PPP-B-636. Carton shall be designed with suitable interior fittings or shall contain sufficient cushioning material to insure against shifting of, or damage to the contents. Closure shall be adequate to prevent spillage of contents under normal handling.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type I, class 2.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix of PPP-B-636. In addition, all seams, including manufacturer's joint, shall be closed by tape conforming to the requirements of PPP-T-76. The tape shall be not less than 2 inches wide. Three strips shall be applied to both top and bottom so that all seams and joints shall be covered their full length. Tape applied over the lengthwise seams shall extend a minimum of 2 inches onto each of the end panels. Tape applied over the end edge seams shall extend a minimum of 2 inches along the side and top, and side and bottom panels.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type I, class 1, fabricated from not less than 275-pound test fiberboard.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. In addition, the letter "L" shall precede the Federal stock number. One space shall be allowed between the letter "L" and the first digit of the stock number. The terms "Stock number" or "FSN" shall not be indicated.

6. NOTES

6.1 Intended use. The shelf assembly, medicine cart, is intended for use with FSN 6530-774-1100, Cart, Medicine, Corrosion-Resisting Steel.

6.2 Ordering data. Procurement document should specify the following:

(a) Number and date of this specification.

(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Custodians:
Army. - MD
Navy. - BuMed
Air Force. - AFMSHB(03)

Review activities:
Army. - MD
Navy. - BuMed
Air Force. - AFMSHB(03)

Preparing activity:
Defense Supply Agency. - DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

FORM APPROVED BUDGET BUREAU NO 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use of the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity.

SPECIFICATION Military Specification, dated 15 August 1964
SHELF ASSEMBLY, MEDICINE CART, CORROSION-RESISTING STEEL MIL-S-36262(DM)

ORGANIZATION	CITY AND STATE

CONTRACT NO.	QUANTITY OF ITEMS PROCURED	DOLLAR AMOUNT
		$

MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?
   ☐ YES ☐ NO IF "YES" IN WHAT WAY?

4. REMARKS (ATTACH ANY PERTINENT DATA WHICH MAY BE OF USE IN IMPROVING THIS SPECIFICATION. IF THERE ARE ADDITIONAL PAPERS, ATTACH TO FORM AND PLACE BOTH IN AN ENVELOPE ADDRESSED TO PREPARING ACTIVITY)

SUBMITTED BY (PRINTED OR TYPED NAME AND ACTIVITY)	DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

1 Under

DEFENSE SUPPLY AGENCY
ENSE MEDICAL SUPPLY CENTER
RD AVENUE AND 29TH STREET
BROOKLYN, N.Y., 11232

OFFICIAL BUSINESS

POSTAGE FEES PAI
DEFENSE SUPPLY AGE

Defense Supply Agency
Defense Medical Supply Center
ATTN: Directorate of Technical Operations (Code T-M5)
3rd Avenue and 29th Street
Brooklyn, N.Y., 11232

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.