MIL-S-36280B
11 June 1975
SUPERSEDING
MIL-S-36280A
12 April 1974

MILITARY SPECIFICATION

STAND, SURGICAL INSTRUMENT

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1 SCOPE

1.1 This specification covers a field-type, knockdown, adjustable stand, suitable for use with a surgical instrument tray (see 6.1).

2 APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein

SPECIFICATIONS

Federal

QQ-A-225/6	Aluminum Alloy Bar, Rod, and Wire; Rolled, Drawn, or Cold Finished, 2024.
QQ-S-637	Steel, Bar, Carbon, Cold Finished (Standard Quality, Free Machining).
QQ-S-698	Steel, Sheet and Strip, Low-Carbon.
TT-C-490	Cleaning Methods and Pretreatment of Ferrous Surfaces for Organic Coatings
TT-E-527	Enamel, Alkyd, Lustreless.
TT-P-636	Primer Coating, Alkyd, Wood and Ferrous metal
VV-L-800	Lubricating Oil, General Purpose, Preservative, (Water-displacing, Low Temperature).
PPP-B-585	Box, Wood, Wirebound.
PPP-B-601	Box, Wood, Cleated-Plywood.
PPP-B-621	Box, Wood, Nailed and Lock-Corner.
PPP-B-636	Boxes, Shipping, Fiberboard.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:20Z
Check the source to verify that this is the current version before use.

Military

MIL-A-8625	Anodic Coatings, for Aluminum and Aluminum Alloys.
MIL-L-10547	Liners, Case, and Sheet, Overwrap, Water Vaporproof or Waterproof, Flexible

STANDARDS

Federal

FED-STD-66	Steel, Chemical Composition and Hardenability
FED-STD-151	Metals; Test Methods
FED-STD-595	Colors

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

10128	Stand, Surgical Instrument, Adjustable, Steel. Olive Drab Enameled, Field

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publication. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids shall apply.

NATIONAL FIRE PROTECTION ASSOCIATION (NFPA)

NFPA	56A Standard for the Use of Flammable Anesthetics

(Application for copies should be addressed to the National Fire Protection Association, 60 Batterymarch Street, Boston, Massachusetts 02110)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:20Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material.

3.1.1 Aluminum rod. Aluminum alloy 2024-T4 in conformance with QQ-A-225/6 shall be used for aluminum rod.

3.1.2 Steel strip. Carbon steel strip shall conform to QQ-S-698.

3.1.3 Steel rod. Carbon steel rod shall conform to QQ-S-637

3.1.4 Corrosion-resisting steel (CRES) rod. CRES rod shall conform to Type 302 or 304 of FED-STD-66

3.2 Construction. The stand shall consist of a base having three legs, an upright, and a tray frame. The design and dimensions shall be in strict accordance with DPSC Drawing No. 10128. The stand shall be so constructed that the tray frame may be pivoted and locked at any height

3.2.1 Salt spray. The stand shall be subjected to a salt spray test in accordance with 4.4.3. There shall be no evidence of corrosion as a result of the test.

3.3 Electrical conductivity. The stand shall meet the requirements of 4694 of NFPA 56A, when tested as specified in 4.4.2.

3.4 Finish

3.4.1 Glides. Steel glides shall have a commercial mill finish

3.4.2 Carbon steel. Carbon steel parts shall be thoroughly cleaned and given a phosphate coating conforming to Type I finish of TT-C-490 Process controls are required in TT-C-490 in conjunction with testing of specimens (panels) to determine effectiveness of process The phosphate coating shall be followed by an overall coating of synthetic primer conforming to TT-P-636, and by two coats of enamel conforming to TT-E-527 Threads shall not be painted but shall be protected with oil conforming to VV-L-800 The color of the enamel shall be olive drab, color No 34087 (see 6.3) conforming to FED-STD-595 The finish shall be baked after the coat of primer and after each coat of enamel. The resultant finish shall be hard smooth, and resistant to abrasion, chipping and cracking.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:20Z
Check the source to verify that this is the current version before use.

3.4.3 Aluminum. Aluminum parts shall be anodized in accordance with MIL-A-8625, type II. Following the anodizing operation, the oxide coating shall be hot-water-sealed.

3.5 Workmanship. Workmanship shall be first class throughout. Stands shall be free from defects which detract from their appearance or may impair their serviceability.

3.5.1 Welding. Welding shall extend the full length of the seam joined and it shall be durable and sound in every detail. Corners shall be rounded and ground smooth to blend with adjacent areas. Prior to welding, all components shall be properly aligned in their proper positions to insure a straight, rigid, and level assembly. Welds shall show good fusion and shall be free from cracks, porosity, gas pockets, and weld spatter.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, and auditory, investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality supplied by the manufacturer of the stand, component, or material may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:20Z
Check the source to verify that this is the current version before use.

4.2 Sampling

4 2 1 For examination Sampling for examination shall be conducted in accordance with MIL-STD-105 and Table I

Table I Sampling for Examination

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	1.0
Total (major and minor combined defects	II	4 0
For dimensional examination	S-3	2 5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in table II.

Table II Sampling for Tests

Component (unit of product)	Characteristic	Requirement	Test procedure	Inspection level	AQL
Metals	Chemical composition	3.1	4.4.1	as req'd	1/
Stand	Electrical conductivity	3 3	4 4 2	S-3	1/
	Salt spray	3 2.1	4 4 3	1 unit per lot	-

1/ Acceptance number shall be zero

4 3 Examination The stand shall be examined to determine compliance with all requirements contained in this specification

4 3 1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects listed in Table III.

Table III Classification of Defects

Categories	Defects
Major:	
101	Stand cannot be assembled from knockdown condition and/or disassembled.
102	Tray frame does not pivot and lock in any position.
cont'd	

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:20Z
Check the source to verify that this is the current version before use.

Table III Classification of Defects (cont'd)

Categories	Defects*
103	Finish not free of abrasion, chipping, cracking, or imbedded foreign matter.
104	Components not free of burrs
105	Component parts improperly aligned
Minor	
201	Threads of tray frame lock handle and adjustable post lock handle are not upset
202	Welds not free from cracks, porosity, gas pockets, and weld spatter
203	Welded corners not rounded, not ground smooth to blend with adjacent areas
204	Threads not protected with preservative oil

*Examination shall not be restricted to the classified possible defects listed above.

4.3.2 Dimensional examination. The stands shall be examined for defects in dimensions. Any dimension not within the specified tolerance shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Chemical composition. Tests for chemical analysis shall be conducted in accordance with method 111.2 or 112.2 of FED-STD-151.

4.4.2 Electrical conductivity. The stand complete with rubber bumpers shall be tested using the equipment described in section 4628(c) of NFPA 56A. One electrode shall be connected to a tray that is in contact only with the rubber bumpers, and the other shall be placed under the glide.

4.4.3 Salt spray. With the unit disassembled and threaded surfaces protected from the spray, painted parts of the stand shall be subjected for 72 hours to the salt spray test described in TT-C-490.

4.5 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table IV.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:20Z
Check the source to verify that this is the current version before use.

Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

Table IV Examination for Preparation for Delivery

Examination	Defect
Containers	Not as specified
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and packing materials	Any component missing or damaged
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application

5 PREPARATION FOR DELIVERY

5.1 Packaging Packaging shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit package. Each stand shall be disassembled. The unit along with the disassembled components shall be wrapped in kraft paper, cushioned with corrugated fiberboard and packaged in a snug fitting box constructed in accordance with PPP-B-636, type CF, class domestic, style optional The box shall contain sufficient cushioning material to insure against shifting of or damage to the contents Box closure shall be as specified in PPP-B-636

5.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose

5.2 Packing Packing shall be level A, B or C as specified (see 6.2)

5.2.1 Level A. Four (4) stands shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3, PPP-B-601, overseas type, PPP-B-621, class 2; or PPP-B-636, class weather resistant Closure and strapping shall be as specified in the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:20Z
Check the source to verify that this is the current version before use.

5 2 1 1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547 Closure and sealing shall conform to applicable paragraphs of appendix thereto Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof as specified in the appendix of PPP-B-636.

5.2 2 Level B Four (4) stands shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3, PPP-B-601, domestic type, PPP-B-621, class 1, or PPP-B-636, class domestic Closure shall be as specified in the appendix of the applicable box specification

5 2 3 Level C The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, to point of delivery called for in the contract or purchase order

5 3 Marking

5 3 1 Unit package and exterior container. Each unit package and exterior container shall be marked in accordance with MIL-STD-129

6 NOTES

6 1 Intended use The STAND, SURGICAL INSTRUMENT, ADJUSTABLE is intended for use with CORROSION-RESISTING STEEL INSTRUMENT TRAY, NSN 6530-00-793-9570

6 2 Ordering data Procurement documents should specify the following

a Number and date of this specification
b Selection of applicable level of packing and packaging (see 5 1 and 5 2)

* 6 3 Availability of color standards The color standard of the color referenced in 3 4 2 is available from

Naval Weapons Plant
Specification Activity
Printed Materials, Supply Division,
Building 197
Washington, DC 20407

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:20Z
Check the source to verify that this is the current version before use.

6.4 This specification does not include all types, classes, grades or sizes of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types which are normally procured to meet Military requirements.

6.5 This specification covers the following item appearing in the Federal Supply Catalog.

National Stock Number	Item Identification
6530-00-551-8681	STAND, SURGICAL INSTRUMENT, Adjustable, Steel, Olive Drab Enameled, Field

6.6 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6530-1354
1S

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:20Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No 22-R255

INSTRUCTIONS The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay or will otherwise enhance use of the document DoD contractors government activities or manufacturers vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side staple in corner and send to preparing activity Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements Attach any pertinent data which may be of use in improving this document If there are additional papers attach to form and place both in an envelope addressed to preparing activity

DOCUMENT IDENTIFIER AND TITLE
MIL-S-36280B-STAND, SURGICAL INSTRUMENT

NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER
	MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1 HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A GIVE PARAGRAPH NUMBER AND WORDING

B RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2 COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3 IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO (If "Yes", in what way?)

4 REMARKS

SUBMITTED BY (Printed or typed name and address - Optional)	TELEPHONE NO
	DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:20Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD 304

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS DEFENSE PERSONNEL SUPPORT CENTER
ATTN DIRECTORATE OF MEDICAL MATERIEL CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:20Z
Check the source to verify that this is the current version before use.