MILITARY SPECIFICATION

TIP ASSORTMENT, UTERINE CANNULA, PLASTIC, DISPOSABLE

1. SCOPE

1.1 This specification covers plastic tips for use with a uterine cannula.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-up, Paperboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, Water Resistant (for Carton Sealing).

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

: FSC 6515 :

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:15Z
Check the source to verify that this is the current version before use.

DRAWING

Defense Medical Supply Center

20543 Tip, Uterine, Cannula

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. Tips shall be molded of high quality polyethylene, suitable for the purpose intended.

3.2 Style, design, and construction. The tips shall be suitable for use in uterotubal insufflation and shall consist of one (1) large tip and four (4) small tips. The tips shall be in accordance with Defense Medical Supply Center Drawing 20543.

3.3 Instructions. Tips shall be provided with instructions describing proper methods for sterilizing. In addition, information cautioning the user against sterilizing methods that would be harmful to the tips shall be included.

3.4 Workmanship. Workmanship shall be first class throughout. Tips shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:15Z
Check the source to verify that this is the current version before use.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling. Sampling for examination shall be conducted in accordance with Military Standard MIL-STD-105.

4.3 Examination. The tips shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) large and four (4) small tips, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I - Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	12 units	288 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:15Z
Check the source to verify that this is the current version before use.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be packaged in an envelope of appropriate size and design, and shall be of sufficient strength to adequately protect the contents. Closure shall be adequate to prevent spillage of contents under normal handling.

5.1.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with Specification PPP-B-566, or PPP-B-676, except that commercial colors shall be acceptable. Closure shall be adequate to prevent spillage of contents under normal handling.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with Federal Specification PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type I, class 2.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to Military Specification MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix of Federal Specification P -B-636. In addition, all seams, including manufacturer's joint, shall be closed by tape conforming to the requirements of Federal Specification PPP-T-76. The tape shall be not less than 2 inches wide. Three strips shall be applied to both top and bottom so that all seams and joints shall be covered their full length. Tape applied over the lengthwise seams shall extend a minimum of 2 inches onto each of the end panels. Tape applied over the end edge seams shall extend a minimum of 2 inches along the side and top, and side and bottom panels.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with applicable paragraphs of appendix thereto.

NOTE: Strapping will not be required for shipment forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:15Z
Check the source to verify that this is the current version before use.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to Federal Specification PPP-B-601, table I, using type III, class 1 plywood, or shall be designed for type 2 load, and constructed in accordance with Federal Specification PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type I, class 1.

5.3 Marking.

5.3.1 Unit package. Each envelope shall be marked with Federal stock number, item identification, quantity, unit of issue, contract number, and name of contractor. In addition, information required by 3.3 shall also be shown.

5.3.2 Intermediate package. Each intermediate package shall be marked in accordance with Military Standard MIL-STD-129. Waterproofing of labels shall be required only when carton utilized is fabricated of water-resistant material. Contract number and name of contractor shall be shown.

5.3.3 Exterior container. Exterior container shall be marked in accordance with Military Standard MIL-STD-129. In addition, the letter "L" shall precede the Federal stock number. One space shall be allowed between the letter "L" and the first digit of the stock number. The terms "Stock number" or "FSN" shall not be indicated.

NOTE: The quantity and unit of issue on unit and intermediate packages, and on exterior containers shall be designated, as follows:

Unit Package	"1 PKG (of 5 tips)"
Intermediate Package	"12 PKG"
Exterior Container	"288 PKG"

6. NOTES

6.1 Intended use. The uterine cannula tips shall be suitable as replacement tips for FSN 6515-890-1105, Cannula, Uterine, Corrosion-Resisting Steel.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:15Z
Check the source to verify that this is the current version before use.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.

(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet military requirements.

Custodians:
Army. - MD
Navy. - BuMed
Air Force. - AFMSHB(03)

Preparing activity:
Defense Supply Agency. - DM.

Review activities:
Army. - MD
Navy. - BuMed
Air Force. - AFMSHB(03)

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:15Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE SUPPLY AGENCY
DEFENSE MEDICAL SUPPLY CENTER
3RD AVENUE AND 29TH STREET
BROOKLYN, N.Y., 11232

OFFICIAL BUSINESS

POSTAGE FEES PAID
DEFENSE SUPPLY AGENCY

Defense Supply Agency
Defense Medical Supply Center
ATTN: Directorate of Technical Operations (Code T-M2)
3rd Avenue and 29th Street
Brooklyn, N.Y., 11232

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:15Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity *(as indicated on reverse hereof)*.

SPECIFICATION

ORGANIZATION *(Of submitter)*	CITY AND STATE

CONTRACT NO.	QUANTITY OF ITEMS PROCURED	DOLLAR AMOUNT
		$

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES.

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO IF "YES", IN WHAT WAY?

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity)*	DATE

DD FORM 1 APR 63 1426 REPLACES NAVSHIPS FORM 4863, WHICH IS OBSOLETE PLATE NO. 15419

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:15Z
Check the source to verify that this is the current version before use.