MIL-T-36263A
22 August 1973
SUPERSEDING
MIL-T-36263 (DM)
17 August 1964

MILITARY SPECIFICATION

TRAYS, MEDICAL INSTRUMENT AND SUPPLY SET CHEST

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE.

1.1 Scope. This specification covers three sizes of lightweight perforated aluminum insert trays intended for use with the chests specified in 6.1.

1.2 Classification. The trays covered by this specification shall be of the following sizes, as specified (see 6.2):

Size	Outside dimensions Length	Width	Depth
A	23-25/32	11-3/4	5-27/32
B	29	16-27/32	5-27/32
C	29	16-27/32	9-27/32

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

QQ-A-250/8	Aluminum Alloy, 5052, Plate and Sheet.
TT-E-527	Enamel, Alkyd, Lustreless.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-T-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (for Carton Sealing).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

Military

MIL-P-8585	Primer Coating, Zinc Chromate, Low-Moisture-Sensitivity.
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Federal

Fed. Test Method Std. No. 151	Metals; Test Methods.
Fed. Std. No. 595	Colors.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection.
MIL-STD-129	Marking for Shipment and Storage.

DRAWINGS

Defense Personnel Support Center

10118	Tray, Medical Instrument and Supply Set Chest, 23-25/32 inches Long by 11-3/4 inches Wide by 5-27/32 inches High.
10119	Tray, Medical Instrument and Supply Set Chest, 29 inches Long by 16-27/32 inches Wide by 5-27/32 inches High.
10120	Tray, Medical Instrument and Supply Set Chest, 29 inches Long by 16-27/32 inches Wide by 9-27/32 inches High.

Each of the above drawings includes a Bill of Materials.

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. Except as specified herein, materials shall conform to the applicable portions of the specifications, standards, and drawings (with bill of materials) specified herein, and shall be capable of meeting the requirements specified therein. In addition, the metals shall conform to the chemical requirements of their applicable specifications when tested as specified in Section 4. When materials to be used are not specifically designated, those used shall be of the highest commercial quality, and entirely suitable for the purpose intended.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

3.1.1 Aluminum alloy. Aluminum alloy sheets shall conform to QQ-A-250/8. All parts shall be temper H34, except sides (including ends) which shall be temper H32.

3.1.2 Primer. Primer shall conform to MIL-P-8585 and shall be color Y.

3.1.3 Enamel. The enamel for identification marking shall conform to TT-E-527. Color shall be black, color 37038 of FED. STD. No. 595.

3.2 Construction. The aluminum insert trays shall conform in all details to the requirements specified herein, and as shown on the applicable drawing. The insert trays shall be fabricated from aluminum alloy sheets as specified in 3.1.1; sheet material for the bottom and sides shall be perforated along the full length (see drawings); sheet material for the end reinforcing angles shall be solid sheets. Perforated sheet shall be continuously perforated with elongated holes 1 by 0.156 inches, on 4-5/8 inches and 3/4 inch centers as shown on the drawings. The bottom shall be of one-piece construction. The sides shall be formed from two pieces with ends joggled, lapped, and joined by spot welding as shown on the applicable drawing. Sides shall be provided with 90 degree flanges with return bends. Bottom shall be spot welded to flanges. Reinforcing angles shall be spot welded to the ends.

3.3 Welding. Welding shall be durable, sound in every detail and in accordance with latest practices recommended by the American Welding Society. All welds shall be produced by qualified welding operators with satisfactory welding equipment using satisfactory filler metal or electrode where applicable to provide the best possible weld. Surfaces to be welded shall be free from grease, oil, oxides or any other foreign matter; any cleaning method which will remove these substances and leave the surface clean is satisfactory. All weld joints shall blend into the adjacent metal in gradual, smooth curves. Welds shall not be oxidized. Spot welds shall be free from surface burning, cracks, and porosity. Spot welds shall be capable of developing the tensile strength of the parent metal.

3.4 Tolerances. Tolerances shall be as specified in the drawing. Particular accuracy shall be maintained in perforating the sheet stock and in jigging and welding insert trays to insure the proper location of the critical holes designated on the drawing.

3.5 Finish. Completed trays shall have a finish characteristic of the cleaning process used in 3.3.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

3.6 Identification marking. The identification marking shall be located on the outer surface of one of the sides of the tray. The area covered by the marking shall be prepared with primer as specified in 3.1.2. The marking shall be permanently and legibly applied with black enamel (see 3.1.3), applied on the primer. The characters shall be of appropriate size and shall consist of the applicable National stock number, size number, (A, B, or C), the item identification, name or registered trademark of the manufacturer, and the contract or order number. As an alternate, the identification marking shall be die stamped on one side of the tray.

3.7 Workmanship. Workmanship shall be first class throughout. Sharp edges and burs shall be removed. All corners shall be broken. Trays shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturers of the aluminum sheet may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in tables I and II (see 4.2.1 and 4.2.2) and 4.2.3.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in table I. Unit of product for examination purposes shall be one tray.

Table I. Sampling for Examination

Examination	Inspection level	AQL (percent defective)
For visual examination		
Major	II	2.5
Minor	II	6.5
For dimensional examination	S-1	4.0

4.2.2 Examination. The tray shall be examined for defects including, but not limited to, those listed in table II.

Table II. Classification of Defects

Categories	Defects
Major	
101	Any component missing.
102	Spotwelds not located as per drawing.
103	Perforations incomplete.
104	Burs on tray.
105	Sharp edges on tray.
106	Marking on tray missing or incomplete.
Minor	
201	Tray not free of scratches.
202	Marking on tray not completely legible.
203	Finish not uniform.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

4.2.3 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, Inspection Level S-1, Acceptable Quality Level 1.0, except that the minimum sample size shall be 3 trays of the same size. Unit of product for test purposes shall be one tray.

4.2.4 For chemical testing of metals. Four ounces of each lot of each metal used in the construction of the trays shall be selected for chemical analysis as specified in 4.3.1.

4.3 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.3.1 Chemical analysis of the metals. The samples selected in accordance with 4.2.4 shall be analyzed in conformance with methods 111 or 112 of FED. TEST METHOD No. 151. In case of dispute, method 111 shall be used for referee analysis.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) tray, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table III shall be packaged in each unit and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table III. Packaging-quantities

Size	Unit package	Intermediate package	Exterior container
A	1 unit	Not required	8 units
B and C	1 unit	Not required	4 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

5.1.3 Level A.

5.1.3.1 Unit package. Unit package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent accidental opening under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification. Special requirements of applicable Table shall apply for PPP-B-636 box.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Fiberboard box shall be closed by method II of PPP-B-636. Special requirements of applicable Table shall apply for PPP-B-636 box.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Intended use. The trays are intended for use with the size of chests as indicated below for the appropriate National stock number and item identification.

NSN 6545-00-925-9200, TRAY, MEDICAL INSTRUMENT AND SUPPLY SET CHEST, 23-25/32 inches Long by 11-3/4 inches Wide by 5-27/32 inches High, for use with NSN 6545-00-914-3460, CHEST, MEDICAL INSTRUMENT AND SUPPLY SET, Field, No. 1, 24 inches Long by 12 inches Wide by 6 inches Deep.

NSN 6545-00-925-9210, TRAY, MEDICAL INSTRUMENT AND SUPPLY SET CHEST, 29 inches Long by 16-27/32 inches Wide by 5-27/32 inches High, for use with NSN's 6545-00-914-3470 CHEST, MEDICAL INSTRUMENT AND SUPPLY SET, Field No. 2, 30 inches Long by 18 inches Wide by 6 inches Deep, 6545-00-914-3490 CHEST, MEDICAL INSTRUMENT AND SUPPLY SET, Field No. 4, 30 inches Long by 18 inches Wide by 12 inches Deep, and 6545-00-914-3500, CHEST, MEDICAL INSTRUMENT AND SUPPLY SET, Field, No. 5, 30 inches Long by 18 inches Wide by 16 Inches Deep.

NSN 6545-00-925-9220 TRAY, MEDICAL INSTRUMENT AND SUPPLY SET CHEST, 29 inches Long by 16-27/32 inches Wide by 9-27/32 inches High, for use with NSN's 6545-00-914-3480, CHEST, MEDICAL INSTRUMENT AND SUPPLY SET, Field No. 3, 30 inches Long by 18 inches Wide by 10 inches Deep, 6545-00-914-3500, CHEST, MEDICAL INSTRUMENT AND SUPPLY SET, Field No. 5, 30 inches Long by 18 inches Wide by 16 inches Deep, and 6545-00-914-3510, CHEST, MEDICAL INSTRUMENT AND SUPPLY SET, Field, No. 6, 30 inches Long by 18 inches Wide by 20 inches Deep.

Copies of Defense Medical Purchase Descriptions covering the above referenced items may be obtained from the procuring activity or as directed by the contracting officer.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Size required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification does not include all types, classes, grades, or sizes, of the commodity indicated by the title of this speficication, or which are commercially available, but is intended to cover the types which are normally procured to meet military requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project Number: 6545-0013

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation shall be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-T-36263A

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.