MIL-T-36264A
9 September 1968
Superseding
MIL-T-36264
4 September 1964

MILITARY SPECIFICATION

TUBE, COLON, SURGICAL, DISPOSABLE, 28 FR, 50s

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a rectal tube, suitable for irrigation of the human colon in medical treatment procedures.

2. APPLICABLE DOCUMENTS

2.1 Specifications, standards and drawing. The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-585	Boxes; Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

DRAWING

Defense Personnel Support Center

20750 TUBE, COLON, SURGICAL, Disposable, 28 FR, 50s

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

U.S. Pharmacopeial Convention, Inc.

Pharmacopeia of the United States.

(Application for copies should be addressed to the Mack Publishing Company, Easton, Pa.).

3. REQUIREMENTS

3.1 Material. All components coming in contact with solutions being administered shall contain not more than 5 parts per million (p.p.m) of lead when tested as specified in 4.4.5.

3.1.1 Plastic tubing. Plastic tubing shall be transparent, medical-grade, polyvinyl, and shall be sterile, pyrogen-free, and pass the animal implantation test when tested as specified in section 4. When specified, the tubing shall be safe (nontoxic),(see 3.2.2), when tested as specified in 4.4.3. Tubing shall be unaffected by oil base lubricants.

3.2 Style, design and dimensions. The colon tubes shall be either with integral connector, (type I) or, with an attached connector (type II), in accordance with Drawing 20750, and as hereinafter specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

3.2.1 Type I. Type I colon tube shall be fabricated with a connector which is an integral part of the plastic tubing (formed from this tubing). The tube shall comply with the requirements of 3.1, 3.1.1, 3.3, 3.4, 3.6 and 3.10.

3.2.2 Type II. Type II colon tube shall be fabricated with an attached connector. The tube shall comply with the requirements of 3.1, 3.1.1, 3.3, 3.4, 3.5, 3.6, 3.9 and 3.10.

3.3 Sterility. Tubes shall be sterile when tested for sterility as specified in 4.4.1.

3.4 Freedom from pyrogen. Tubes shall be pyrogen-free when tested for freedom from pyrogens as specified in 4.4.2.

3.5 Safety (nontoxicity). Tubes shall be safe (nontoxic) when tested for safety as specified in 4.4.3.

3.6 Animal implantation test. Plastic tubing only shall be tested and shall show no evidence of toxicity or adverse tissue reaction upon gross and microscopic examination, when tested as specified in 4.4.4.

3.7 Certification for sterility, freedom from pyrogen, safety, and animal implantation test. A certificate (s) listing all lot or control numbers, shall state that representative sample (s) from each lot or control have been tested and have been found to be sterile, free from pyrogens, safe and have passed the animal implantation test. This certificate (s) shall include the specific tests performed and the test results. The record of the animal implantation test and its results shall be certified by the Pathologist who performed or supervised performance of the animal implantation test. For destination inspection certificates shall be submitted by the contractor directly to Defense Personnel Support Center, Directorate of Medical Materiel, Technical Operations Division, 2800 South 20th Street, Philadelphia, Pa. 19101 attention: Quality Control Branch, and one copy of each shall be forwarded with each shipment to the consignee. For source inspection, certificates shall be furnished to the procuring activity via the government inspector.

3.8 Manufacturer's record. The manufacturer shall place a code or lot number on the paper label and on the printed carton, which can be verified against the manufacturer's records of all tests made on nontoxic, animal implantation test, sterile, pyrogen-free sets. These records shall be made available for inspection by any duly authorized Government representative. No shipment shall be made until all tests have been completed, verified, and recorded in the permanent records of the manufacturer.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

3.9 Security of assembly. The connector of type II tube shall not separate or loosen from the plastic tubing when tested as specified in 4.4.6.

3.10 Sterilization. When tested as specified in 4.4.7, tubes shall show no signs of brittleness, blistering, swelling, peeling, cracking or any other damage or deformation.

3.11 Workmanship. Workmanship shall be first class throughout. Tubes shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilizechis own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

4.2.2 For tests.

4.2.2.1 For pyrogen, sterility, and safety tests. Sampling procedures for pyrogen, sterility, and safety tests shall be conducted as indicated in the current edition of the Pharmacopeia of the United States (U.S.P.), for transfusion and infusion assemblies.

4.2.2.2 For implantation test. Two samples shall be cut from every manufacturer's lot of tubing for test.

4.2.2.3 For security of assembly. Sample size for test shall be as indicated for inspection level S-3 and AQL of 1.5 percent defective of MIL-STD-105. The unit of product for inspection purposes shall be one tube.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

4.2.2.4 Sterilization tests. Sampling for each test shall be conducted in accordance with MIL-STD-105, inspection level S-1, except that the acceptance number shall be zero (0) for all samples sizes, and a minimum of four (4) samples shall be taken regardless of lot size. Unit of product shall be one tube.

4.2.2.5 For lead test. Sample shall be taken from each lot of plastics as indicated in 4.4.5.

4.2.3 Sample unit usage. Where possible, the same unit shall be used for the determination or two or more test characteristics.

4.3 Examination. The tubes shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Sterility. Tubes shall be tested for sterility in accordance with the current edition of the U.S.P., for transfusion and infusion assemblies.

4.4.2 Pyrogen. Tubes shall be tested for freedom from pyrogens in accordance with the current edition of the U.S.P., for transfusion and infusion assemblies.

4.4.3 Safety test. The tubes shall be tested for safety in accordance with the current edition of the U.S.P., for transfusion and infusion assemblies.

4.4.4 Animal implantation test. Samples of the tube shall be tested in accordance with the current edition of the U.S.P. for Implantation Test, Plastic Containers, for class VI plastics, except after "72 hours" add "and not to exceed 144 hours"; and delete 'Examine macroscopically the area of the tissue surrounding the center portion of each implant strip. Use a magnifying lens if necessary' and substitute: 'The muscle tissue surrounding the implants shall be examined, both grossly and microscopically, by a pathologist certified by the American Board of Pathology or the American College of Veterinary Pathologists'."

4.4.5 Test for lead. Ten (10) grams of sample shall be taken, ignited to char, and then further ignited in a muffler at 700° Centigrade (C.) to a white ash. The ash shall be boiled in 14 milliliters (ml.) of 10 percent nitric acid filtered into a volumetric flask and made up to 100 ml. An aliquot portion, equivalent to one gram of sample, shall be taken and placed in a separator. The dithizone test for lead, as given in the current edition of the U.S.P., shall be applied using a control of 5 p.p.m. or 12 p.p.m. of lead solution as given in the U.S.P. Test. If the sample solution is darker than the limit, smaller aliquots shall be used until the color of the control is matched by the sample solution and the concentration of lead in p.p.m. is determined.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

4.4.6 Security of assembly. The tubes shall be conditioned at 240° Fahrenheit (F.) and 100 percent relative humidity for 20 minutes. The security of assembly of connector of the type II shall be tested by gradually applying a five-pound weight to connector and tubing held in a clamping device so that the connector hangs vertically downward.

4.4.7 Sterilization.

4.4.7.1 Steam sterilization. Tubes shall be steam autoclaved at 15 p.s.i.g. and 250° Fahrenheit (F.) for 15 minutes and then allowed to cool at room temperature (68° F. to 72° F.).

4.4.7.2 Boil sterilization. Tubes shall be boiled in distilled water for 30 minutes and then allowed to cool at room temperature (68° F to 72° F.).

4.4.7.3 Cold sterilization. Tubes shall be immersed in a chemical disinfectant for 20 minutes at room temperature (68° F to 72° F) and then followed by five (5) water rinses.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One package (50 tubes), as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in the following table shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table — Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not required	4 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit packaging.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

5.1.3.1.1 Individual package. Each tube shall be packaged straight so that it will not kink or be permanently deformed, and shall be packaged in a bag fabricated from seamless polyethylene tubing having an average wall thickness of not less than 0.002 inch, with no wall thickness at any point below 0.00125 inch. Excess air shall be expelled or evacuated, and closure effected by heat sealing. The completed package shall be adequate to maintain sterility in accordance with the requirements of 3.3.

5.1.3.1.2 Unit package. Each unit shall be packaged in a box of appropriate size constructed in accordance with PPP-B-636. Closure shall be adequate to prevent accidental opening.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type I, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant. Closure and strapping shall be in accordance with appendix of applicable box specification.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table III, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, type 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic. Closure shall be in accordance with appendix of applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Individual package. Each individual package shall be marked with the name of the manufacturer, commercial description of contents, lot (control) number, the legend "Sterility not guaranteed if package is broken or damaged", (or similar statements).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

5.3.2 Unit package. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number, name of contractor and lot (control) number shall be shown.

5.3.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.

(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification is intended to cover the types which are normally procured to meet military requirements. It does not include all types, etc., as indicated by the title nor all types which are commercially available.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL CODE: ATS
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
AUG 66

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements.

SPECIFICATION

MIL-T-36264A

ORGANIZATION

CITY AND STATE | CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)* | DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.