MIL-T-36267 (DM)
19 August 1964

MILITARY SPECIFICATION

TUBE, IRRIGATOR, SURGICAL, VALENTINE
RUBBER, 7 Ft.

1. SCOPE

1.1 This specification covers rubber tubing of the Valentine design, suitable for irrigating purposes.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

ZZ-T-831	Tubing; Rubber.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up, Paperboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (for Carton Sealing).

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

Military

MIL-L-10547 Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3.1. Material.

3.1.1 Proximal and distal ends. The proximal and distal end of the irrigator tube shall be pure gum tubing in accordance with grade L, type I, light wall rubber of ZZ-T-831.

3.1.2 Tubing proper. The tubing proper shall be red rubber in accordance with grade A, light wall rubber of ZZ-T-831.

3.2 Style, design and dimensions. The irrigator tube shall be of the Valentine design, 7 feet ± 2 inches in overall length, consisting of an expanded proximal end, a middle section (tube proper), and a partially reinforced distal end.

3.2.1 Proximal end. The proximal end shall be 4-1/2 ± 1/4 inches long. The proximal end shall be funnel-shaped, having an inside diameter of 13/16 ± 1/32 inch at the top, and the base being of same dimension as tubing proper to which this funnel-shaped end shall be securely vulcanized. The wall thickness of the proximal end shall be 1/16 ± 1/64 inch when measured as specified in 4.4.1.

3.2.2 Tubing proper. The tubing proper shall be 73-1/2 ± 1/2 inches long, shall have an outside diameter of 3/8 ± 1/64 inch, and a wall thickness of 1/16 ± 1/64 inch.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

3.2.3 Distal end. The distal end of the irrigator tube shall be 6 ± 1/4 inches long and shall be securely vulcanized to the other end of the tubing proper. This distal end shall be of the same dimensions as the tubing proper, except for the distal 1 inch of this end of the irrigator tube, which shall be reinforced by being of double-wall thickness. This reinforced portion of the distal end shall have a wall thickness of 1/8 ± 1/32 inch, when measured as specified in 4.4.1, and an inside diameter of 3/16 ± 1/64 inch. The remaining 5 inches of the distal end shall be of the same dimensions as the tubing proper.

3.3 Workmanship. Workmanship shall be first class throughout. Tubes shall be free from defects which detract from their appearance or may impair their serviceability.

3.4 Identification marking. Each tube shall be permanently marked with the name or registered trademark of the contractor and the date of cure (month and year). As an alternate, a coding system may be used in marking each tube, provided the code represents the month (or quarter of the year) and year of cure and the procuring agency is advised of the coding system being used prior to delivery of the items.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 Sampling for examination. Sampling procedure for examination shall be conducted in accordance with the procedures set forth in MIL-STD-105.

4.2.2 Sampling for tests. Sample size for each test shall be as indicated for inspection level S1 in MIL-STD-105, except that the acceptance number shall be zero (0) for all sample sizes. Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The tubes shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Wall thickness measurements. Measurements for wall thickness shall be made at a minimum of three locations. The average of these measurements shall apply.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) tube, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	12 units	144 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be formed into a coil of such diameter that it will not kink or be permanently deformed. Unit package shall be a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors shall be acceptable. Closure shall be adequate to prevent spillage of conents under normal handling.

5.1.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors shall be acceptable, or PPP-B-636. Closure shall be adequate to prevent spillage of contents under normal handling.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant.

5.2.1.2 Case liner. Each level A wood box shall be lined with a water-proof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix of PPP-B-636. In addition, all seams, including manufacturer's joint, shall be closed by tape conforming to the requirements of PPP-T-76. The tape shall be not less than 2 inches wide. Three strips shall be applied to both top and bottom so that all seams and joints shall be covered their full length. Tape applied over the lengthwise seams shall extend a minimum of 2 inches onto each of the end panels. Tape applied over the end edge seams shall extend a minimum of 2 inches along the side and top, and side and bottom panels.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. In addition, the letter "L" shall precede the Federal stock number. One space shall be allowed between the letter "L" and the first digit of the stock number. The terms "Stock number" or "FSN" shall not be indicated.

6. NOTES

6.1 Intended use. The irrigator tube is intended for use with FSN 6530-782-2075 IRRIGATOR, Corrosion-Resisting Steel, 2 Qt. and FSN 6515-385-3900 TUBE, CONNECTING, Surgical Apparatus, Straight Glass.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.

(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet military requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

Custodians:
Army. - MD
Navy. - MS
Air Force. - 03

Review activities:
Army. - MD
Navy. - MS
Air Force. - 03

Preparing activity:
Defense Supply Agency. - DM.

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

FORM APPROVED BUDGET BUREAU NO. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use of the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity.

SPECIFICATION TUBE, IRRIGATOR, SURGICAL, VALENTINE RUBBER, 7 Ft. MIL-T-36267 (DM)

ORGANIZATION	CITY AND STATE

CONTRACT NO.	QUANTITY OF ITEMS PROCURED	DOLLAR AMOUNT $

MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?
   ☐ YES ☐ NO IF "YES" IN WHAT WAY?

4. REMARKS (ATTACH ANY PERTINENT DATA WHICH MAY BE OF USE IN IMPROVING THIS SPECIFICATION. IF THERE ARE ADDITIONAL PAPERS, ATTACH TO FORM AND PLACE BOTH IN AN ENVELOPE ADDRESSED TO PREPARING ACTIVITY)

SUBMITTED BY (PRINTED OR TYPED NAME AND ACTIVITY)	DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold Under

DEFENSE SUPPLY AGENCY
DEFENSE MEDICAL SUPPLY CENTER
3RD AVENUE AND 29TH STREET
BROOKLYN, N.Y., 11232

OFFICIAL BUSINESS

POSTAGE FEES PAI
DEFENSE SUPPLY AGE

Defense Supply Agency
Defense Medical Supply Center
ATTN: Directorate of Technical Operations (Code T-MS)
3rd Avenue and 29th Street
Brooklyn, N.Y., 11232

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:18Z
Check the source to verify that this is the current version before use.