MIL-W-36258D
SEPTEMBER 27, 1976
SUPERSEDING
MIL-W003628C (DSA-DM)
18 January 1974 and
MIL-W-36258B (DSA-DM)
4 May 1973

MILITARY SPECIFICATION

WATER TESTING KIT, BACTERIOLOGICAL

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers one type of water analysis kit, suitable for determining the potability of drinking water.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

FEDERAL

W-C-596	-	Connector, Plug, Electrical; Connector, Receptacle, Electrical.
SS-P-196	-	Pencil; Spiral-Paper Form, Wax Glass- and China-Marking.
PPP-B-601	-	Boxes, Wood, Cleated-Plywood.
PPP-B-636	-	Boxes, Shipping, Fiberboard.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Hq, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document, or by letter.

FSC 6665

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-C-4150 - Case, Carrying, Water-Vaporproof.
MIL-M-7298 - Manual, Technical; Commercial Equipment.
MIL-L-10547 - Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

FEDERAL

FED-STD-66 - Steel; Chemical Composition and Hardenability.

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

20887 Water Testing Kit, Bacteriological.

(Copies of specifications, standards and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publication. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

UNDERWRITERS' LABORATORIES, INC. (UL)

UL 62 - Flexible Cord and Fixture Wire.

(Application for copies should be addressed to the Underwriters' Laboratories, Inc., Walt Whitman Road, Melville, L.I. NY 11746, 207 East Ohio Street, Chicago, IL 60611, and 1566 Scott Boulevard, Santa Clara, CA 95050.)

3. REQUIREMENTS

3.1 Material. Material shall be of the highest quality and as herein specified.

3.1.1 Corrosion-resisting steel. Corrosion-resisting steel shall conform to FED-STD-66, 300 series.

3.1.2 Fiberglass plastic. The fiberglass plastic material for the case shall conform to MIL-C-4150.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

3.2 Style, design and construction. Water testing kit shall be in accordance with drawing 20087. Kit shall be capable of accommodating sufficient equipment and materials to accomplish 24 bacteriological water analysis under field conditions, using the membrane filter technique. The complete kit shall be enclosed in a fiberglass-reinforced plastic carrying case and shall be of such size and weight as to permit handcarrying by one man. Kit shall provide an incubator with operating controls, including an incubator rack, a filter holder assembly and accessories, repair parts and tools.

* 3.2.1 Carrying case. The carrying case shall be of blue fiberglass-reinforced plastic, with overall nominal dimensions (without handle) of 14-3/8 inches long by 9-7/8 inches wide by 9-13/16 inches high, All corners and edges shall be rounded. Case shall be moisture- and mold-resistant throughout (see 3.1.2), and shall be capable of withstanding rough handling under field conditions. A plastic, collapsible carrying handle shall be securely mounted on top of the cover. Openings of the cover and body shall have an aluminum alloy metal edging designed so as to provide a rainproof and splashproof closure, when tested as specified in 4.4.5. All perforations through the case for rivets, screws, etc., shall be sealed to prevent seepage of moisture. A plastic holding device (insert) shall be mounted on the inside of the cover to hold the following components securely in place as shown in the drawing: filter holder assembly (F); graduated sampling cup (C); metal syringe (A); 4-ounce (B) and 8-ounce (E) plastic bottles; package of Endo Broth (D). The holding device (insert) shall be so designed as to be easily removed and interchanges among similar containers. Components shall be located for convenient and easy removal and shall be so held as to prevent movement or damage in transit. All components except Endo Broth shall be furnished with kit. Endo Broth shall not be furnished (see 6.3).

Cover shall not be removable and shall be attached on one long side by an aluminum or corrosion-resisting steel piano hinge or a latching -type hinge. The case shall be provided with a sufficient number of pulldown corrosion-resisting metal clasps of suitable size and tension to afford a rainproof and slashproof closure, when tested in accordance with 4.4.5. The carrying case shall be so designed that the cover may be opened to a maximum position of 90 degrees without the front end of the case lifting from the surface, when tested as specified in 4.4.4.

3.2.2 Incubator. A plastic liner, with provision for accommodating a plastic or aluminum incubator cavity and a conveniently located control panel, shall be suitably fitted into the body of the carrying case in accordance with the drawing. The incubator cavity shall

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

have a clear acrylic-plastic lid, not less than 7-1/8 inches square by 5/16-inch thick and actuated by a corrosion-resisting steel piano hinge (see 3.1.1). Edges of the lid shall be so designed as to provide a close fit with the mouth edges of the cavity. The lid shall be held closed by a spring-action latch. A temperature-indicating device shall be so located and secured as to measure the median temperature in the cavity of the incubator. The incubator shall be capable of maintaining an operating temperature of 36° Centigrade (°C.) ±1° (see 3.2.5) as measured on the surface of a nutrient-saturated filter disk in a sealed culturing dish (see 6.2, 6.3 and 6.4) within the incubator cavity in ambient temperatures varying from -30° Fahrenheit (°F.) to 98.5°F., when tested as specified in 4.4.1.

3.2.3 Incubator rack. A square corrosion-resisting wire rack shall be furnished in accordance with the drawing and shall be capable of accommodating 24 Petri culture dishes (see 6.2) for simultaneous incubation. The rack shall also be capable of accommodating 60 millimeter (mm.) Petri dishes. The frame shall be made of 0.134 inch diameter (No. 10 Birmingham Wire (B.W.) gauge) wire, and cross wires shall be made of 0.083 inch diameter (No. 14 B.W. gauge) wire. Gauges specified are minimum. Cross wires shall be so located that wells are provided to accommodate not less than forty-eight 13-by-100-mm. test tubes in an upright position. The bottom shelf shall support the tubes within the rack. Wires shall be securely welded or brazed in position. All wire ends shall be finished smooth and shall be without sharp edges.

3.2.4 Filter holder assembly. The filter holder assembly shall consist of a funnel assembly with locking ring; a filter support assembly with removable screen, secured to a supporting base of aluminum alloy; and a vacuum filter flask (beaker).

3.2.4.1 Funnel assembly. The funnel assembly shall consist of a corrosion-resisting steel 100 ml.-capacity (minimum) funnel, 0.036 inch minimum thick of general funnel design, the base of which shall be collared with an aluminum locking ring with nylon plastic rollers (wheels) suitably positioned on the inside. These rollers shall facilitate a twist-type locking action of the locking ring. The filter support assembly shall be so designed that turning the locking ring clockwise shall cause the funnel assembly to lock securely in place and afford a leakproof seal. When tested as specified in 4.4.3, a vacuum of 25-26 inches of mercury in the system shall not drop below 20 inches of mercury within a period of 20 minutes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

3.2.4.2 Filter support assembly. The filter support assembly shall be fabricated of smooth finish corrosion-resisting steel (see 3.1.1) and shall consist of a removable screen (100-mesh minimum) assembly and the filter base upon which it shall be centrally positioned as shown the drawing.

* 3.2.4.3 Supporting base. The supporting base shall be fabricated of aluminum alloy of sufficient weight to afford stability for the complete filter holder assembly, and shall be in accordance with the drawing. An asbestos wick of 1/4-inch nominal width shall be contained in a recess around the top of the base and shall serve to sterilize the filtering components by means of formaldehyde gas (produced by the incomplete combustion of methyl alcohol). Two Buna-N-Rubber O-rings shall be contained in grooves around the perimeter of the base to afford a complete seal with the vacuum filter flask (see 4.4.3). An outlet hole of luer taper design shall connect to the bottom inner well of the supporting base. Three or four nylon plastic feet shall be provided in a manner to afford stability to the supporting base. There shall be two release grooves on the perimeter to provide assistance in removing the support base from the vacuum filter flask.

3.2.4.4 Vacuum filter flask. The vacuum filter flask shall be a corrosion-resisting steel (see 3.1.1) beaker, 0.031 inch thick (No. 22 U.S. Standard gauge) minimum and shall be not less than 590 ml. capacity. The rim of the filter flask shall be so formed that it permits formaldehyde gas sterilization and also forms a leakproof seal with the two O-rings on the sides of the supporting base.

3.2.5 Controls. The incubator shall be capable of operating on 6,12 and 24 volts, direct current; and 115 and 230 volts, 50 to 60 cycles, alternating current, when tested as specified in 4.4.2. A voltage selector switch, a temperature control dial, a pilot light and a fuse holder shall be located on a corrosion-resisting steel (see 3.1.1) plate mounted in the case. A male "twist-lock" type receptacle for the line cord shall be mounted on the outside of the case. The receptacle shall be gasket-mounted and recessed, and shall be covered with a device for protection when not in use. The thermostat shall be bellows type, and the bulb shall be located in the bottom of the incubator cavity and protected by a metal shield. Recesses shall be provided in the front and rear of the control panel to house the electrical cord and container for accessories and repair parts.

3.2.5.1 Ground potential. Design and construction shall be such that all external metal parts of control panel shall be at ground potential.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

3.2.5.2 Design and construction of circuits. Design and construction of electrical circuits and components shall include features which shall result in stable operation and which shall assure maxium quality and reliability.

3.3 Accessories, repair parts and tools. The following accessories, repair parts and tools shall be furnished with the kit:

One sampling cup made of corrosion-resisting steel 0.0007 to 0.017 inch thick, with a capacity of 250 ml. (cc.) and graduated as shown in the drawing.

One screwcap, polyethylene plastic bottle of 4-ounce capacity.

One polypropylene plastic bottle with phenol-formaldehyde screwcap of 8-ounce capacity.

One corrosion-resisting steel (see 3.1.1) straight forceps (tweezers) with flattened tips, of 4-1/2-inch nominal length.

One 7-foot, 10-ampere, 3-conductor cord (one ground) No. 18 AWG Type SVT or SV conforming to UL 62 terminating in a 3-conductor (male) plug at one end, and a female "twist-lock" receptacle at the other end conforming to W-C-596 (F), style 41 (closed), suitable for mating with the recessed socket in case conforming to W-C-596 (S) (closed), 2-prong plug adapter to be furnished).

One red wax pencil for glazed surface marking conforming to SS-P-196.

One translucent, flexible, plastic container holding one each of the following, unless otherwise indicated:

a. Lamp, indicator, bayonet type, 14.4 volts, G 3-1/2, 53 clear with red indicating lens.

b. Fuses, cartridge, AGC5, 5 amperes, 250 volts, 5 each.

c. Adapter, plug, continental.

d. Allen wrench for control knobs.

e. Battery adapter, "Alligator" clip assembly.

f. Envelope contining two nylon lockwheels, two No. 4-40 thread by 1/4-inch long flathead screws, two Allen setscrews, and one Allen wrench for setscrews.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

3.4 Finish.

3.4.1 Case. All surfaces of the case shall have a smooth finish.

3.4.2 Corrosion-resisting steel components. The vacuum filter flask (beaker), the graduated sampling cup and the funnel shall have a No. 7 commercial finish (mirror). The control metal plate shall have a satin finish.

3.4.3 Supporting base. The aluminum supporting base of the filter holder assembly shall have a clear, anodized, satin finish.

3.4.4 Aluminum edgings. Aluminum tongue and groove edgings shall have a clear, anodized, satin finish.

3.4.5 All metal components and parts. All metal components, unless otherwise specified, shall be suitably coated or treated to resist corrosion.

3.5 Identification marking.

3.5.1 Case. Each case shall have a pressure-sensitive nameplate securely mounted, and permanently and legibly marked with the following information:

Item Identification
National Stock Number (NSN)
Manufacturer's name or registered trademark
Contract or purchase order number
"U.S."

3.5.2 Metal control plate. The metal plate mounted on the control section on the inside of the case shall be marked as shown in the drawing

3.5.3 "Caution" plate. A pressure-sensitive plate shall be securely mounted on the outside of the fiberglass case alongside the receptacle and shall be permanently and legibly marked in red, with the following or a similar legend:

"CAUTION: Set voltage selector switch prior to plugging in power cord."

3.5.4 Metal syringe. Instructions for use shall be clearly and legibly marked along the lengthwise surface of the barrel of each metal syringe.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

3.6 Service data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for installation, operation and maintenance and lists of component parts. The manual shall comply wtih the requirements of MIL-M-7298. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the contracting officer, via the cognizant Government inspector, two preliminary copies of his proposed technical manual, for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximun period of 30 days for review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract or purchase order. In addition, the contractor shall furnish 15 copies of the approved technical manual to the contracting officer for distribution. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of contract and propose to furnish manuals identical with those previously accepted, the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the Contracting Officer. Waiver of submission of 15 additional copies shall be at the option of the Government.

3.7 Operating instructions. A card containing detailed operating instructions including requirements for power source, temperature adapters, etc., shall be furnished. These instructions shall be printed on white paper which shall be sealed between clear, plastic sheets.

3.8 Literature. Literature in the form of a booklet shall be furnished with each kit describing in detail procedures to be followed in the use of the test kit for the purpose of determining the potability of water.

3.9 Workmanship. Workmanship shall be first class throughout. Kit shall be free from defects listed in Table III and other defects which detract from its appearance or may impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requriements as specified herein. Except as otherwise specified in the contract or purchase order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.3 Certificate of quality. Certificate of quality, supplied by the manufacturer of the fiberglass and corrosion-resisting steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with Table I and MIL-STD-105.

Table I Sampling for examination

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.5
Minor defects	II	4.0
For dimensional examination	S-1	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with Table II and MIL-STD-105. Acceptance number shall be zero.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

Table II Sampling for tests.

Component (unit of product)	Characteristic	Requirement	Test procedure	Inspection level
End item	Incubator temperature test	3.2.2	4.4.1	S-1
End item	Operation test	3.2.5	4.4.2	100%
Funnel assembly	Leakage test	3.2.4.1	4.4.3	S-3
End item	Stability test	3.2.1	4.4.4	S-1
End item	Rain-and splash-proof test	3.2.1	4.4.5	S-1

4.3 Examination. The kit, with components, shall be examined to determine compliance with all requirements contained in this specification. Examination shall be in accordance with, but shall not be limited to, the defects listed in Table III.

Table III Classification of defects

Categories	Defects
Major	
101	Component missing.
102	Instruction card missing, illegible and incomplete.
103	Literature booklet missing.
104	"O" rings not free of cuts, cracks and tears.
105	Wick not set securely in groove.
106	Supporting base not stable on flat, smooth surface.
107	Vacuum filter flask not stable on flat, smooth surface.
108	Plastic feet at bottom of support base not secure.
109	Pin in piano hinge not secure (slides out from hinge during use).
110	Items in holding device (cover insert) not held securely in plate.
111	Plastic lid for incubator cavity not clear; not free of cracks, chips, checks, crizzles and warp.
112	Close fit not provided with edges of lid and mouth edges of cavity.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

Table III Classification of defects (cont'd)

Categories	Defects
Major	
113	Lid not held securely by spring latch.
114	Hardness on case not securely attached.
115	Marking on metal control plate incomplete and missing (see 3.5.2 and drawing).
116	Plastic holding device (cover insert) not secured in place; not free of cracks, holes or deformation.
117	Plastic bottles with screwcaps not free of deformation, cuts, cracks, holes and tears.
118	Thermometer not secured in place; not free of chips, cracks and tears.
119	Markings and graduation on the thermometer missing, incomplete and illegible.
120	Wires of rack not welded or brazed at each point of contact.
121	Wires of rack not free of burrs and sharp edges.
122	Plastic liner in body of case not free of deformation, cracks, cuts and holes.
123	Fiberglass plastic case not free of cracks, holes and indentations.
124	Fiberglass container not free of rocking, when placed on a flat, smooth surface.
125	Flexible plastic container (containing repair parts, etc.) not free of deformation, cracks and tears.
126	Item identification on case illegible, incomplete and missing.
127	Plate bearing caution legend missing, incomplete, illegible and not marked in red.
Minor	
201	"O" rings loose in grooves.
202	Plastic rollers (wheels) not secure.
203	Plastic lid for incubator cavity not free of scratches, mars, dents and pitting.
204	Identification marking on case missing, incomplete and illegible.
205	Welding of wires in rack not sound and clean.
206	Plastic bottles not free of dirt and foreign matter.
207	Plastic knobs on metal control plate not secure.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

4.3.2 Dimensional examination. The unit kit, including components, shall be examined for defects in dimensions. Any dimension not within the tolerance specified herein shall be classified as a defect.

4.3.3 Packaging inspection. The inspection of the unit package, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Incubator temperature test. The kit shall be placed in a conditioned chamber maintained at -30 F. $\pm$ 1°. The incubator cavity shall contain the rack with 24 disposable Petri dishes. Each Petri dish shall contain a nutrient-saturated filter disk and an absorbant pad. For 23 of the Petri dishes, a filter disk and 2 ml. of distilled water may be substituted for the nutrient-saturated filter disk. The kit shall be operated in accordance with the manufacturer's instructions. Test shall be repeated at 98.5°F. + 0°, -2°F. For -30°F. test for sample size of two, one kit shall be run at 6 volts and one at 115 volts. Testing at other voltages specified in 3.2.5 is not required for this sample size. If more than two kits are tested (see Table II), -30°F. testing of additional samples shall be done at voltages other than 6 and 115, and no voltage shall be repeated until all voltages specified in 3.2.5 have been covered. 98.5°F. test shall be run at any of the voltages specified in 3.2.5 except for the one which was used for the -30°F. testing. No more than one test voltage is required for each sample for the 98.5°F. test.

4.4.2 Operation test. Each sample shall be operated at room temperature on each of the specified voltages to determine that the unit operates properly.

4.4.3 Leakage test. Filter holder assembly shall be secured in vacuum filter flask and shall be connected to a vacuum pump capable of obtaining a vacuum of 25 inches of mercury. Connection shall be by means of a tube which fits into the evacuator hole without permitting leakage. After it has been completely saturated with water, a filter disk with a pore size of 0.45 microns shall be placed in the filter holder. The vacuum pump shall be turned on and allowed to run until a vacuum of 25-26 inches of mercury is obtained inside the filter holder assembly. Connecting tube shall be clamped so that there is no flow of air through the evacuator hole, and filter pump shall be turned off. After 20 minutes the vacuum inside the filter holder assembly shall be measured.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

4.4.4 Stability test. A fully equipped unit kit shall be set on a firm, smooth, level surface and the cover shall be opened to its maximum 90-degree position. The incubator lid shall be opened and the rack containing 24 Petri dishes (empty) shall be removed from the cavity. Any lifting from the surface of the front end of the case shall be carefully noted.

4.4.5 Rainproof and splashproof test. A fully equipped water testing kit shall be set on a suitable horizontal surface. A water spray device shall be located 1 to 2 feet directly above the cover of the kit and shall be capable of passing 4000 ml./minute, ± 20 percent, of water on to the top surface of the case. Prior to performance of the test, the inside of the kit shall be checked to insure that all internal surfaces are dry. Kit shall then be properly closed before testing. The spray device water shall be turned on, so that a minimum of 8000 ml. of water shall be allowed to fall on the top surface of the kit. At the end of testing, outside surface of case shall be wiped dry of any surplus water. Cover shall be opened to a height of 2 inches and the inside of the case shall be carefully observed for evidence of leakage. Care shall be exercised to insure that any water present on the cover or sealing surfaces shall not drip inside upon opening the lid. The lid shall be completely opened and all surfaces on the inside of the case shall be examined for evidence of water. Any presence of water inside the case resulting from the test spray shall be cause for rejection.

5. PACKAGING

* 5.1 Packaging. Packaging shall be Level A or C as specified (see 6.1).

5.1.1 Unit of issue. One water testing kit, as specified, constitutes one unit of issue.

5.1.2 Level A. Components comprising the test kit shall be positioned and secured within their respective compartments in the carrying case. The case shall be closed and the closure secured.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

* 5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

5.2.1 Level A. Each unit shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-601, overseas type, or PPP-B-636, class weather-resistant. The exterior container shall be designed with suitable interior fittings or shall contain sufficient cushioning material to prevent marring, scratching or damage to the case and contents. Closure of box shall be as specified in the appendix of the applicable box specification.

* 5.2.1.1 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof as specified in the appendix of PPP-B-636.

* 5.2.2 Level B. Each unit shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-601, domestic type, or PPP-B-636, type CF, class domestic. Closure of boxes shall be in accordance with appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size and kind commonly used for the purpose constructed so as to insure acceptance and safe delivery called for in the contract or purchase order.

5.3 Marking. Each exterior container shall be marked as specified in MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

a. Title, number and date of this specification.
b. National Stock Number (NSN).
c. Selection of applicable level of packaging and packing (see 5.1 and 5.2).

6.2 The Petri culture dish referred to in 3.2.2 and 3.3 is NSN 6640-00-299-8689, DISH, CULTURE, PETRI, Top and Bottom, Complete, Disposable, 100s.

6.3 The Endo Broth referred to in 3.2.1 and 3.2.2 is NSN 6505-00-687-8436, ENDO BROTH MEMBRANE FILTER, 2 cc., 24s.

6.4 The filter disk referred to in 3.2.2 is National Stock Number 6640-00-299-8692, DISK, FILTERING, MICROPOROUS, Hydrosol Type, 47 mm. Diameter, 100s.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

6.5 This specification does not include all types, classes, grades or sizes of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the type procured to meet Military requirements.

6.6 This specification covers the following item as listed in the Federal Supply Catalog:

National Stock Number	Item identification
6665-00-682-4765	WATER TESTING KIT, BACTERIOLOGICAL

6.7 The margins of this specification are marked with an askerisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6665 - 0379

☆U.S. GOVERNMENT PRINTING OFFICE:1976-703-020/555

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-W-36258D WATER TESTING KIT, BACTERIOLOGICAL

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

☆ GPO: 1974— 603-680/ 6126 2-1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:17Z
Check the source to verify that this is the current version before use.