MILITARY SPECIFICATION

ANESTHESIA KIT, LARYNGOTRACHEAL, DISPOSABLE

This specification is approved for use by the Defense Personnel Support Center, Defense Logistics Agency, and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a disposable, laryngotracheal anesthesia kit for the administration of topical anesthesia to the interior of the larynx and trachea.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bid or request for proposal form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-636 - Boxes, Shipping, Fiberboard.

MILITARY

MIL-S-36586 - Sterilization Test Strip Set, Bacterial Spore.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:36Z
Check the source to verify that this is the current version before use.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

24395 - Anesthesia Kit, Laryngotracheal.

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bid or request for proposal shall apply.

U. S. PHARMACOPEIAL CONVENTION, INC. (USP)

Pharmacopeia of the United States

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE,
FOOD AND DRUG ADMINISTRATION (FDA)

Federal Food, Drug, and Cosmetic Act and Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, DC 20204.)

(Technical society and technical association specifications and standards are generally available for reference from libraries. They are also distributed among technical groups and using Federal agencies.)

3. REQUIREMENTS

3.1 Materials.

3.1.1 Vial injector. The barrel shall be made of polypropylene. The metal cannula in the barrel shall be made of an austenitic stainless steel. The plastic cannula attached to the barrel shall be made of polycarbonate resin.

3.1.2 Vial. The vial shall be made of USP Type I borosilicate glass. The rubber stopper (plunger) shall be inert to the contained medication.

3.1.3 Medication. The medication contained in the vial shall be sterile Lidocaine Hydrochloride Solution, USP, 4%. It shall not contain any preservative.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:36Z
Check the source to verify that this is the current version before use.

3.2 Style and design. The laryngotracheal anesthesia kit shall consist of a sterile, plastic dispensing injector (vial injector) with integral stainless steel cannula, a 5 ml glass vial containing 4 ml of sterile Lidocaine Hydrochloride Solution, USP, 4% and instructions for use. The stainless steel cannula shall be ground to a point at the proximal end to facilitate penetration of the rubber stopper.

3.2.1 Dimensions. Dimensions shall be as indicated in DPSC Drawing No. 24395.

3.3 Security of assembly. There shall be no separation of the plastic cannula from the vial injector when tested as specified in 4.4.1.

3.4 Ease of ejection. There shall be a complete ejection of the Lidocaine within three (3) seconds when tested as specified in 4.4.2. There shall be no leakage at the junction of the plastic cannula and vial injector or past the rubber stopper when tested as specified in 4.4.2.

3.5 Sterility. The Lidocaine and vial injector shall be sterile when tested with a biological indicator as specified in 4.4.3.

3.5.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include.

a. Item identification.
b. National Stock Number (NSN).
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.).
h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of a laboratory or a designated qualified alternate.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Defense Personnel Support Center, Directorate of Medical Materiel, Technical and Quality Assurance Division, ATTN: DPSC-ATQ, 2800 S. 20th Street, Philadelphia, PA 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:36Z
Check the source to verify that this is the current version before use.

3.5.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and sterilized simultaneously in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.2. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

3.6 Instructions. Each laryngotracheal anesthesia kit shall be supplied with instructions for dispensing the anesthetic and precautions to be observed in using the kit.

3.7 Expiration dating period. The anesthesia kit shall have an expiration period of not less than three (3) years.

3.7.1 Delivery. Not more than six (6) months shall have elapsed from the time of manufacture to the date of delivery to the Government.

3.8 Workmanship. Workmanship shall be first class throughout. The kits shall be free from defects that detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three (3) years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:36Z
Check the source to verify that this is the current version before use.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the Lidocaine, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/ material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in table I. The unit of product for examination purposes shall be one kit.

TABLE I. Sampling for examination.

Type defect	Inspection level	AQL (percent defective)
Major	S-3	1.0
Minor	S-3	1.5

4.2.2 For tests. Sampling for tests shall be in accordance with MIL-STD-105 and as indicated in table II. The unit of product for test purposes shall be one kit. Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

TABLE II. Sampling for tests.

Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Dimensions	3.2.1	--	S-1	1.5
Security of assembly	3.3	4.4.1	S-1	1.5
Ease of ejection	3.4	4.4.2	S-1	1.5
Sterility	3.5	4.4.3	S-2	1.0

4.3 Examination.

4.3.1 Examination of the laryngotracheal anesthesia kit. The laryngotracheal anesthesia kit shall be examined to determine compliance with all the requirements of this specification. Examination shall be conducted as specified in tables I and II. Examination, however, shall not be restricted to defects listed in table III. Any laryngotracheal anesthesia kit in the sample containing one or more defects shall be rejected, and if the number of defective laryngotracheal anesthesia kits in any sample exceeds the acceptance number for that sample, the lot represented by that sample shall be rejected.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:36Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects.

Categories	Defects
Major	
101	Presence of burrs, flashing on plastic cannula.
102	Vial leaking in kit.
103	Expiration date of Lidocaine less than 2 years and 6 months from date of delivery.
104	Vial or component of kit, other than instructions, missing.
Minor	
201	Instructions missing.

4.3.2 Packaging inspection. The inspection of the unit package, packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Security of assembly. A tensile load of 25 pounds shall be applied for one minute to the vial injector to observe any separation at the plastic cannula-barrel junction.

4.4.2 Ease of ejection. The cap shall be removed from the vial and the vial screwed into the vial injector. The vial shall be forcefully compressed into the vial injector to discharge the Lidocaine. This compression shall be accomplished in less than three (3) seconds.

4.4.3 Sterility test. The contents of the laryngotracheal anesthesia kit shall be tested for sterility in accordance with the method specified in the USP for liquids and parenteral devices using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in 4.4.3.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for Type I sterilization and 30 for Type II sterilization including inoculated carrier and product test units. The inspection level shall be as indicated in Table II.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:36Z
Check the source to verify that this is the current version before use.

4.4.3.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55° - 65°C
Dry Heat	B. stearothermophilus, or	55° - 65°C
	B. subtilis var. niger (or globigii)	30° - 35°C
Gas	B. subtilis var. niger (or globigii)	30° - 35°C
Radiation	B. pumilus	30° - 35°C

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

4.4.3.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used, shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.3 and 4.4.3.1. The method of sterilization will control the one organism in the spore strip to be tested for, as shown in 4.4.3.1.

When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

4.4.3.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein, shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

4.4.3.4 Product testing. Testing for sterility of the product after receipt by the Government shall be without a biological indicator and in accordance with the USP.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C, as specified (see 6.2).

5.1.1 Unit of issue. One package containing twenty-four individually packaged kits constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Individual package. Components comprising each kit shall be clean, dry, neatly arranged and packaged sterile, in a sealed, tamperproof, peel-open type container, capable of maintaining sterility of contents unless the container is damaged or opened.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:36Z
Check the source to verify that this is the current version before use.

5.1.2.2 Unit package. Twenty-four individual packages shall be packaged in a box conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.1.3 Level C. Components comprising each kit shall be clean, dry, neatly arranged and packaged sterile, in a sealed, peel-open type container, capable of maintaining sterility of contents unless the container is damaged or opened. Twenty-four individually packaged kits shall be further packaged in a manner that will afford adequate protection against deterioration and physical damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.2).

5.2.1 Level A. Two unit packages shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-636 class weather-resistant. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

5.2.2 Level B. Two unit packages shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The packaged kits shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the carrier's rules and regulations.

5.3 Marking.

5.3.1 Individual container. In addition to any commercial markings, each container shall be marked in accordance with the requirements of the Federal Food, Drug, and Cosmetic Act. Marking shall also include the item identification, lot or control number, expiration date and the legend or similar statement "STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED".

5.3.2 Unit package. Each unit package shall be marked in accordance with MIL-STD-129. Type I shelf life marking, as specified in MIL-STD-129, shall be shown. The lot or control shall be shown. The method of sterilization (G-Gas, R-Radiation, S-Steam, C-Cold, D-Dry Heat) shall be shown following the level of packaging and date of packaging. Example: "A 1/78-G". Final sterilization shall be on or after indicated packaging date.

5.3.3 Shipping container. Shipping containers shall be marked as specified in MIL-STD-129. Type I shelf life marking, as specified in MIL-STD-129, and the lot or control number, shall be shown. The method of sterilization shall be shown following the levels of packaging and packing and date packed. Example: "A/B 1/78-G".

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:36Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Intended use. The laryngotracheal anesthesia kit is used for topical anesthesia of the laryngotracheal tissues prior to intubation.

6.2 Ordering data. Procurement documents should specify the following:

a. Title, number and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
c. Size or type as applicable (see 2.1).

6.3 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6515-01-003-9783	ANESTHESIA KIT, LARYNGOTRACHEAL, Disposable, 24s

Preparing activity:
DLA-DM

Project No. 6515-2107

☆U.S. GOVERNMENT PRINTING OFFICE: 1970 — 703-122/1627

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:36Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-A-37842(DLA-DM) ANESTHESIA KIT, LARYNGOTRACHEAL, DISPOSABLE

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

☆ GPO 1974—603-660/6126 2-1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:36Z
Check the source to verify that this is the current version before use.