MILITARY SPECIFICATION

PASTE, PERIODONTAL PROTECTIVE, Surgical, Soft Dressing

This specification is approved for use by the Defense Personnel Support Center, Defense Logistics Agency, and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a soft, protective, periodontal paste.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bid or request for proposal form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-566 - Boxes, Folding, Paperboard.
PPP-B-585 - Boxes, Wood, Wirebound.
PPP-B-601 - Boxes, Wood, Cleated-Plywood.
PPP-B-621 - Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636 - Boxes, Shipping, Fiberboard.
PPP-B-676 - Boxes, Setup.
PPP-C-186 - Containers, Packaging and Packing for Drugs, Chemicals, and Pharmaceuticals.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:39Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-L-10547 - Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

(Copies of specifications and standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bid or request for proposal shall apply.

U.S. PHARMACOPEIAL CONVENTION, INC.

The United States Pharmacopeia (USP)

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

The National Formulary (NF)

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

AMERICAN DENTAL ASSOCIATION (ADA)

Ad Interim Recommended Standard Practices for Toxicity Tests on Dental Materials.

(Application for copies should be addressed to the Council on Dental Materials and Devices, American Dental Association, 211 East Chicago Avenue, Chicago, IL 60611.)

3. REQUIREMENTS

3.1 Materials. Materials shall be of the highest quality throughout and shall be suitable in all respects for the purpose intended. Where applicable (see tables I and II and 3.2.3), material shall be in accordance with the requirements of the United States Pharmacopeia (USP) or the National Formulary (NF). Any material used in the periodontal paste which is not included in the USP or NF shall be of a quality suitable for dental use.

3.2 Composition. Unit of issue shall consist of a base and accelerator which shall be pastes and a retarder which shall be a liquid. When the pastes are mixed in accordance with manufacturer's instructions, they shall form a mass with a soft plastic consistency which adapts easily to teeth and tissue. Mixture shall resist dislodgement by the movement which is associated with eating and speaking and shall be easy to apply to and easy to remove from a wound. Use of retarder shall increase the working time and shall not change the other characteristics of the mixture. No component shall contain eugenol, asbestos, or cotton fibers.

3.2.1 Base. Base shall consist of the ingredients listed in table I.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:39Z
Check the source to verify that this is the current version before use.

TABLE I. Base ingredients.

Ingredient	Specification
Rosin	USP
Cellulose Nitrate	USP
Natural gums and wax	---
Fatty Acid	---
Chlorothymol	---
Spearmint Oil	---
Zinc Acetate	USP
Alcohol	USP
Purified Water	USP

3.2.2 Accelerator. Accelerator shall consist of the ingredients listed in table II.

TABLE II. Accelerator ingredients.

Ingredient	Specification
Zinc Oxide	USP
Vegetable Oil	---
Chlorothymol	---
Colloidal Silicon Dioxide	NF
Magnesium Oxide	USP
Synthetic Resin	---
Coumarin	---
Parachlorometaxylenol (see 6.3)	---
Red Dye	---

3.2.3 Retarder. Retarder shall consist of USP glycerine.

3.3 Color. The base and accelerator shall be distinctly different in color. Resultant mixture shall be pink.

3.4 Performance. When tested as specified in 4.4.1, mixing time shall be 45 ± 15 seconds. When no retarder is used, peel time shall be 3 ± 1 minutes and final set time shall be 15 ± 5 minutes. When no retarder is used, the working time shall be a minimum of 6 minutes. Working time is defined as final set time minus peel time. Instructions shall contain information for the extension of working time by use of a retarder. When retarder is used, minimum working time shall be increased to at least 15 minutes.

3.5 Instructions. Adequate and accurate instructions for proportioning and manipulating pastes shall accompany each unit package.

3.6 Immediate container. Tubes for base and accelerator shall conform to group B, class 1, type a, closure A tubes of PPP-C-186. Two keys (one for each tube) shall be supplied for forcing the paste from the tubes. Bottle for retarder shall be glass with a screw cap closure.

3.6.1 Capacity. Net weight of accelerator shall be at least 90 grams; net weight of base shall be at least 90 grams; and net weight of retarder shall be at least 4 grams. Volume after mixing a tube of accelerator with a tube of base shall be at least 125 milliliters.

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:39Z
Check the source to verify that this is the current version before use.

3.7 Acceptance. Contractor shall furnish proof that offered periodontal paste has been accepted by the Council on Dental Materials and Devices of the American Dental Association. Compliance with this acceptance requirement does not absolve the contractor from complete compliance with the requirements of this specification.

3.8 Council on Dental Materials and Devices requirements. Periodontal paste shall meet the marking, toxicity, irritation, and inflammatory reaction requirements of the Ad Interim Recommended Standard Practices for Toxicity Tests on Dental Materials, type II, class I, periodontal packs.

3.9 Shelf life. The shelf life of all the components of the periodontal paste shall be no less than 18 months from the date of manufacture. The inspection test date shall be 18 months from the date of manufacture.

3.9.1 Date of delivery. Not more than 2 months shall have elapsed from the date of manufacture of periodontal paste components to the date of delivery to the Government.

3.9.2 Date of manufacture. The date of manufacture shall be the date when the component is placed within the immediate container.

3.10 Marking. Each tube and bottle shall be marked with the commercial item name, manufacturer's name or registered trademark, and a lot, control, or batch number which shall refer to the manufacturer's records for the particular lots or batches. Bottle shall be marked "Retarder", accelerator tube shall be marked with a "1", and base tube shall be marked with a "2".

3.11 Workmanship. Workmanship shall be first class throughout. Units shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component or material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:39Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table III. Unit of product for examination purposes shall be the unit of issue.

TABLE III. Sampling for examination.

	Inspection level	AQL (percent defective)
For visual examination		
Major A	S-2	2.5
Major B	II	2.5
Minor	II	4.0
For weight and volume examination	S-1	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MII-STD-105 and table IV. The unit of product for test purposes shall be the unit of issue and the acceptance number shall be zero.

TABLE IV. Sampling for tests.

Characteristic	Requirement	Test procedure	Inspection level
Reaction time	3.4	4.4.1	S-2

4.3 Examination. The periodontal paste shall be examined to determine compliance with all requirements of this specification. Examination shall include, but shall not be limited to defects listed in table V.

TABLE V. Classification of defects.

Categories	Defects
Major A:	
101	Accelerator and base pastes not different colors.
102	Pastes not free of dirt and grit.
103	Liquid (retarder) not free of dirt and grit.
104	Pastes not free of lumps.
105	Pastes not smooth and homogeneous.
Major B:	
110	Unit package does not contain a base, accelerator, and retarder.
111	Unit package does not contain instructions.
112	Instructions not complete, legible, and correct.
113	Tube not collapsible.
114	Tube not cylindrical.
115	Tube not free of seams.
116	Tube not free of holes.
117	Closure not a rigid plastic screwcap.

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:39Z
Check the source to verify that this is the current version before use.

TABLE V. Classification of defects. - Continued

Categories	Defects
118	Closure does not have a liner.
119	Closure not tightly closed.
120	Tube not free of leakage
Minor:	
201	Unit package does not contain two keys.
202	Base and accelerator cannot be differentiated by marking on tubes.

4.3.1 Weight and volume. Components shall be examined for compliance to weight requirements. Any component which does not meet weight specifications shall be classified as a defect. In addition, if volume of mixture of accelerator and base does not meet specifications, it shall be classified as a defect.

4.3.2 Packaging inspection. The inspection of the packaging, packing, and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Reaction times. Accelerator and base shall be mixed in accordance with package instructions. Mixing shall be considered complete when mixture is a homogeneous pink color. Mixture shall be placed on nonabsorbent parchment paper and shall be left on paper until it can be removed without sticking. After removal from the paper, ropes shall be formed from the mixture. Peel time is the period of time from the beginning of the mixing to the time when mixture can be removed from the paper without sticking. Final set time is the period of time from the beginning of the mixing to the time when 3 or 4 millimeter diameter ropes break. Working time is the difference between final set time and peel time.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C, as specified (See 6.2).

5.1.1 Unit of issue. One package containing one tube of base, one tube of accelerator, one bottle of retarder, and instructions for use, as specified, constitutes one unit of issue.

5.1.1.1 Level A.

5.1.1.1.1 Unit packaging. Each unit of issue shall be packaged in a box of appropriate size and constructed in accordance with PPP-B-566 or PPP-B-676. Box design shall include interior fittings or shall contain sufficient cushioning material to insure against shifting of, or damage to the contents. Closure of the box shall be in accordance with the appendix of the applicable box specification.

5.1.1.2 Level C. Each unit of issue shall be packaged in a manner that will afford adequate protection against deterioration and physical damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B. or C. as specified (see 6.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:39Z
Check the source to verify that this is the current version before use.

5.2.1 Level A. Twelve (12) units of issue shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard boxes shall be waterproof as specified in the appendix of PPP-B-636.

5.2.2 Level B. Twelve (12) units of issue shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The packaged paste shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carrier's rules and regulations.

5.3 Marking. Each unit, intermediate package, and exterior container shall be marked as specified in MIL-STD-129. Marking shall include the lot (control) number. Type II shelf life marking, as specified in MIL-STD-129, shall be shown.

6. NOTES

6.1 Intended use. The soft, periodontal protective paste is intended to cover and protect the surface of wounds caused by periodontal surgery.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents:

a. Title, number, and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 Parachlorometaxylenol is available from the Ferro Corporation, Ottawa Chemical Division, as "Ottosept Extra".

6.4 This specification covers the following item listed in the Federal Supply Catalog:

National Stock Number	Item Identification
6520-01-047-3577	PASTE, PERIODONTAL PROTECTIVE, Surgical, Soft Dressing

Preparing activity:
DLA-DM

Project No. 6520-1502

Source: https://assist.dla.mil -- Downloaded: 202640-21T11:39Z
Check the source to verify that this is the currentversion before use.