MILITARY SPECIFICATION

ROD, INSECT BAR

This specification is approved for use by all Departments and Agencies of the Department of Defense

1 SCOPE

1 1 Scope This specification covers an aluminum rod with a crossbar, to support an insect net or an intravenous bottle or bag, for use with an adjustable hospital bed.

2 APPLICABLE DOCUMENTS

2 1 Issues of documents The following documents of the issue in effect on date of invitations for bid or requests for proposals, form a part of this specification to the extent specified herein

SPECIFICATIONS

FEDERAL

TT-E-529	- Enamel, Alkyd, Semi-gloss
TT-P-1757	- Primer Coating, Zinc Chromate, Low Moisture Sensitivity
PPP-B-566	- Boxes, Folding, Paperboard
PPP-B-585	- Boxes, Wood, Wirebound
PPP-B-601	- Boxes, Wood, Cleated-Plywood
PPP-B-621	- Boxes, Wood, Nailed and Lock-Corners
PPP-B-636	- Boxes, Shipping, Fiberboard
PPP-B-676	- Boxes, Setup.
PPP-T-60	- Tape Packaging, Waterproof.

MILITARY

MIL-C-5541	- Chemical Conversion Coatings on Aluminum and Aluminum Alloys
MIL-L-10547	- Liners, Case, and Sheet, Overwrap, Water-Vaporproof, or Waterproof, Flexible.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 7210

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.

STANDARDS

FEDERAL

FED. TEST METHOD STD NO. 151	- Metals Test methods
FED-STD-595	- Color

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

25024	- Bed, Folding, Hospital, Field

(Copies of specifications standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer)

3 REQUIREMENTS

3 1 First article When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4 1 2 1 and 6 2).

3 2 Material Except as hereinafter specified, materials shall conform to the applicable portions of the specifications, standards, and drawing listed herein Materials shall be subjected to the tests and shall meet the requirements specified in these documents The metals shall conform to the chemical and physical requirements of the applicable specifications Materials, not specifically designated herein, but used in the fabrication of the rods, shall be suitable for the purpose intended

3 3 Design and construction The rod shall be constructed in conformance with the applicable portions of DPSC Drawing NO 25024 and as specified herein (Sheet No 9 of the drawing and pages 7 8 and 10 of the Bill of Materials thereto contain all references to the rod)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.

3.4 Finish. All surfaces shall be thoroughly cleaned and chemically degreased prior to chemical treatment. All aluminum parts shall be finished with a chemical film conforming to MIL-C-5541. After degreasing and chemical treatment, a zinc-chromate primer conforming to TT-P-1757 shall be applied to all metal parts. The primer coat shall be followed by two baked-on spray coats of synthetic, semi-gloss, enamel in accordance with TT-E-529 Color shall be olive drab, No 24708 of FED-STD-595. The resultant finish shall be hard, smooth, and highly resistant to abrasion, chipping and cracking. When tested as specified in 4.5.2, enamel shall not peel off to expose any of the following

1. Any area which has any dimension greater than 1/8 inch
2 More than one area exceeding 1/16 inch in any dimension.
3 More than five areas of any size

3.5 Identification marking Each unit shall be marked permanently and legibly with the contract or purchase order number, the name or registered trademark of the manufacturer, and the National Stock Number. A decal, firmly attached with a durable adhesive, shall be acceptable.

3 6 Workmanship Workmanship shall be first class throughout. Item shall be free from defects which detract from its appearance or impair its serviceability All burrs and sharp edges shall be removed from the item

4 QUALITY ASSURANCE PROVISIONS

4 1 Responsibility for inspection Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

4.1.2.1 Classification of inspection The inspection requirements specified herein are classified as follows

a First article inspection.
b Quality conformance inspection.

4.1.2.1.1 First article inspection. First article inspection shall be performed on the rod when a first article sample is required (see 3.1) This inspection shall include examination of 4.4 and tests of 4.5

4.1.2.1.2 Quality conformance inspection Quality conformance inspection shall include the examination of 4.4 and tests of 4.5

4.1.3 Certificates of quality Certificates of quality, supplied by the manufacturer of the metals may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government, and provided also that the test methods and the results obtained comply with those described in this specification The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements

4.2 Inspection lot

4.2.1 Raw materials An inspection lot for raw material inspection shall be all material produced under the same conditions and controlled to the degree that uniformity of product within the designated lot can be assured Lot identification of the raw material by the manufacturer is one method of control, as cited above, which is considered acceptable.

4.2.2 End product An inspection lot for inspection of end product shall consist of all rods of the same type produced under essentially the same conditions and offered for inspection at one time

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.

4.3 Sampling.

4.3.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I. The unit of product for sampling purposes shall be one rod.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and Minor combined)	II	4.0
For dimensional examination	S-1	2.5

4.3.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II, for an acceptance number of zero.

TABLE II. Sampling for tests

Component (Unit of product)	Characteristic	Requirement	Test procedure	Inspection level
Rod	Enamel	3.4	4.5.2	1/

1/ Two rods shall be selected randomly from each day's enameling

4.3.3 Raw materials. A sufficient quantity of each lot of each metal used in the manufacture of the rod shall be selected for chemical analysis.

4.4 Examination. The rod shall be examined to determine compliance with all requirements contained in this specification.

4.4.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table III. Examination shall not be restricted to the classified possible defects listed below.

TABLE III. Classification of defects

Categories	Defects
Major	
101	Any component part missing, broken, or otherwise unusable
102	Any sharp edge or burr.
103	Components not properly assembled and aligned
104	Any joint not secure and as specified.
Minor	
201	Identification marking missing, incomplete, or illegible.
202	Finish chipped, cracked, scratched, or rough.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.

4.4.2 Dimensional examination. The rod shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4 4 3 Packaging inspection. The inspection of the unit package, intermediate package, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129

4.5 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.5 1 Chemical composition Four ounces from each lot of material shall be tested in accordance with FED TEST METHOD STD 151.

4.5.2 Enamel adhesion Unless otherwise specified a minimum of two completed units from each day's finishing shall be tested. Adhesion or the painted test specimens should generally be determined within a maximum of 72 hours drying Press a 2 inch length of a somewhat longer piece of water-resistant, pressure sensitive adhesive tape (3/4 inch width) conforming to PPP-T-60. (transparent), firmly onto a flat or cylindrical surface of the item, rubbing out all air bubbles under the tape. Allow approximately 10 seconds for the test area to return to room temperature. Grasp a free end of the tape and at a rapid speed strip it from the item by pulling the tape back upon itself at 180 degrees (in such a manner that the tape is folded back to back during the procedure) Observe for bared areas where the enamel is removed. Disregard flecks of enamel or tape where the underlying metal or phosphate coating is not visibly exposed If any test specimen indicates failure all items processed since last acceptance shall be rejected and corrective action taken.

5 PACKAGING

5.1 Packaging Packaging shall be level A or C, as specified (see 6.2)

5 1 1 Unit of issue Each insect bar rod, as specified, constitutes a unit of issue

5 1.2 Level A

5 1 2.1 Unit package. Each insect bar rod shall be collapsed and folded to its smallest dimension and wrapped in flexible corrugated fiberboard and the wrap secured with kraft gummed reinforced tape As an alternate each unit shall be packaged in a box conforming to PPP-B-566 PPP-B-636 class domestic or PPP-B-676 Closure shall be as specified in the appendix of the applicable box specification

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.

5.1.2.2 Intermediate package. Twelve unit packages shall be packaged in a box conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5 1.2 Level C Each collapsed insect bar rod shall be packaged in a manner that will afford adequate protection against physical damage during shipment from the supply source to the first receiving activity

5 2 Packing Packing shall be level A, B or C, as specified (see 6.2)

5 2 1 Level A Four intermediate packages shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-585, class 3, style 3; PPP-B-601, overseas type, PPP-B-621, class 2 or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5 2.1.1 Case liner Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547 Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes Each fiberboard box shall be waterproofed as specified in the appendix of PPP-B-636

5 2.2 Level B Four intermediate packages shall be packed in a shipping container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3, PPP-B-601, domestic type; PPP-B-621, class 1, or PPP-B-636, class domestic Closure shall be as specified in the appendix of the applicable box specification

5.2 3 Level C The packaged insect bar rods shall be packed in shipping containers in a manner to afford adequate protection against damage during direct shipment from the supply source to the first receiving activity These packs shall conform to the carrier's rules and regulations

5.3 Marking Each unit package, intermediate package and shipping container shall be marked as specified in MIL-STD-129.

6 NOTES

6 1 Intended use The rod is intended for use with adjustable bed, NSN 6530-01-015-5189 or 6530-00-935-4184

6 2 Ordering data Purchasers should select the preferred options permitted herein and include the following information in procurement documents.

a Title, number and date of this specification.
b Selection of applicable levels of packaging and packing (see 5.1 and 5.2)
c When a first article is required (see 3.1. 4 1.2 1 and 6.3)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.

6.3 First article. When a first article is required, it shall be tested and approved under the appropriate provisions of the Defense Acquisition Regulation/ Armed Services Procurement Regulation. The contract or purchase order specifies the quantity, procedure for delivery and other specific instructions for inspection and approval of the first article.

6.4 This specification covers the following item(s) appearing in the Federal Supply Catalog.

National Stock Number	Item Identification
7210-00-118-0011	ROD, INSECT BAR, Aluminum

Custodians.
Army - MD
Navy - MS
Air Force - 03

Review activity.
DoD - MB

Preparing activity
DLA-DM

Project No. 7210-0233

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS This form is provided to solicit beneficial comments which may improve this document and enhance its use DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government Fold on lines on reverse side, staple in corner, and send to preparing activity Attach any pertinent data which may be of use in improving this document If there are additional papers, attach to form and place both in an envelope addressed to preparing activity A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed

NOTE This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

DOCUMENT IDENTIFIER *(Number)* AND TITLE
MIL-R-37828 ROD, INSECT BAR

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1 ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID RESTRICTIVE LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW

A GIVE PARAGRAPH NUMBER AND WORDING

B RECOMMENDED WORDING CHANGE

C REASON FOR RECOMMENDED CHANGE(S)

2 REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*

TELEPHONE NO

DATE

DD FORM 1 OCT 76 1426 EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.