MIL-S-37813A
25 July 1978
SUPERSEDING
MIL-S-37813(DLA-DM)
5 December 1977

MILITARY SPECIFICATION

STOPCOCK, INTRAVENOUS THERAPY, 3-WAY, PLASTIC, AYER TYPE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a disposable, plastic, sterile, three-way stopcock (Ayer type), suitable for use in intravenous therapy.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bid or request for proposal form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-585 - Boxes, Wood, Wirebound.
PPP-B-601 - Boxes, Wood, Cleated-Plywood.
PPP-B-621 - Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636 - Boxes, Shipping, Fiberboard.

MILITARY

MIL-L-10547 - Liners, Case, And Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.
MIL-S-36586 - Sterilization Test Strip Set, Bacterial Spore.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:33Z
Check the source to verify that this is the current version before use.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

23311 Stopcock, Intravenous Therapy, 3-Way, Plastic, Ayer Type.

(Copies of specifications, standards, drawings and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bid or request for proposal shall apply.

U. S. PHARMACOPEIAL CONVENTION, INC.

Pharmacopeia of the United States (USP)

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)

ANSI Z70.1 Glass and Metal Luer Tapers for Medical Applications, Dimensions of.

(Application for copies should be addressed to the American National Standards Institute, 1430 Broadway, New York, NY 10018.)

U. S. DEPARTMENT OF HEALTH, EDUCATION AND WELFARE
FOOD AND DRUG ADMINISTRATION (FDA)

Federal Food, Drug, and Cosmetic Act and Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, U. S. Department of Health, Education and Welfare, Washington, DC 20204.)

(Technical society and technical association specifications and standards are generally available for reference from libraries. They are also distributed among technical groups and using Federal agencies.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:33Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material. All components shall be fabricated from suitable rigid plastics, of the best quality, and suitable in all respects for the purpose intended. All materials of the finished stopcock shall be sterile, safe (non-toxic) and pyrogen-free. In addition, the plastics shall meet the requirements of the USP for Class II plastics, when tested as specified in 4.4.3. The total combined quantity of barium, cadmium and lead for each plastic component shall not exceed 5 parts per million (ppm) when tested as specified in 4.4.1.

3.2 Style, design and dimensions. Shall be a three-way plastic stopcock, of the Ayer type, suitable for use in administering and controlling the direction of flow of blood and solutions for simultaneous and/or alternate intravenous therapy. In addition, the stopcock shall be suitable for use in techniques where higher than normal pressures are encountered such as procedures involving the spine. The stopcock shall have three fluid paths, with only two of the three fluid paths integrally connected at any open flow setting. All fluid paths shall be closed to flow when the rotating plug member is positioned between any of the ports at an approximate angle of 45°. The rotating plug member (handle) shall clearly indicate the path that is closed to flow by the marking "OFF" or "CLOSED". The rotating plug shall have a maximum travel of 180°, constrained by stops on the body and rotating plug. Stopcock shall be in accordance with DPSC Drawing 23311.

3.2.1 Stopcock body. The stopcock body shall have two female Luer-lock type ports, 90° apart, each possessing an internal female Luer taper in accordance with figure 2, female end, of ANSI Z70.1, except that the taper shall be 0.060 $\pm$ .007 inch per inch and as indicated on the drawing. The third arm of the stopcock shall be a male Luer connector in accordance with figure 2, male end, of ANSI Z70.1, except that the taper shall be 0.060 $\pm$ .007 inch per inch and as indicated on the drawing. All female and male Luer tapers shall meet the requirements specified in 3.2.2 and 3.2.3, respectively.

3.2.2 Taper and depth of female Luer cavities. The two female Luer cavities of the stopcock shall be tested for proper taper of Luer cavity as specified in 4.4.6. When the Luer cavities are fitted onto the plug gage, there shall be no movement of the Luer cavities when they are pressed lightly from side to side in all directions. In addition, the "GO" position (excessive depth position) of the feeler gage shall not be exceeded (i.e., the plug gage shall not bottom into the Luer cavity without allowing passage of the feeler gage between the hub cavity and the shoulder of the plug gage).

3.2.3 Taper of male Luer connector. The male Luer connector of the stopcock shall be tested for proper taper as specified in 4.4.7. When the ring gage is fitted onto the male Luer connector, there shall be no movement of the ring gage on the tapered connector tip when the gage is pressed lightly from side to side in all directions. In addition, the tip of the connector shall not protrude beyond the "NO GO" position (outermost surface of the gage) of the ring gage. The male connector shall be designed to form a leakproof connection with standard Luer hypodermic devices.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:33Z
Check the source to verify that this is the current version before use.

3.3 Safety and freedom from pyrogens. All components of the stopcock shall be safe and pyrogen-free when tested as specified in 4.4.4 and 4.4.5, respectively.

3.4 Leakage of stopcock. When tested as specified in 4.4.8, the female Luer cavities, male Luer connector and the junction of the rotating spigot and stopcock body shall not show leakage exceeding the formation of one clinging air bubble.

3.5 Clear passage of fluid paths. All fluid paths of the stopcock shall be clean and free from particles and shall be tested for clear passage of all fluid paths as specified in 4.4.9. Stopcocks which are sealed or constricted shall be rejected.

3.6 Security of assembly. When tested as specified in 4.4.10, the rotating plug shall not pull out of, or separate from, the stopcock body.

3.7 Sterility. The stopcocks shall be sterile when tested with a biological indicator as specified in 4.4.2.

3.7.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item identification.
b. National stock number (NSN).
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.).
h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time and dates of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of laboratory or a designated qualified alternate.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Defense Personnel Support Center, Directorate of Medical Materiel, Technical Operations Division, ATTN: DPSC-ATQ, 2800 South 20th Street, Philadelphia, PA 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

3.7.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and sterilized simultaneously in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.2. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:33Z
Check the source to verify that this is the current version before use.

3.8 Certificate of freedom from pyrogens, safety and biological tests. A certificate listing all lot or control numbers shall state that representative samples from each lot or control have been tested in accordance with the methods described in the USP and found to be free from pyrogens and safe and have passed the biological tests. This certificate shall include the specific tests performed and the test results. This certificate shall be submitted by the contractor in the manner as specified in 3.7.1.

3.9 Manufacturer's record. The manufacturer shall place a code or lot number on the immediate container or unit package which can be verified against the manufacturer's records of all tests made on sterile, pyrogen free and nontoxic units. These records shall be made available for inspection by any duly qualified medical officer of the Armed Forces, National Institutes of Health, or any official of the Federal Food and Drug Administration. No shipment shall be made until all the tests have been completed, verified, and recorded in the permanent records of the manufacturer.

3.10 Instructions. A detailed instruction sheet for proper use of the stopcock shall be provided with each stopcock.

* 3.11 Workmanship. Workmanship shall be first class throughout. Stopcocks shall be free from defects which detract from their appearance or affect their serviceability. The stopcock shall be free from cracks, holes, pits or foreign matter. In addition, stopcocks shall be free from flash on, or in, all Luer surfaces.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for examination purposes shall be one stopcock.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:33Z
Check the source to verify that this is the current version before use.

TABLE I. Sampling for examination.

Type defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

4.2.2 For tests.

4.2.2.1 For sterility test. Sampling for sterility test shall be in accordance with the USP for parenteral devices and as specified in 4.4.2.

4.2.2.2 For pyrogen and safety tests. Sampling for pyrogen and safety tests shall be as indicated in the USP for transfusion and infusion assemblies. Samples shall be selected which are to be representative of each lot of each plastic used in the manufacture of the stopcocks. As an alternate, sampling shall be on a sterilizer lot basis when sterilizer lots are smaller than, or equal to, production lots.

4.2.2.3 For other tests. Sampling for other tests shall be conducted in accordance with MIL-STD-105 and table II. The unit of product for testing purposes shall be one stopcock. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

TABLE II. Sampling for tests.

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Stopcock	Material - Class II	3.1.1	4.4.3	1/	2/
	Extractable metals	3.1.1	4.4.1	1/	2/
	Dimensions	3.2	-	S-1	2.5
	Taper and depth of female Luer cavities	3.2.2	4.4.6	S-1	1.0
	Taper of male Luer connector	3.2.3	4.4.7	S-1	1.0
	Leakage test	3.4	4.4.8	S-2	1.0
	Clear passage of fluid paths	3.5	4.4.9	S-1	1.0
	Security of assembly	3.6	4.4.10	S-1	1.0

1/ Sufficient samples to perform the biological tests and extractable metal tests shall be selected from each lot and shall be representative of each batch of plastic used in the manufacture of the stopcocks.
2/ No defects permitted.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:33Z
Check the source to verify that this is the current version before use.

4.3 Examination.

4.3.1 Examination of stopcock. The stopcocks shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in tables I and III. Examination, however, shall not be restricted to defects listed in table III. Any stopcocks in the sample containing one or more defects shall be rejected, and if the number of defective stopcocks in any sample exceeds the acceptance number for that sample, the lot represented by that sample shall be rejected.

TABLE III. Classification of defects.

Categories	Defects
Major	
101	Stopcock not of the design specified.
102	Stopcock not free from cracks, holes, pits, burrs, slivers, flash or foreign matter.
103	Rotating spigot (handle) does not exhibit smooth operation without sticking or binding when turned in all positions.
104	Fluid paths obstructed or not free from foreign matter.
105	Silicone or other lubricants present in Luer cavities or on male Luer connector.
106	Rotating spigot exhibits play when turned instead of friction-type motion.
107	More than two of the fluid paths integrally connected at any open flow setting.
108	Rotating spigot does not possess "OFF" or "CLOSED" position marking.
Minor	
201	Instructions for use missing or incomplete.
202	Slight flashing on surfaces other than Luer.

4.3.2 Packaging inspection. The inspection of the individual and unit package, packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:33Z
Check the source to verify that this is the current version before use.

4.4.1 Extractable metals. The stopcock components shall be tested for extractable metals. The specimens (all components of the stopcocks) shall be cut into small segments and eluted for 72 hours at 50°C, 24 hours at 70°C, or 60 minutes at 121°C. Each gram of specimen shall be covered with 5 milliliters (mls) of distilled water (should the total weight of the components be 10 grams, the pieces shall be covered with 50 mls of distilled water). At the end of the elution period, the parts shall be agitated and the eluate removed and submitted for analysis for lead, cadmium and barium by atomic absorption spectrophotometry. The absorbances of the solution shall be determined at the recommended instrument parameters using Hollow-Cathode Pb, Ba, and Cd lamps on a spectrophotometer such as the Perkin-Elmer Model 303 or Varian Model AA-6 Atomic Absorption Spectrophotometer. The absorbances of the sample solutions shall be determined using the same conditions as the standards. Standard solutions for each element shall be prepared of 1, 2 and 5 ppm.

Calculations:

A x F x dilution ÷ W = ppm of the element

A = Absorbance
F = Factor = (Concentration standard in ppm) * ÷ Absorbance
W - Sample weight in grams

* The three standards are averaged to yield ppm/unit A.

4.4.2 Sterility test. All fluid paths of the stopcock shall be tested for sterility in accordance with the method specified in the USP for parenteral devices using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in 4.4.2.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for type I sterilization and 30 for type II sterilization including inoculated carrier and product test units.

4.4.2.1 Inoculated product or simulated product. For each type of sterilization the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55° - 65°C
Dry Heat	B. stearothermophilus, or	55° - 65°C
	B. subtilis var. niger (or globigii)	30° - 35°C
Gas	B. subtilis var. niger (or globigii)	30° - 35°C
Radiation	B. pumilus	30° - 35°C

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:33Z
Check the source to verify that this is the current version before use.

4.4.2.2 Inoculated carrier. Inoculated carriers (paper spore strips), if used, shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.2 and 4.4.2.1. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.4.2.1.

When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

4.4.2.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

4.4.2.4 Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP.

4.4.3 Biological tests for Class II plastics. All plastic components used in the manufacture of the stopcocks shall be tested in accordance with the requirements of the USP for Class II plastics.

4.4.4 Safety (toxicity) test). Samples of the finished stopcock shall be tested for safety in accordance with the USP for transfusion and infusion assemblies.

4.4.5 Freedom from pyrogens. Samples of the finished stopcock shall be tested for freedom from pyrogens in accordance with the USP for transfusion and infusion assemblies.

4.4.6 Taper and depth of female Luer cavities. The female Luer cavities shall be tested with the male Luer taper plug gage in accordance with figure 5 of ANSI Z70.1 (this gage shall be as shown in the ANSI except that it shall be constructed with a suitable pneumatic path to perform the leakage test as specified in 4.4.8). The female Luer cavities shall be slipped onto this gage with a straight motion, without turning or twisting the plug gage into the cavity. Once the female Luer cavity is slipped onto the gage, no additional force shall be exerted to force the plug gage into the Luer cavity. The cavities shall be observed for movement when they are pressed lightly from side to side on the plug gage. The feeler gage, in accordance with figures 5 and 6 of ANSI Z70.1, shall then be inserted between the Luer cavities and the shoulder of the plug gage to insure that the "GO" position is not violated.

4.4.7 Taper of male Luer connector. The taper of the male connector shall be tested with the female Luer ring gage in accordance with figure 7 of ANSI Z70.1. The male connector shall be slipped into this gage with a straight motion, without turning or twisting the male connector into the gage. Once the male connector is slipped into the gage, no additional force shall be exerted to force the male connector into the gage. The connector shall be tested for movement (wobble) and protrusion beyond the "NO GO" position of the gage.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:33Z
Check the source to verify that this is the current version before use.

4.4.8 Leakage test. All fluid paths of the stopcock shall be subjected to an internal air pressure of 10 pounds per square inch gage (psig). The position of the rotating plug shall be in line with the male connector so that only two fluid paths shall be integrally connected. The modified plug gage specified in 4.4.6 shall be inserted into one of the female Luer connectors with a straight motion, without turning or twisting the plug gage into the cavity with a force not to exceed 5 pounds. The feeler gage shall then be placed onto the plug gage to assure that the "GO" position is not violated. Another plug gage (not modified with a pneumatic path) in accordance with figure 5 of ANSI Z70.1 shall then be inserted into the other female Luer connector with a straight motion, without turning or twisting the plug gage in the female Luer cavity and with a force not to exceed 5 pounds. The feeler gage shall be placed onto this plug gage to again assure that the "GO" position is not violated. A hypodermic needle, with the lumen sealed, shall be used to seal the male Luer connector in the following manner: The hub of the hypodermic needle shall be tested to comply with acceptable Luer cavity as specified in 4.4.6, except that in addition, the Luer hub cavity of the needle shall also pass the "NO GO" position of the feeler gage. A needle possessing a hub cavity that passes the above requirements for diameter and depth of Luer cavity shall then be used to seal the Luer connector of the stopcock. Air shall then be internally admitted to the testing arrangement in a leakproof manner, with internal air pressure of 10 psig and the entire testing stopcock assembly completely immersed into a suitable container of water with the air pressure applied for at least 3 minutes. The entire stopcock test assembly shall be observed for leakage as specified in 3.4. The two remaining fluid paths shall be tested in turn by keeping the above arrangement essentially intact, but the position of the rotating plug shall be placed in line with the female Luer connector containing the plug gage that does not possess the pneumatic path (i.e., making sure that the modified plug gage with the pneumatic inlet is part of the pneumatic path circuit). It is recommended that the plug gage containing the pneumatic path be placed in the female Luer connector (directly opposite the male Luer connector) so that no reversing of the plug gages shall be necessary.

* 4.4.9 Clear passage of fluid path. The stopcocks shall be tested with an appropriate testing fluid applied internally to the stopcock with the stopcock placed in a horizontal position. Each fluid path shall be tested, in turn, using testing assembly fixtures as specified in 4.4.8 (or similar arrangement), except that one port shall be opened for fluid exit during the test for each path. The stopcock shall allow free passage of the testing fluid within 3 seconds after an internal pressure of 1.3 psig (3 feet of water) has been applied to the stopcock.

* 4.4.10 Security of assembly test. The stopcock shall be tested by applying a dynamic axial force of 25 pounds to the rotating plug member in a suitable tensile-strength device with the stopcock body appropriately secured.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C, as specified (see 6.1).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:33Z
Check the source to verify that this is the current version before use.

5.1.1 Unit of issue. One package containing fifty stopcocks constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Individual container. Each stopcock along with instructions for use shall be packaged sterile in a sealed, peel-open type container, capable of maintaining sterility of contents unless the container is damaged or opened. The container shall be designed with tabs to facilitate opening. The instructions may be printed on the container or may be inserted within the container.

5.1.2.2 Unit package. Fifty individually packaged stopcocks shall be packaged in a box conforming to PPP-B-636, type CF, class domestic. Closure of the box shall be as specified in the appendix of PPP-B-636.

5.1.3 Level C. Stopcocks shall be clean, dry and individually packaged sterile along with instructions for use. Fifty individually packaged stopcocks shall be further packaged in a manner that will afford adequate protection against deterioration and physical damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.1).

5.2.1 Level A. Twelve unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-585, class 3 style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in the appendix of PPP-B-636.

5.2.2 Level B. Twelve unit packages shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1, or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The packaged stopcocks shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carrier's rules and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:33Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Individual package. In addition to any commercial markings, each individual package shall be marked in accordance with the requirements of the Food, Drug and Cosmetic Act. The sterility lot (control) number and the legend "STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED", or similar statement, shall be shown.

5.3.2 Unit package. Each unit package shall be marked in accordance with MIL-STD-129. Sterility lot (control) number, contract or purchase order number and name of contractor shall be shown. The method of sterilization (G-Gas, R-Radiation, C-Cold, S-Steam, D-Dry Heat Sterilization) shall be shown following the level of packaging and date of packaging. Example: "A-5/74-G". Final sterilization shall be on or after indicated packaging date.

5.3.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Sterility lot (control) number and method of sterilization as specified in 5.3.2 shall be shown.

6. NOTES

6.1 Ordering data. Procurement documents shall specify the following:

a. Number, title and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
c. NSN.

6.2 This specification covers the following item appearing in the Federal Supply Catalog:

NSN	Item Identification
6515-01-035-7962	STOPCOCK, INTRAVENOUS THERAPY, 3-Way, Plastic, Disposable, 50s.

* 6.3 Changes from previous issue. The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Offerors and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activity:
DOD - MB

Preparing activity:
DLA-DM

Project No. 6515-2146

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:33Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS: This form is provided to solicit beneficial comments which may improve this document and enhance its use. DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity. A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed.

NOTE: This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

DOCUMENT IDENTIFIER *(Number)* AND TITLE

MIL-S-37813A STOPCOCK, INTRAVENOUS THERAPY, 3-WAY, PLASTIC, AYER TYPE

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1. ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID, RESTRICTIVE, LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW.

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDED WORDING CHANGE

C. REASON FOR RECOMMENDED CHANGE(S)

2. REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*

TELEPHONE NO.

DATE

DD FORM 1 OCT 76 1426

EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:33Z
Check the source to verify that this is the current version before use.