MILITARY SPECIFICATION

STAND, BED SERVICE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope This specification covers an adjustable bed service stand, with folding legs, for use with an adjustable hospital bed.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

TT-E-529	- Enamel, Alkyd, Semi-gloss
TT-P-1757	- Primer Coating, Zinc Chromate, Low Moisture Sensitivity.
PPP-B-636	- Boxes, Shipping Fiberboard.
PPP-T-60	- Tape Packaging, Waterproof.

MILITARY

MIL-C-5541	- Chemical Conversion Coatings on Aluminum and Aluminum Alloys

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.

STANDARDS

FEDERAL

FED-TEST METHOD-STD-NO. 151 - Metals: Test Methods
FED-STD-595

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

25024 - Bed, Folding, Hospital, Field.

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3 REQUIREMENTS

3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4.1.2.1 and 6.2).

3.2 Material. Except as hereinafter specified, materials shall conform to the applicable portions of the specifications, standards, and drawing listed herein. Materials shall be subjected to the tests and meet the requirements specified in these documents. The metals shall conform to the chemical and physical requirements of the applicable specifications. Materials, not specifically designated herein, but used in the fabrication of the stand, shall be suitable for the purpose intended.

3.3 Design and construction. The stand shall be constructed in conformance with the applicable portions of DPSC Drawing No. 25024 and as specified herein. (Sheet No. 8 of the drawing, and pages 6, 7, 9, and 10 of the Bill of Materials thereto, contain all references to the stand.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.

3.4 Finish. All surfaces to be painted shall be thoroughly cleaned and chemically degreased prior to chemical treatment. All aluminum parts shall be finished with a chemical film conforming to MIL-C-5541. After degreasing and chemical treament a zinc-chromate primer conforming to TT-P-1757 shall be applied to all metal parts. The primer coat shall form a grounding bond and base for the finishing enamel. The primer coat shall be followed by two baked-on spray coats of synthetic, semi-gloss, enamel in accordance with TT-E-529 Color shallbe oliver drab, No 24708 of FED-STD-595. The resultant finish shall be hard, smooth, and highly resistant to abrasion, chipping, and cracking. When tested as specified in 4.5.2, enamel shall not peel off to expose any of the following:

1. Any area exceeding 1/8 inch average diameter.
2. More than one area exceeding 1/16 inch average diameter.
3. More than five areas of any size.

3.5 Identification marking. Each unit shall be marked permanently and legibly, with the contract or purchase order number, and the name or registered trademark of the manufacturer, and the National Stock Number.

3.6 Workmanship Workmanship shall be first class throughout. Item shall be free from defects which detract from its appearance or impair its serviceability All burrs and sharp edges shall be removed from the item.

3.6.1 Welding Welding shall be durable, sound in every detail and in accordance with the latest practices recommended by the American Welding Society All welds shall be produced by qualified welding operators with satisfactory welding equipment using satisfactory filler metal or electrode where applicable to provide the best possible weld Surfaces to be welded shall be free from grease, oil, oxides or any other foreign matter. Any cleaning method which will remove these substances and leave the surface clean is satisfactory All weld joints shall blend into the adjacent metal in gradual, smooth curves. Beads shall be smooth and free of slag, excessive undercut, or excessive splatter. Sufficient metal shall be added to provide a suitable fillet or reinforcement unless otherwise specified Welds shall be oxidized All welds shall be continuous with no visible voids, pin holes, or cracks.

4 QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

4.1.2.1 Classification of inspection. The inspection requirements specified herein are classified as follows:

a. first article inspection.
b. Quality conformance inspection.

4.1.2.1.1 First article inspection. First article inspection shall be performed on the stand when a first article sample is required (see 3.1). This inspection shall include examination of 4.4 and tests of 4.5.

4.1.2.1.2 Quality conformance inspection. Quality conformance inspection shall include the examination of 4.4 and tests of 4.5.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the metals may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government and provided also that the test methods and the results obtained comply with those described in this specification. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Inspection lot.

4.2.1 Raw materials. An inspection lot for raw material inspection shall be all material produced under the same conditions and controlled to the degree that uniformity of product within the designated lot can be assured. Lot identification of the raw material by the manufacturer is one method of control, as cited above, which is considered acceptable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.

4.2.2 End product. An inpsection lot for inspection of end product shall consist of all stands of the same type produced under essentially the same conditions and offered for inspection at one time.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I. The unit of product for sampling purposes shall be one stand.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and Minor combined)	II	4.0
For dimensional examination	S-1	2.5

4.3.2 For tests Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II, for an acceptance level of zero.

TABLE II. Sampling for tests

Component (Unit of product)	Characteristic	Requirement	Test procedure	Inspection level
Stand	Enamel	3.4	4.5.2	1/

1/ Two stands shall be selected randomly from each day's enamelling.

4.3.3 Raw materials. A sufficient quantity of each lot of each metal used in the manufacture of the stand shall be selected for chemical analysis.

4.4 Examination. The stand shall be examined to determine compliance with all requirements contained in this specification

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.

4.4.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table III. Examination shall not be restricted to the classified possible defects listed below.

TABLE III. Classification of defects.

Categories	Defects
Major	
101	Any component part missing, broken or otherwise unusable.
102	Exposed edge not free of burrs or sharpness.
103	Components not properly assembled or aligned.
104	Any weld not as specified
105	More than one spot weld missing
Minor	
201	Identification marking missing, incomplete, or illegible
202	Finish chipped, cracked, scratched, or rough
203	Spot weld spatter present.

4.4.2 Dimensional examination. The stand shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect

4.4.3 Packaging inspection. The inspection of the unit package, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.5 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.5.1 Chemical composition. Four ounces from each lot of material shall be tested in accordance with Fed Test Method Std. No. 151

4.5.2 Enamel adhesion. Unless otherwise specified, a minimum of two stands from each day's production shall be run through all steps of the regular production process, including enameling. Adhesion of the painted test specimens should generally be determined within a maximum period of 72 hours drying. Press a 2 inch length of a somewhat longer piece of water-resistant pressure-sensitive adhesive tape (3/4 inch width) conforming to PPP-T-160, (transparent) firmly onto a flat or cylindrical surface of the item, rubbing out all air bubbles under the tape. Allow approximately 10 seconds for the test area to

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.

return to room temperature. Grasp a free end of the tape and at a rapid speed strip it form the item by pulling the tape back upon itself at 180 degrees (in such a manner that the tape is folded back to back during the procedure). Observe for bared areas where the enamel is removed. Disregard flecks of enamel on tape where the underlying metal or phosphate coating is not visibly exposed. If any test specimen indicates failure, all items processed since last acceptance shall be rejected and corrective action taken.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C, as specified (see 6.1).

5.1.1 Unit of issue. One stand, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Unit package. Each stand shall be folded, wrapped in kraft paper, and packagedin a snug fitting box constructed in accordance with PPP-B-636, type CF, class domestic. The unit shall be cushioned within the box with fiberboard pads, in such a manner so as to prevent shifting of or damage to the stand. Closure shall be as specified in the box specification.

5 1 3 Level C. Each stand shall be folded and packaged in a manner that will afford adequate protection against physical damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.1).

5.2 1 Level A. Six unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

5.2.2 Level B. Six unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The packaged stands shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct sipment from the supply source to the first receiving activity These packs shall conform to the carrier's rules and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.

5.3 Marking. Each unit package and shipping container shall be marked as specified in MIL-STD-129.

6. NOTES

6.1 Intended use. The stand is intended for use with adjustable bed, NSN 6530-01-015-5189 or 6530-00-935-4184.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents.

a. Title, number and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
c. When a first article is required (see 3.1, 4.1.2.1 and 6.3).

6.3 First article. When a first article is required, it shall be tested and approved under the appropriate provisions of the Defense Acquisition Regulation/Armed Services Procurement Regulation. The contract or purchase order specifies the quantity, procedure for delivery and other specific instructions for inspection and approval of the first article.

6.4 This specification covers the following item appearing in the Federal Supply Catalog.

National Stock Number	Item Identification
6530-00-115-6028	STAND, BED SERVICE, Aluminum, 24 by 11-5/8 by 11-5/8 inches.

Custodians
Army - MD
Navy - MS
Air Force - 03

Review activity:
DoD - MB

Preparing activity
DLA-DM

Project No. 6530-1439

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS This form is provided to solicit beneficial comments which may improve this document and enhance its use DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government Fold on lines on reverse side, staple in corner, and send to preparing activity Attach any pertinent data which may be of use in improving this document If there are additional papers, attach to form and place both in an envelope addressed to preparing activity A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed

NOTE This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

DOCUMENT IDENTIFIER *(Number)* AND TITLE
MIL-S-37826 STAND, BED SERVICE

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1 ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID RESTRICTIVE LOOSE OF AMBIGUOUS? PLEASE EXPLAIN BELOW

A GIVE PARAGRAPH NUMBER AND WORDING

B RECOMMENDED WORDING CHANGE

C REASON FOR RECOMMENDED CHANGE(S)

2 REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*	TELEPHONE NO
	DATE

DD FORM 1 OCT 76 1426 EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

POSTAGE AND FEES PAID

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

DEFENSE PERSONNEL SUPPORT CENTER
DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:34Z
Check the source to verify that this is the current version before use.