FED. SUP CLASS
6640



PART NUMBER	CAPACITY (ml)
MS36065-1	125
MS36065-2	250
MS36065-3	500

MATERIAL: Borosilicate glass in accordance with Federal Specification DD-G-541, Type I, Class A.

DESIGN: Flasks shall be in accordance with Federal Specification NNN-F-240, Type IV, Class 2, Style III.

Mouth of the flask shall have a flared neck to provide for a liquid seal.

Each flask shall have a marking area located in accordance with commercial practice.

Each flask shall be permanently and legibly marked with manufacturer's name or trademark and capacity.

Pennyhead stoppers shall be ground joint size number 22 and in accordance with Federal Specification DD-S-750, Type I, Composition 1, Grade A.

NOTES:
1. Flasks are supplied with stoppers.
2. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements as specified herein.
3. For design feature purposes this standard takes precedence over procurement documents referenced herein.
4. Referenced documents shall be of the issue in effect on date of invitations for bid.

USER ACTIVITIES: NAVY - AS, OS, YD
ARMY - ME, PS
NAVY - MC

APPROVED 29 Jun 1960 REVISED (A) 15 Aug 1969

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

(A) ENTIRE STANDARD REVISED

P.A. MU Other Cust MS 82	TITLE FLASK, ERLENMEYER, IODINE DETERMINATION	MILITARY STANDARD MS36065	
Procurement Specification NNN-F-240	SUPERSEDES:	SHEET 1 OF 1	

DD FORM 1 SEP 63 672 (Coordinated) EDITION OF 1 SEP 57 MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:33Z
Check the source to verify that this is the current version before use.