REVIEW ACTIVITIES: OTHER - NSA
USER ACTIVITIES: ARMY - ME

FED. SUP CLASS
6640



| PART NUMBER | DIMENSIONAL DATA mm | | | | | | |
|---|---|---|---|---|---|---|---|
| | TOP OD | ID AT PLATE | DIAMETER OF PERFORATED AREA | H ABOVE PLATE | H OVERALL | STEM LG | STEM OD |
| MS36093-5 | 51 ± 3 | 41 to 47 | 32 ± 2 | 24 ± 2 | 89 ± 3 | 46 ± 3 | 13 ± 2 |
| MS36093-1 | 67 ± 3 | 55 to 61 | 45 ± 3 | 30 ± 2 | 101 ± 3 | 46 ± 3 | 13 ± 2 |
| MS36093-2 | 106 ± 4 | 90 to 98 | 80 ± 3 | 46 ± 3 | 160 ± 10 | 71 ± 10 | 20 ± 2 |
| MS36093-3 | 142 ± 5 | 125 to 134 | 115 ± 5 | 59 ± 3 | 202 ± 10 | 88 ± 10 | 22 ± 2 |
| MS36093-6 | 171 ± 5 | 150 to 160 | 139 ± 5 | 61 ± 4 | 221 ± 10 | 103 ± 10 | 25 ± 2 |
| MS36093-4 | 213 ± 5 | 185 to 195 | 173 ± 5 | 70 ± 4 | 270 ± 10 | 115 ± 10 | 26 ± 2 |

MATERIAL: Vitreous, ceramic, whiteware.

FINISH: Glazed inside and outside, except for the rim.

DESIGN: Funnels shall be of the conventional Buchner type with a fixed perforated filter plate and beveled tip terminating at a 45° angle.

NOTES:
1. For design feature purposes this standard takes precedence over procurement document referenced herein.
2. Referenced document shall be of the issue in effect on date of invitations for bid.

APPROVED 28 Jun 1961 REVISED (A) 4 JUN 69

(A) ENTIRE STANDARD REVISED

| | | | |
|---|---|---|---|
| P.A. | MU | TITLE: FUNNEL, FILTERING, LABORATORY porcelain, Buchner | MILITARY STANDARD |
| Other Cust | MS 82 | | MS 36093 |
| Procurement Specification DD-F-786 | | SUPERSEDES: | SHEET 1 OF 1 |

DD FORM 1 SEP 65 672 (Coordinated)

EDITION OF 1 SEP 57 MAY BE USED.

☆ U. S. GOVERNMENT PRINTING OFFICE: 1969-341-520/A-7231

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:34Z
Check the source to verify that this is the current version before use.