FED. SUP CLASS
6640

USER ACTIVITIES: ARMY - EL, MF

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.



PART NUMBER	NOMINAL DIMENSIONS in.			WELL DATA	
	LENGTH	WIDTH	HEIGHT	TUBE QUANTITY	NOMINAL DIAMETER mm
MS36163-3	13	2-7/8	3	20	22
MS36163-1	Canceled without replacement				
MS36163-2	Canceled without replacement				

MATERIAL: Aluminum in accordance with Federal Specification QQ-A-250.

DESIGN: Rack shall have two rows of ten holes each with the upper and lower shelves perforated.

NOTES:
1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements as specified herein.
2. Referenced document shall be of the issue in effect on date of invitations for bid.

APPROVED 30 Dec 1966 REVISED Ⓐ 20 June 1969

Ⓐ ENTIRE STANDARD REVISED

P.A. MU Other Cust MS	TITLE RACK, TEST TUBE, LABORATORY folding, aluminum	MILITARY STANDARD MS 36163	
Procurement Specification None	SUPERSEDES:	SHEET 1	OF 1

DD FORM 1 SEP 63 1072 (Coordinated)

EDITION OF 1 SEP 47 MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:35Z
Check the source to verify that this is the current version before use.