FED. SUP CLASS
6640



PART NUMBER	JOINT SIZE DESIGNATION	D TUBING END OUTSIDE DIAMETER approx mm
MS36268-1	10/30	14.5
MS36268-2	14/35	18.5
MS36268-3	18/38	22.75
MS36268-4	24/40	28.5

CANCELLED AFTER 16 NOV 1983 · NO SUPERSEDING STANDARD

Specification DD-G-541, Type I,

Joints shall be outer part, full length, which is medium length and shall conform cation DD-G-675, Type I.

Clearance shall be - 0.15 mm at gaging point.

design feature purposes, this standard takes precedence er procurement documents referenced herein.

Referenced documents shall be of the issue in effect on date of invitations for bid.

APPROVED 15 DEC 1961 REVISED (A) 16 NOV 1983

(A) CANCELLATION

P.A.	CmlC	TITLE	MILITARY STANDARD
Other Cust	BuMed DAY	GROUND JOINT, LABORATORY GLASSWARE Standard Taper, outer member	MS36268
Procurement Specification DD-G-675		SUPERSEDES:	SHEET 1 OF 1

DD FORM 1 SEP 57 672 (Coordinated) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE.

Cancellation
Proj. No. 6640-1211

NOTICE—When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility, nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:37Z
Check the source to verify that this is the current version before use.