FED. SUP CLASS
6640

MS36573-3

PART NUMBER	CAPACITY (mL)	NOMINAL TUBE LENGTH (in.)
MS36573-1	CANCELED WITHOUT REPLACEMENT	
MS36573-2	CANCELED WITHOUT REPLACEMENT	
MS36573-3	15	8

(A) MATERIAL: MS36573-3 shall be of soda-lime glass in accordance with ASTM E 438.

Bulb shall be of oil-resistance rubber or neoprene.

DESIGN: Bulb shall be constructed with heavy walls, and an inner ring to fit around the collar of the pipet to prevent slippage and provide a tight, leak-proof fitting.

Pipet shall be ungraduated and have the upper end of the tube flanged to fit the inner ring of the bulb. The end of the pipet shall be smoothly finished. use with evacuating type desiccator, MS36215-1.

The tip of the pipet shall be straight and plain.

The bulb shall be capable of filling at least two-thirds of the length of the pipet when it is compressed and released.

NOTES:
1. Dropping pipet is supplied with bulb.
(A) 2. In the event of a conflict between the text of this standard and the reference cited herein, the text of this standard shall take precedence.
(A) 3. FSCM: 22527

APPROVED 14 JUL 1969 REVISED (A) 22 JAN 1988

USER ACTIVITIES:
ARMY - MD, ME
REVIEWER ACTIVITIES:

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document

P. A. EA / Other Cust MS 99	INTERNATIONAL INTEREST	TITLE: PIPET, DROPPING general purpose	MILITARY STANDARD MS36573
PROCUREMENT SPECIFICATION: None	SUPERSEDES:		SHEET 1 OF 1

DD FORM 672-1 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-1466-2
REVISION

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:43Z
Check the source to verify that this is the current version before use.

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.