PART NUMBER	DESCRIPTION	ELECTRODE LENGTH (CM)	TEMP RANGE °C	pH RANGE	APPLICATION (MEASUREMENTS OF)
MS36583-1	Glass Body Ceramic Junction	13	-5 to 80	0-14	pH and Titration
MS36583-2	Plastic Body Liquid Filled Ceramic Junction	13	+10 to 80	0-14	pH and Titration
MS36583-3	Plastic Body Gelled Filled Plastic Junction	13	+10 to 80	0-14	pH and Titration
MS36583-4	Plastic Body Flat End Ceramic Junction	15	+10 to 80	0-14	pH on Moist Flat Surfaces

MATERIAL Electrodes shall be made with internal elements and external materials to meet the requirements for temperature, pH range and resistance to chemical attack.

DESIGN. Electrodes may be either plastic, glass, or ceramic and with either a side arm or a side filling hole. The lead may be detachable or permanently attached and shall be a minimum of 76 cm in length. The electrode plugs shall be designed so that they are readily adaptable for use with most pH meters.

PURPOSE These electrodes incorporate both a pH sensing element and reference electrode and are used for pH determinations and titration measurements of aqueous or small samples They are readily adaptable for field or classroom use

NOTE: Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements as specified herein

USER ACTIVITIES ARMY - CE
NAVY - NC

REVIEWER ACTIVITIES ARMY - MD

This military standard is approved for use by all Departments and Agencies of the Department of Defense Selection for all new engineering and design applications and for repetitive use shall be made from this document

P A EA
Other Cust MS 99
Procurement Specifications

INTERNATIONAL INTEREST

TITLE
ELECTRODE, HYDROGEN METER (COMBINATION TYPE)

SUPERSEDES
MS36465

MILITARY STANDARD
MS36583
SHEET OF

DD FORM 672 (COORDINATED)

PREVIOUS EDITIONS OF THIS FORM ARE [illegible]

PROJECT NUMBER

STANDARDIZATION DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:43Z
Check the source to verify that this is the current version before use.