
FED. SUP CLASS
6640



PART NUMBER	NOMINAL DIMENSIONS (mm)				
	A OVERALL LENGTH OF SUPPORT ARM	B WIDTH OF SUPPORT ARM	C THICKNESS OF SUPPORT ARM	D WIDTH OF SLOTS	E MAX DIAMETER ROD ACCOMMODATED
MS36021-1	305 - 405	64	16	13	17

MATERIAL: Support arm shall be of hardwood such as rosewood, walnut, birch, or beech, oiled or treated to prevent splitting or absorption.

Clamp shall be cast iron, aluminum alloy, or equivalent with a corrosion-resisting finish.

Thumbscrew shall be of aluminum or brass.

DESIGN: Threads shall be in accordance with Federal Handbook H28, Unified National Coarse Thread Series.

Thumbscrew shall have a thread length of 25 to 38 mm.

Holes shall be equal in size and of sufficient size and conical shape to accommodate standard laboratory funnels. They shall be equally spaced along the center line which travels the length of the support arm.

NOTES: 1. Use only stand, laboratory apparatus support (MS36019-1 or MS36020-1) with support arm, funnel, laboratory (MS36021-1) for design and engineering.

© ENTIRE STANDARD REVISED

APPROVED 23 Dec 1959 REVISED (A) 15 Dec 1967 (B) 3 Dec 1975 (C) 4 OCT 1983

USER ACTIVITIES: ARMY - MD, DM NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME AIR FORCE - 82

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		SUPPORT ARM, FUNNEL, LABORATORY	MS36021
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 2

PROJECT NUMBER 6640-1205

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:33Z
Check the source to verify that this is the current version before use.

2. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

3. For design feature purposes, this standard takes precedence over procurement documents referenced herein.

4. Referenced documents shall be of the issue in effect on date of invitations for bid.

5. FSCM: 20512 and 22527.

USER ACTIVITIES:

REVIEWER ACTIVITIES:

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

APPROVED REVISED

P.A. Other Cust	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD MS 36021
Procurement Specifications	SUPERSEDES:		SHEET 2 OF 2

PROJECT NUMBER 6640-1205

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:33Z
Check the source to verify that this is the current version before use.