
FED. SUP CLASS
6640



PART NUMBER	NUMBER OF WELLS	WORKING VOLUME (mL)	DIMENSIONS (mm) LENGTH	WIDTH	HEIGHT
MS36243-1	60	0.01	83.8	56.4	12.7
MS36243-2	96	0.30	127.8	85.6	15.8

MATERIAL: Plates shall be made of polystyrene plastic.

DESIGN: Plates shall be optically clear, sterile, rigid and disposable. They shall be supplied with a friction fit lid. MS36243-2 shall be individually wrapped.

USE: Plates are used for tissue culture quality for cultivation of human and animal cells in monolayer. Also as a mechanism for micro-assay in virology for virus neutralization and isolation procedures and in cell biology for cell assays and replicating studies.

NOTES:
1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.
2. FSCM: 11712 and 08455.

USER ACTIVITIES: ARMY - MD, EM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 82

APPROVED 20 JUL 1983 REVISED

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		PLATE, TISSUE CULTURE	MS36243
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-1189

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:36Z
Check the source to verify that this is the current version before use.

☆U.S. GOVERNMENT PRINTING OFFICE: 1983 — 606-034/4060