
FED. SUP CLASS
6640



PART NUMBER	APPROXIMATE DIMENSIONAL DATA inches			
	DIAMETER OF BASE	DIAMETER OF SPIDER	OVERALL HEIGHT	TYPE OF GAS USED
MS36342-1	3-1/4	3-1/4	5	Acetylene

MATERIAL: The base and spider shall be cast iron. The inlet tube, tip post and mixer, tip body, nozzle and lock nut shall be brass.

FINISH: The base shall have either a black-crackle or a smooth, baked-enamel finish. The spider shall be nickel or chromium plated.

DESIGN: The spider shall be round in shape and shall be of such design that a flat surface will rest evenly on it without slipping across the top. The free end of the inlet tube shall be fitted for the attachment of a 4 foot length of rubber tubing, furnished with the burner.

NOTES:

1. Burner shall be in accordance with Military Medical Purchase Description for Burner, Gas, Laboratory, Bunsen, Acetylene, and the applicable paragraphs of Federal Specification GG-B-511.
2. Item referenced herein is for use by Military Medical Services only.
3. For design feature purposes, this standard takes precedence over procurement documents referenced herein.
4. Referenced documents shall be of the issue in effect on date of invitations for bid.

APPROVED 22 JUN 1962 REVISED

P.A. CmlC Other Cust Buried MAA	TITLE BURNER, GAS, LABORATORY Bunsen type, acetylene	MILITARY STANDARD MS36342
Procurement Specification See Note 1	SUPERSEDES:	SHEET 1 OF 1

DD FORM 1 SEP 57 672 (Coordinated) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE.

NOTICE—When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility, nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

Source: https://assist.dla.mil -- Downloaded: 202640-23T17:38Z
Check the source to verify that this is the current version before use.