A-A-50736
15 November 1987

COMMERCIAL ITEM DESCRIPTION

GENERATOR, ELECTRONIC MARKER, DIGITAL SIGNAL

The General Services Administration has authorized the use of this Commercial Item Description.

Scope. This Commercial Item Description (CID), describes a digital signal generator. This CID is meant as a minimum requirement for the digital signal generator in which only those manufacturers that meet or surpass the following requirements are qualified per this CID.

Classification.. Equipment covered by this CID may be commercially available equipment modified to the extent necessary to meet the following description. The equipment shall be type III, class 5, style F and color R in accordance with MIL-T-28800. This CID makes provision for the bid sample testing in accordance with MIL-T-28800.

Applicable documents.

Specifications and standards. The following specifications and standards form a part of this CID to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DODISS) and supplement thereto, cited in the solicitation.

SPECIFICATIONS

MILITARY

MIL-T-28800	Test Equipment for use with Electrical and Electronic Equipment, General Specification fo.

(Copies of specifications and standards, publications, and other Government documents required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting officer.)

Other publications. The following document forms a part of this CID to the extent specified herein. Unless otherwise specified, the issue of the documents which are DOD adopted shall be those listed in the issue of the DODISS specified in the solicitation. Unless otherwise specified, the issues of documents not listed in the DODISS shall be the issue of the nongovernment documents which is current on the date of the solicitation.

PSC 6625

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

AMERICAN NATIONAL STANDARDS INSTITUTE

ANSI-C62.41-1980

(Application for copies of ANSI publications should be addressed to the American National Standards Institute, 1430 Broadway, New York, NY 10018.)

Order of precedence. In the event of a conflict between the text of this CID and the references cited herein (except for associated detail specifications, specification sheets or MS standards) the text of this specification shall take precedence. Nothing in this CID, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

Salient characteristics. The equipment shall be capable of operation within the accuracies, limits, and specifications herein.

1. Safety. The equipment shall be UL (Underwriters Laboratories) listed and approved and/or shall comply with the type III safety requirements of MIL-T-28800.

a. Leakage current protection. Protection shall be provided from leakage current in excess of 3.5 milliamperes (mA) rms (AC and DC) from any accessible conductive parts of equipment (including control shafts with knobs broken or removed and recessed calibration or adjustment controls) to either pole of the power source for any position of the power switch(es).

b. Radio frequency interference (RFI) capacitors. Any capacitor used in RFI filters shall be rated for at least four (4) times the operating voltage.

c. Type III ventilation openings.

(1) Ventilation openings shall be designed to prevent accidental contact with live electrical circuits and mechanical moving parts.

(2) Top ventilation openings shall incorporate small mesh screening or similar design to prevent the expulsion of hot electronic particles due to failure of electronic devices such as capacitors.

2. PARTS, MATERIALS, AND PROCESSES. Except for marking, the digital signal generator shall comply with the type III parts, materials, and process requirements of MIL-T-28800. Nonstandard parts approval requests shall be in accordance with MIL-T-28800. Front panel marking shall be in accordance with MIL-T-28800.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

a. Restricted material. The equipment shall comply with the restricted material requirements of MIL-T-28800, including the requirement for mercury and radioactive materials, and shall contain no combination of materials which cause deterioration of any material contained in the equipment due to the effects of outgassing.

3. Design and construction. Unless otherwise specified herein, the equipment shall comply with the type III design and construction requirements of MIL-T-28800. The equipment shall be of solid-state construction. Encapsulation and embedment (potting) of subassemblies may be used only for high voltage circuits where construction is susceptible to leakage or corona, indicators and associated controls shall be in close proximity to each other, all potentiometer controls shall have at least a 270 degree rotation range (4.71 radians).

a. Type of construction. Construction of the digital signal generator shall be of a monolithic or a separate component design.

b. Modular design restrictions. The modular design concept for the digital signal generator is acceptable with the following restrictions:

(1) Each module shall contain a power circuit protection fuse.

(2) Each module plug-in card shall be keyed and mated with the back plane receptacle so that modules may not be randomly placed into the mainframe.

(3) Modules shall not be removable by access through the front or rear of the mainframe without the unaided hand.

(4) The digital signal generator shall come with one(1) blank space cover and one(1) extender board if deemed necessary.

(5) A package part number shall be assigned to the digital word generator as a whole system. This system part number shall be marked on the mainframe identification plate and all plug-in(s).

(6) All module interface shall be done in the back of the mainframe.

c. Signal input and output connectors. Unless otherwise specified, all signal input and output connectors shall be Type BNC female.

d. Internal cooling. The equipment shall be capable of operating continuously for twenty four(24) hours at 55°C. If forced air is used, it shall be as specified in MIL-T-28800.

(1) Forced air cooling motor. DC brushless motors shall be utilized for the forced air cooling motor in the equipment.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

(2) Air filter. Equipment utilizing a cooling scheme which exhausts air from the equipment through the cooling fan shall not require an externally accessible air filter.

(3) Thermal switch. A thermal switch shall be provided which will interrupt circuit power to the digital signal generator when the internal temperature rises to the degree of possible damage to the electronic components. A minimum of 10°C hysteresis drop from this tripping temperature must occur before circuit power is again applied.

4. Electrical power sources. The equipment shall operate from nominal commercial, military, and shipboard power sources of 100, 110, 120, 200, 220 and 240 volts ac (±10%) at line frequencies of 50 Hz to 400 Hz.

a. Electrical power connectors. The equipment shall operate with a ground at the primary source on either side of the single phase input.

b. Power consumption. Maximum power consumption of all components combined shall not exceed 630 watts.

c. Fuses. Fuse(s) shall be provided, on the exterior rear panel for replacement without entering the enclosure. The fuse(s) shall automatically disconnect the equipment from the power source in the event of excessive current requirements.

d. Surge and transient voltage protection. The digital signal generator shall contain surge and transient voltage protection internal of the enclosure. The transient voltage protection circuit shall be so designed as to prevent electronic device damage within the system per type III solid state tolerance bands of MIL-T-28800, the surge voltage protection circuit shall conform to ANSI/IEEE Std. C62.41-1980; 2.5KY, category A. The surge protection circuit shall drain excess power due to surges directly to the ground conductor for personnel protection. Normal operation of the digital signal generator may be inhibited and false operational signals may be generated under input voltage transients greater than ±20 percent and frequency transient conditions greater than ±5 percent from within the nominal type III steady-state tolerance bands of MIL-T-28800, damage (other than fuse failure) to the digital signal generator shall not occur during voltage transient conditions greater than ±20 percent and frequency transient conditions greater than ±5 percent within the maximum limits (30 percent voltage transients and 10 percent frequency transients) specified in MIL-T-28800. Normal operation of the digital signal generator shall be automatically resumed when the transient has passed and any blown fuses are replaced. Blown fuses shall be externally accessible. Testing shall be in accordance with MIL-T-28800.

5. Dimensions and weight. The size and weight of individual components shall be consistent with current commercial capabilities, refer to paragraphs 5.a and 5.b of this CID.

a. Equipment dimensions. The nominal dimensions per component for the digital signal generator should be approximately 7.75 inches height x 17 inches wide x 34 inches deep.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

b. Equipment weight. The weight including all accessories (excluding manual) per component, shall not exceed 53 lbs.

6. Enclosure requirements. Unless otherwise specified herein, the equipment shall comply with the style F enclosure requirements of MIL-T-28800.

a. Rackmount. Rackmount capability shall be available as an option for the equipment specified herein.

b. Tilt bail handle. A tilt-bail handle shall be provided in lieu of the hinged metal handles in accordance with MIL-T-28800. Exception to this paragraph may be acceptable as this is not a prime consideration of the equipment.

c. Reusable pouch or container. An attached accessory pouch may be used for storage of all supplied accessories in lieu of front-cover storage.

7. Marking and identification.

a. Marking. The visible surface adjacent to panel facilities such as variable controls, indicators, jacks, connectors and fuse holder(s), shall be marked with a suitable word, phrase, or abbreviation thereof, indicating the purpose or use of the part.

(1) Panel marking. The front panel marking shall comply with the type III parts, materials and processes panel marking requirements of MIL-T-28800; however, marking identification processes other than the type III requirements specified in MIL-T-28800 may be used, provided the marking and identification processes are capable of passing the front panel marking test when tested in accordance with this CID.

b. Items to be marked (if applicable).

(1) Volume and weight (including accessories) shall be marked on three external surfaces of the enclosure.

(2) The type designation of the AN nomenclature of the digital signal generator shall be stenciled on the exterior or main frame on at least three adjacent surfaces, on the transient case on two opposite sides. Marking shall be in block letters approximately 2 inches in height.

8. Environmental requirements.

a. Temperature requirements. Unless otherwise specified herein, the equipment shall meet its specified performance characteristics and other requirements when subjected to the class 5 temperature requirements specified by MIL-T-28800 and as listed below.

(1) Non-operating temperature. The equipment shall meet its performance and accuracy requirements in an non-operating environment of -40°C to +65°C.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:46Z
Check the source to verify that this is the current version before use.

(2) Operating temperature. The equipment shall meet its performance and accuracy requirements in an operating environment of 0°C to 55°C.

b. Humidity requirements. The equipment shall meet its required performance and accuracy requirements under humidity requirements for class 5 equipment as specified by MIL-T-28800.

c. Altitude requirements. The equipment shall meet its required performance accuracy requirements under the altitude requirements for class 5 equipment as specified by MIL-T-28800.

d. Vibration requirements. The equipment shall meet the class 5 vibration requirements as specified by MIL-T-28800.

e. Mechanical shock requirements.

(1) Shock, functional. The equipment shall meet the class 5 shock, pulse requirement as specified by MIL-T-28800.

(2) Transit drop. The equipment shall meet transit drop requirement as specified by MIL-T-28800.

(3) Bench handling. The equipment shall meet the requirements as specified by MIL-T-28800.

(4) Shock, high impact. The equipment shall meet the requirements as specified by MIL-T-28800.

f. Electromagnetic interference (EMI). The equipment shall comply with MIL-T-28800.

g. Magnetic environment. The equipment shall comply with the magnetic environment requirements of MIL-T-28800.

9. Reliability. Reliability verification testing, shall be in accordance with the type III, class 5 environmental conditions of MIL-T-28800.

a. Mean time between failures (MTBF). The equipment mean time between failures (MTBF) shall be 2,000 hours.

b. Pre-conditioning requirement. Each deliverable equipment shall be capable of being subjected to the environmental screening effects of burn in and thermal cycling, with thermal cycling being accomplished in the sample testing phase.

(1) Burn-in. Burn-in shall expose the equipment to extended periods of elevated temperatures, (40°C for a minimum of 100 hours). Each equipment which fails during burn-in shall be repaired and fully retested, until failure free. Failures which occur during the burn-in test shall be noted and reported but shall not count toward the establishment of MTBF. Prior to and following burn-in, a satisfactory operation test (SOT) shall be performed. A satisfactory operation check (SOC) shall be conducted at selected times during the burn-in test.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

(2) Thermal cycling. Thermal cycling shall expose the equipment to hot and cold cycles for class 5 equipment per MIL-T-28800. Each equipment which fails during thermal cycling shall be repaired and full cycling test shall be noted and reported but shall not count toward the cycling, a satisfactory operation check (SOC) shall be conducted during the last two upper and lower temperature limits.

10. Calibration and maintenance adjustments. The design of the system shall provide for readily accessible calibration adjustments and maintenance adjustments. These adjustments shall be provided by variable value components which are adjustable by the use of simple means. The calibration by substitution of selected components or parts is unacceptable unless specifically approved. The calibration adjustments, wherever possible, shall be accessible without removal of the instrument case or modules.

a. Maintainability. The equipment shall meet the type III maintainability requirements specified by MIL-T-28800.

11. Interchangeability. The equipment shall meet the type III interchangeability requirements specified by MIL-T-28800.

12. Digital interface. The equipment shall meet the digital interface requirements per MIL-T-28800.

13. Performance characteristics. Unless otherwise specified, all vernier controls shall be in the calibrated position and all performance requirements shall apply after a 30 minute warm up period in a laboratory ambient operating temperature.

a. General specifications. The digital signal generator shall have at least eight parallel channels with a minimum bit rate per channel of 100 Mbits-Hz.

b. Electrical function specification. Unless otherwise specified the digital signal generator shall meet the following minimum requirements:

(1) ECL drivers: The ECL drivers shall drive to a logic "low" of less than or equal to -1.60V and shall drive to a logic "high" of greater than or equal to -1.02V.

(2) Rise and fall times: The Rise and Fall times for inputs shall not exceed 3 ns.

(3) Skew: The skew between channels shall be less than or equal to 6 ns.

(4) Internal oscillator: The internal oscillator shall be programmable with a max frequency of 100 MHz.

(5) Internal clock range and resolution: The resolution shall not be less than the following:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

Resolution	Range
20 ns	20ns - 980ns
1 us	10 us - 999 us
100 us	1 ms - 99.9 ms
1 ms	.1s - .999s

(6) Programmability: The period for each state shall be programmable and shall be independent of all other channels.

(7) External clock: The external clock input shall have a frequency range from 8 Hz to 100 MHz.

(8) External clock resolution: External clock resolution shall not deviate more than 1 percent of the range.

c. Data capacity. The digital signal generator shall have a minimum capacity of 1024 bits per channel for all eight parallel channels.

d. System Integration. The digital signal generator shall be system integratable through an interface bus which shall comply with the requirements of MIL-T-28800D section 3.3.4 and shall be rack mountable to comply with the requirements of MIL-T-28800.

e. Device Interface. The digital signal generator shall include a minimum of one probe per channel plus one probe which shall be used for external clock trigger, or enable. All probes shall have a frequency response of 100 MHz or greater and shall have a retractable hook tip.

14. Calibration interval. After calibration, the equipment shall meet each performance requirement within the tolerances specified for a period of at least 9 months, based on an operating time of 1500 hours.

15. Operating documentation. Each unit shall include one copy of the operation and maintenance manual and one copy of calibration and testing procedures if not included in the O&M manuals.

Packaging, packing, and marking. The items shall be packaged in accordance with normal commercial practice and packed to assure acceptance by common carrier and provide product protection against loss and damage during multiple shipments, handling, and storage. The shipping container shall be in compliance with national motor freight classification and uniform freight classification. Shipping container shall be marked as specified in the contract or order.

Certification. The contractor shall certify that the product offered meets the salient characteristics of this description and that the product conforms to the producer's own drawings, specifications, standards and quality assurance practices and is the same product offered for sale in the commercial marketplace. Small businesses that cannot offer products with established commercial market acceptability, may compete by submission of bid samples. (The Government reserves the right to require proof of such conformance before first delivery and thereafter as may be otherwise provided for under the provisions of the contract).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

Known acceptable products: The following product(s) are known to have previously met the requirements of this CID and need not resubmit bid samples or certification. All other manufacturers must meet the requirements of this CID and submit to bid testing and certification as required.

Manufacturer	Model	NSN
HP	8175A-C01	
Interface Technology	RS-680/Opt 680-001	6625-01-225-8193

Custodian:
Air Force - 99

Preparing activity:
Air Force - 82

(Project 6625-F881)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL

FIRST CLASS PERMIT NO. 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

SA-ALC-MMEDO
Kelly AFB, TX 78241-5990

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:46Z
Check the source to verify that this is the current version before use.