


FED. SUP CLASS
6640




MS36487-1




MS36487-2




MS36487-3




MS36487-4




MS 36487-5



MS 36487-6



MS 36487-7



MS36487-8

USER ACTIVITIES:

REVIEWER ACTIVITIES:
ARMY - MD
NAVY - OS
AIR FORCE - 03

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

APPROVED 12 FEB 1982 REVISED

| | | | |
|---|---|---|---|
| P.A. EA<br>Other Cust MS 99 | INTERNATIONAL INTEREST | TITLE<br>DISPENSER, REAGENT<br>Plastic | MILITARY STANDARD<br>MS36487 |
| Procurement Specifications None | SUPERSEDES: | | SHEET 1 OF 2 |

DD FORM 1 MAY 73 672 (COORDINATED)

PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

Project Number 6640-0796
STANDARD DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:41Z
Check the source to verify that this is the current version before use.

| PART NUMBER | DESCRIPTION |
|---|---|
| MS36487-1 | Siphon type, made entirely of polyethylene, comprised of 152 cm (5 ft) length, 1 cm (.375 in.) ID, 12 mm (.50 in.) OD tubing with priming bulb and needle valve on one end. |
| MS36487-2 | Siphon type flexible polyethylene tube, 228.6 cm (7.5 ft) long, priming bulb on one end. |
| MS36487-3 | Bellows type, flexible discharge tube, polyethylene, 83 mm buttress thread, max delivery rate, 16.5 liter per minute. |
| MS36487-4 | Hand pump, polyethylene, siphon tube 86 cm long, friction fit spout can be turned up to eliminate dripping. Delivery rate 600 ml per stroke. |
| MS36487-5 | Bulb type, polyethylene, flexible discharge tube and no drip liquid return. For 4 to 20 liter containers. |
| MS36487-6 | Bulb type, polyethylene, flexible discharge tube and no drip liquid return. For 20 and 24 liter containers. |
| MS36487-7 | Battery operated type, powered by two "D" alkaline flashlight cells. PVC suction tube 33 cm long and a flexible polyethylene discharge tube 53 cm long. |
| MS36487-8 | Pump (polyethylene bottle) with polyethpropylene siphon and delivery tubes and top. Fits a 2.4 liter or 3.8 liter acid bottle. |

MATERIAL: The material shall be as specified in the table.

DESIGN: Acid pump dispensers shall be designed and constructed to perform the intended function.

NOTES:
1. Items differing from this military standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements as specified herein.
2. All of the above may be used as acid and chemical reagent dispensers.

USER ACTIVITIES:

REVIEWER ACTIVITIES:

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

APPROVED 12 FEB 1982 REVISED

| P.A. / Other Cust | INTERNATIONAL INTEREST | TITLE | MILITARY STANDARD |
|---|---|---|---|
| | | DISPENSER, REAGENT Plastic | MS36487 |
| Procurement Specifications | SUPERSEDES: | | SHEET 2 OF 2 |

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:41Z
Check the source to verify that this is the current version before use.