MIL-V-6753A
AMENDMENT 2
24 NOVEMBER 1958

SUPERSEDING
AMENDMENT 1
26 SEPTEMBER 1957

MILITARY SPECIFICATION

VOLTMETERS, 0-150 VOLT, 330 TO 1200 CYCLE, ALTERNATING CURRENT

This amendment forms a part of Military Specification MIL-V-6753A, 23 March 1956, and has been approved by the Department of Defense and is mandatory for use by the Departments of the Army, the Navy, and the Air Force.

Page 1, title. In the title and wherever else referenced delete "330 to 1200" and substitute "300 to 1200 and 350 to 450".

Page 1, paragraph 1.1. Delete and substitute:

"**1.1 Scope.** This specification covers two types of 0-150V moving iron voltmeters.

"**1.2 Classification.** The voltmeters shall be classified as follows:

MS24332–1	300-1200 cycles
MS24332–2	350-450 cycles".

Page 2, paragraph 3.3.4. Delete the last sentence and substitute "The case shall be fabricated from metal or of a suitable thermosetting-plastic molding material."

Page 3, figure 1. In the dial portion of the figure add a reference line at the "0" graduation and delete one of the reference lines between "50" and "90".

Page 4, paragraph 3.5.2.3. Delete and substitute:

"3.5.2.3 *Frequency influence.* The readings shall not vary more than one percent of full scale value at frequencies of 300 to 1200 cycles for MS24332–1 meters and 350 to 450 cycles for MS24332–2 meters, from the readings observed at 400 cycles."

Page 5, paragraph 3.12, line 2. Delete "1.5" and substitute "1.2".

Page 9, paragraph 4.6.3, line 3. Add the following "shall be performed on samples in accordance with the following tabulation".

Custodians:
Army—Signal Corps
Navy—Bureau of Aeronautics
Air Force—AF

Other Interest:
Army—TO
Navy—MC

Preparing Activity
Air Force

FSC 6625

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T07:25Z
Check the source to verify that this is the current version before use.