FEDERAL SPECIFICATION

GAUZE, ABSORBENT

This amendment, which forms a part of Federal Specification CCC-G-101c, dated January 15, 1957, was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

Page 1, paragraph 1.2.2: At end of paragraph, add:

"Note: One hundred yard bolt is not available in class 2, sterile."

Page 1, paragraph 2.1: Under *"Federal Specifications"* add:

PPP-T-0045—Tape, Gummed, Paper, Reinforced.

PPP-T-76 — Tape, Pressure-Sensitive Adhesive, Paper, Water Resistent.

PPP-B-636—Boxes, Fiber.

Page 1, paragraph 2.1: Delete *"Military Specification:*

JAN-P-108—Boxes, Fiberboard, Corrugated and Solid (Military Type)."

Page 1: In lower right-hand corner delete "FSC Class 6510" and substitute "FSC 6510".

Page 3: Following paragraph 4.3.1 add new paragraph 4.4 as follows:

4.4 Supplier responsibilities for inspection. —Such examinations and tests as are set forth in this specification, or as shall otherwise be appropriate or necessary to insure that the supplies conform to specification requirements, shall be performed by and at the expense of the supplier. Suppliers who do not have facilities adequate for such tests shall arrange for the use of test facilities acceptable to the Government. Records of examinations and tests performed by the supplier shall be maintained by the supplier and made available to the Government, upon the Government's request, at any time, or from time to time, for a period of 2 years after delivery of the supplies to which such records relate.

Pages 3 and 4, paragraph 5.1.3.2: Delete and substitute:

5.1.3.2 *Intermediate package.*—Intermediate package shall be a carton of appropriate size constructed in accordance with Federal Specification PPP-B-636, class 2, compliance symbol W5c or W5s, style RSC, except that the manufacturer's joint shall be effected by the application of a water-resistant tape, or by gluing the integral lap to outer face of end or side panel using water-resistant adhesive in lieu of metal stitches. Closure shall be effected by means of water-resistant, gummed, or pressure-sensitive tape. The tape shall be not less than 2 inches wide, and shall be applied to both top and bottom so that the center seams are covered their full length. Tape shall extend at least 2 inches onto each of the end panels.

Page 4, paragraph 5.2.1.1: Delete and substitute:

5.2.1.1 *Exterior container.*—Exterior container shall be a box of appropriate size constructed in accordance with Federal Specification PPP-B-636, class 2, compliance symbol V3c or V3s, style RSC. Closure shall be as specified in 30.2.2.1 thereof and, in addition,

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:28Z
Check the source to verify that this is the current version before use.

the top and bottom center seams shall be closed by means of tape conforming to the requirements of Federal Specification PPP–T–0045 or PPP–T–76. The tape shall be not less than 3 inches wide and shall extend at least 2 inches onto each of the end panels.

Page 4, paragraph 5.3.2: Delete and substitute:

5.3.2 *Intermediate package.*—Each intermediate package shall be marked in accordance with section 5 of Military Standard MIL–STD–129.

Page 4: Under "5.3.2" add:

5.3.3 *Exterior container.* — Exterior container shall be marked in accordance with Military Standard MIL–STD–129, except that 6.2(c), 6.2(d), 6.5, and 7.5.1 shall not be applicable. Service color and service symbol markings are required only on direct shipments from the contractor's plant to overseas activities."

MILITARY INTERESTS:
Army—M
Navy—Md
Air Force.

Source: https://assist.dla.mil -- Downloaded: 202640-26T12:28Z
Check the source to verify that this is the current version before use.