MILITARY SPECIFICATION

CANNULA, NASAL, OXYGEN, PLASTIC, DISPOSABLE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a flexible, plastic nasal cannula to be used for attachment to an oxygen supply at one end and delivery of oxygen to the nostrils of the patient at the other end.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

Military

MIL-L-10547	Liners, Case, and Sheet Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)
No. 24242 CANNULA, NASAL, OXYGEN, PLASTIC, DISPOSABLE

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.

(Copies of specifications, standards, and drawing, required by suppliers in connection with specific procurement functions, should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.

3.1.1 Cannulae. The cannula (facial portion) shall be fabricated from polyvinyl chloride plastisol or other suitable plastic. It shall readily be adaptable to varied facial contours. The cannulae shall be suitable for insertion into the nostrils.

3.1.2 Tubing. The tubing shall be made of polyvinyl chloride or other suitable clear plastic and tinted green. The green color shall approximate the color of, and shall be at least as intense as, the standard specified in 4.4.

3.1.3 Connector. The female connector attached to the distal end of the tubing shall be green, made of polyvinyl chloride plastic or other suitable material.

3.2 Construction. The nasal oxygen cannula shall be in accordance with any of the four types of Defense Personnel Support Center Drawing No. 24242. An adjustable headband shall be attached for convenience of intranasal administration of oxygen. The surfaces of the nasal oxygen cannula, especially those making contact with the patient, shall be smooth and free from burrs and feathers. The proximal end of the tubing shall be permanently attached to the cannula. The distal end of the tubing shall be permanently attached to the female connector. The connector shall have a tapered design for attachment of the connector to a nipple from a standard oxygen supply source.

3.3 Workmanship. The finished nasal oxygen cannula shall conform to the highest standards of commercial fabrication. The cannula shall be free from defects which detract from its appearance or may impair its serviceability. The nasal oxygen cannula shall not cause irritation and shall be free from objectionable odor at the time of delivery, during use and in storage.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as indicated in Table I.

TABLE I. Sampling Criteria for Examination

Type of defect	Inspection level	AQL* (percent defective)
Major	S-3	1.0
Minor	S-1	2.5

* If acceptance number is not exceeded, lot shall be acceptable. If rejection number is equaled or exceeded, then inspection on new samples at level II may be performed to determine the final rejection or acceptance.

4.3 Examination.

4.3.1 Examination of cannulae. The nasal cannulae shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted as specified in Tables I and II. Examination, however, shall not be restricted to defects listed in Table II. Any nasal cannulae in the sample containing one or more defects shall be rejected and, if the number of defective nasal cannulae in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:16Z
Check the source to verify that this is the current version before use.

TABLE II. Classification of Defects

Categories	Defects
Major	
101	Head band does not hold adjustment.
102	Dimensions not in agreement with drawing.
103	Surfaces of nasal cannula in contact with patient not free from roughness, burrs, slivers, etc.
104	Lumen of nasal cannula not free of foreign matter.
105	Lumen of nasal cannula not free from obstruction.
Minor	
201	Surfaces not free of discoloration.
202	Feathers or burrs present on parts of nasal cannula, other than parts in contact with patient.
203	Unit package does not maintain cannula in clean condition.

4.3.2 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table III. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

TABLE III. Examination for Preparation for Delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and packing materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.

4.4 Green color standard. A solution shall be made up of 0.20 gram nickelous chloride ($NiCl_2.6H_2O$) in 22 milliliters of distilled water. This solution shall be placed in a borosilicate glass test tube of 16 x 150 millimeter size and shall constitute the color standard. It shall be viewed diametrically through the test tube held against a sheet of white paper.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit package. Each cannula shall be cleaned and formed into a coil of such diameter that it will not kink or be permanently deformed. Cannula shall be further packaged in a transparent cellophane or plastic bag of appropriate size. As an alternate, the coiled cannula may be assembled on an interior fitting that will maintain the cannula ready for use in its coiled position, prior to insertion in the bag.

5.1.1.2 Intermediate package. Fifty cannulae shall be packaged in a box of appropriate size constructed in accordance with PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent accidental opening.

5.1.2 Level C. Cannulae shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Two hundred (200) cannulas shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636. class water resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof as specified in the appendix of PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.

5.2.2 Level B. Two hundred (200) cannulae shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1 or PPP-B-636, type CF, class domestic. Closure shall be as specified in the appendix for the applicable box specification.

5.2.3 Level C. Cannulae shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Unit package. Each unit package shall be marked in accordance with MIL-STD-129. Marking shall include directions for use in addition to the legend: "Clean Ready For Use". Marking shall be printed on the interior fittings or directly on the bag in such a manner as to be readily identified. As an alternate, the unit package may be marked by inserting a label with the required markings, inside the bag in such a manner as to permit ready identification of the item.

5.3.2 Intermediate package and exterior container. Each intermediate package and exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. These cannulae are to be used for the intranasal administration of oxygen in low to medium concentrations.

6.2 Ordering data. Procurement documents shall specify the following:

(a) Number, title and date of this specification.

(b) Applicable National Stock Number (NSN)

(c) Selection of applicable level of packaging and packing required (see 5.1 and 5.2).

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:16Z
Check the source to verify that this is the current version before use.

6.3 National Stock Number coverage. This specification covers the following item listed in the Federal Supply Catalog:

National Stock Number	Item Identification
6515-00-246-3782	CANNULA, NASAL, OXYGEN, Plastic, Disposable

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity
Defense Supply Agency - DM

Project No. 6515-1956

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or *manufacturers/ vendors who are prospective suppliers of the product* are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-C-37243 - Cannula, Nasal, Oxygen, Plastic, Disposable

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:16Z
Check the source to verify that this is the current version before use.

FOLD

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.