MILITARY SPECIFICATION

STAND, SURGICAL INSTRUMENT

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers a folding surgical instrument stand packed in a polyester-foam lined, canvas carrying case, suitable for use in field medical units.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-636	-	Boxes, Shipping, Fiberboard.

MILITARY

MIL-M-7298	-	Manual, Technical, Commercial Equipment.

STANDARDS

FEDERAL

Fed. Std. No. 66	-	Steel, Chemical Composition and Hardenability
Fed. Test Method Std. No 151	-	Metals, Test Methods

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:13Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-STD-105	-	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	-	Marking for Shipment and Storage.

DRAWINGS

Defense Personnel Support Center (DPSC)

4061-4350	-	STAND, Surgical Instrument.
3061-4348	-	CASE, Surgical Instrument Stand.

(Copies of Military specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN WELDING SOCIETY (AWS)

Terms and Definitions A3.0

(Application for copies should be addressed to the American Welding Society, 2501 Northwest 7th St , Miami, FL 33125)

NATIONAL FIRE PROTECTION ASSOCIATION (NFPA)

NFPA 56A	Standard for the Use of Inhalation Anesthetics.

(Application for copies should be addressed to the National Fire Protection Association, 470 Atlantic Avenue, Boston MA 02210.)

3 REQUIREMENTS

3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4.1.2.1 and 6.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:13Z
Check the source to verify that this is the current version before use.

3.2 Materials. All materials cited on DPSC Drawings 4061-4350 or 3061-4348 shall conform to the specifications listed thereon, or to the specific characteristics set forth on the drawings. All corrosion-resisting steel shall comply with AISI 302 or 304 of Fed Std No 66 when tested as specified in 4.4.1.

3.3 Design and construction. The surgical instrument stand shall consist of a removable corrosion-resisting steel tray, resting on a frame supported by a telescoping upright The stand base shall be equipped with two electrically-conductive, full swivel casters, and two floor-contacting conductive skids. The height of the tray top rim from the floor shall be adjustable, with provision for automatic locking by foot lever action. Overall dimensions of the stand, when set up, shall not exceed 21-1/2 inches wide by 19 inches long. The stand shall be supplied folded and packed in a polyester-foam lined canvas carrying case. The weight of the stand alone shall not exceed 30 pounds and the overall weight of the stand and case shall not exceed 43 pounds. Style, design, material, and construction shall be in accordance with DPSC drawings 4061-4350 and 3061-4348, and bills of material thereto. The case shall be resistant to mildew.

3 4 Electrical conduction. The stand shall be electrically conductive throughout in accordance with NFPA 56A when tested as specified in 4.4.2

3.5 Performance under load and stability. When set up for operation, the stand shall withstand, without permanent deflection or damage, the application of a static weight of 50 pounds to the instrument tray for 5 minutes. Each stand shall be capable of satisfactory operation with the weight attached when tested for static weight and performance under load in accordance with 4.4.3. The stand shall exhibit no evidence of instability when tested in accordance with 4.4.4.

3.6 Identification marking. Each stand and case shall be permanently marked with the manufacturer's name or registered trademark, or the contract or purchase order number.

3.7 Welding. Welding shall be of the heliarc type and shall be durable, sound in every detail and in accordance with latest practices recommended by the American Welding Society (AWS). All welds shall be produced by certified welding operators (AWS Terms and Definitions - A3.0). Filler metal shall be the same composition as the base metal. All weld joints shall blend into the adjacent metal in gradual, smooth curves. Beads shall be smooth and free of slag, undercut, or spatter. Sufficient metal shall be added to provide a suitable fillet or reinforcement unless otherwise specified. Welds shall not be oxidized. All welds shall be continuous with no visible voids, pin holes, or cracks. Exposed spot welds shall be polished smooth; other welds shall be polished wherever specified on the drawing.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:13Z
Check the source to verify that this is the current version before use.

3.8 Finish. Corrosion-resisting steel shall have a Commercial Standard Number 4 finish.

3 9 Service data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance and lists of component parts. The manual shall comply with the requirements of MIL-M-7298.

3 10 Workmanship Workmanship shall be first class throughout. Stands and cases shall be free from defects which detract from their appearance or may impair their serviceability. Exposed surfaces shall be free of burs or sharp edges. Stand shall rest evenly on a flat surface. Tray shall rest evenly and securely on support. Tray support shall lock securely into position

4 QUALITY ASSURANCE PROVISIONS

4 1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4 1 1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4 1 2 Inspection Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

4 1 2.1 Classification of inspection. The inspection requirements specified herein are classified as follows:

a First article inspection.
b. Quality conformance inspection.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:13Z
Check the source to verify that this is the current version before use.

4.1.2.1.1 First article inspection. First article inspection shall be performed on the stand when a first article sample is required (see 3.1). This inspection shall include examination of 4.3. and tests of 4.4.2, 4.4.3 and 4.4.4.

4.1.2.1.2 Quality conformance inspection. Quality conformance inspection shall include the examination of 4.3 and tests of 4.4.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the metals specified, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component or material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Minor defects	II	4.0
For dimensional examination	S-3	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with table II and MIL-STD-105. The acceptance level shall be zero.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:13Z
Check the source to verify that this is the current version before use.

TABLE II. Sampling for tests.

Characteristic	Requirement	Test Method	Inspection level
Chemical composition	3.2	4.4.1	4.4.1
Electrical conduction	3.4	4.4.2	100%
Performance under load	3.5	4.4.3	II
Stability	3.5	4.4.4	S-1

4.3 Examination. The sample shall be examined to determine compliance with all the requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table III. Examination shall not be restricted to the classified possible defects listed below.

TABLE III. Classification of defects.

Categories	Defects
Major	
101	Any component part missing, broken, or otherwise unusable.
102	Exposed edges not free of burs or sharpness.
103	Stand does not rest evenly on a flat surface.
104	Service data missing, incorrect, or not legible.
105	Tray does not rest evenly or securely on support.
106	Height of tray not easily adjustable.
107	Tray support does not lock securely in position.
108	Carrying case not free of holes, tears, or other damage.
109	Casters do not function smoothly.
110	Foam not properly secured in carrying case.
111	Welding not as specified.
Minor	
201	Identification marking not complete, permanent, or legible.
202	Finish not free of cracks, scratches, or roughness.

4.3.2 Dimensional examination. The unit shall be examined for defects in dimensions. Any dimension not within the tolerance specified herein shall be classified as a defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:13Z
Check the source to verify that this is the current version before use.

4.3.3 Packaging inspection. The inspection of the packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with this specification.

4.4.1 Chemical composition. Four ounces of each lot of metal used in the stands shall be tested for chemical composition in accordance with method 111 or 112 of Fed. Test Method Std. No 151.

4.4.2 Electrical conduction. The complete stand shall be tested in accordance with NFPA 56A. Test points shall be the tray, located on the stand, and each of the skids Conductivity shall be measured from each of these points to a metal plate located under each of the casters (each caster measured separately).

4.4.3 Performance under load. A weight of 50 pounds shall be centered on the instrument tray of the stand. The stand shall then be raised to its uppermost position After 5 minutes, the weight shall be removed and the stand examined for any permanent deformation If there is no visible deformation, the weight shall again be placed on the tray and the stand shall be operated two times throughout the complete range of adjustment of the stand. The stand shall be examined again for any permanent deformation.

4.4.4 Stability The telescoping extension shall be extended to its full height and the casters shall be blocked to prevent rolling. A gradually increasing force (up to 1-1/2 pounds) shall be applied to the vertical upright 40 inches above floor level and in the direction opposite to the tray frame and legs, and the stand shall be checked for stability. The fully extended stand shall then be placed on a 13^{0} inclined plane with the caster wheels parallel to the leg centerline, pointing fully away from the legs, and blocked to prevent rolling. The legs shall point uphill. The stand shall then be checked for stability.

5. PACKAGING

5.1 Packing. Packing shall be level A or C as specified (see 6.2).

5.1.1 Level A. Each stand shall be positioned within the carrying case and the case shall be closed. The stand shall be further packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, class weather-resistant. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:13Z
Check the source to verify that this is the current version before use.

5.1.2 Level C. Each stand shall be positioned within the carrying case and the case shall be closed. The stand shall be further packed in a substantial commercial container of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.2 Marking. Marking shall be as specified in MIL-STD-129.

6. NOTES

6.1 Intended use. Stand is for use in surgery to hold surgical supplies and equipment, and is suitable for use in field medical units.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents.

a. Title, number and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
c. When a first article is required (see 3.1, 4.1.2.1 and 6.3).

6.3 First article. When a first article is required, it shall be tested and approved under the appropriate provisions of 7-104.55 of Armed Services Procurement Regulation. The contract or purchase order specifies the quantity, procedure for delivery and other specific instructions for inspection and approval of the first article.

6.4 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6530-00-027-5252	STAND, SURGICAL INSTRUMENT, Adjustable, Folding, Corrosion-Resisting Metal

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6530-1347

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:13Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D.C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Logistics Services Center
ATTN: DPSC-LS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:13Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21110:132
Check the source to verify that this is the current version before use.