MILITARY SPECIFICATION

TABLE, SURGICAL INSTRUMENT, AND SURGICAL DRESSING

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers a collapsible, adjustable, elevating, surgical instrument table with carrying cases; fabricated of corrosion-resisting steel.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

V-T-285	Thread, Polyester.
CCC-C-419	Cloth, Duck, Cotton, Unbleached Plied-Yarns, Army and Numbered.
CCC-D-950	Dyeing and Aftertreating Processes for Cotton Cloths.
MMM-A-130	Adhesive, Contact.
PPP-B-601	Box, Wood, Cleated-Plywood.
PPP-B-621	Box, Wood, Nailed and Lock-Corner.
PPP-B-640	Boxes, Fiberboard, Corrugated, Triplewall.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof, or Waterproof, Flexible.
MIL-W-530	Webbing, Textile, Cotton, General Purpose, Natural or in colors.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.

STANDARDS

Federal

FED-STD-66	Steel Chemical Composition and Hardenability.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage

DRAWINGS

Defense Personnel Support Center (DPSC)

24181	Table, Surgical Instrument and Dressing, Collapsible, Back with Containers

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

NATIONAL FIRE PROTECTION ASSOCIATION (NFPA)

No. 56A	Standard for the use of Inhalation Anesthetics

(Application for copies should be addressed to the National Fire Protection Association, 60 Batterymarch Street, Boston, MA 02510.)

3. REQUIREMENTS

3.1 Material. Materials shall be as hereinafter specified. Materials which are not specifically described herein shall be of the highest quality throughout and suitable in all respects for the purpose intended.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.

3.1.1 Steel. Steel shall be corrosion-resisting, conforming to class 302 of FED-STD-66.

3.1.2 Thread. Thread shall conform to type I, class 1, subclass B of V-T-285.

3.1.3 Contact Adhesive. Contact adhesive shall conform to MMM-A-130.

3.1.4 Cloth. Cloth used in construction of the cases shall be No. 8 hard texture before finishing and shall conform to type I of CCC-C-419. Cloth shall be dyed Olive Drab No. 7. Finished cloth shall conform to type I, Class B of CCC-D-950. Mildew resistance treatment shall be limited to 2.2 methylene-bis-(4 chlorophenol) or copper 8-quinolinolate.

3.1.5 Webbing. Webbing used in the fabrication of the carrying handles and for reinforcement shall be dyed Olive Drab No. 7 and shall conform to type IIb, class 4 or 7, one inch width of MIL-W-530.

3.2 Dimensions. Thickness of steel sheeting and the wall thickness and diameter of tubing referenced in this specification are nominal standard gage sizes. Commercial tolerances shall apply.

3.3 Construction. The table shall be constructed so that the unit may be set up for use or disassembled for packing and shipment without the use of special tools. When assembled, the unit shall be stable and all moving parts shall operate smoothly. When disassembled the unit shall be stored in two carrying cases (see 3.3.5). The table, when extended to its maximum height, shall support an evenly distributed load of 100 pounds without failure or damage to any component (see 4.4.3). The overall weight of the table without case shall not exceed 90 pounds. The table top shall have a minimum vertical adjustment range of 39 to 60 inches from the floor. The table, when extended to its full height and with the casters locked, shall not tip when a force of seven pounds is applied to a leg at a height of 40 inches from the floor (see 4.4.2). The table shall be in accordance with this specification and DPSC Drawing 24181. The total weight of the table and carrying cases shall not exceed 155 pounds.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.

3.3.1 Top. Top shall be constructed, as shown on drawing 24181, of No. 18 United States Standard (U.S.S.) gage (0.049 inch thick) class 302 corrosion-resisting steel. The dimensions of the top shall be 23 inches by 40 inches. The back edge shall be formed up as shown on the drawing. The corners shall be fully welded. A No. 14 U.S.S. gage (0.078 inch thick) corrosion-resisting steel, hat-shaped, reinforcing channel, constructed as shown on Drawing 24181, shall be spotwelded to the top on nominal 6 inch centers. The channel shall also be welded to the 1/2-inch flange at each end of the table.

3.3.2 Table legs. The supporting legs of the table shall consist of two telescoping, square, tubular members. The top section shall be 2 inches square and the bottom section 1-3/4 inches square, both shall be No. 16 U.S.S. gage (0.062 inch thick) corrosion-resisting steel. Welded to each upper leg section shall be two triangular support gussets of No. 11 U.S.S. gage (0.125 inch thick) corrosion-resisting steel. The gussets shall be welded to the upper leg. Holes to mate with the dowel pins of the hat channel shall be drilled in the gussets. The gussets shall be secured to the hat channel with 500 pound, fast acting, toggle clamps. The table leg bases shall each consist of a 20 inch length of 2 inch by 2 inch, No. 11 U.S.S. gage (0.125 inch thick) corrosion-resisting steel channel. The ends of the channel shall be closed with No. 16 U.S.S. gage corrosion-resisting steel caps. The center of the channel shall accomodate the lower tubular leg section which protrudes through the top of the base. The upright shall be welded to the bottom edge of the base. Each end of the base shall be equipped with one 3 inch, electrically conductive, ball bearing, swivel caster. The table shall have a total of four casters. The two front casters shall have foot operated wheel locks. The top of the upper leg section shall be capped to conceal the elevating mechanism.

3.3.3 Elevating mechanism. Elevating mechanism shall be in accordance with Drawing 24181

3.3.4 Elevating handle. Two elevating handles shall be supplied with the table. The handles shall be constructed of corrosion-resisting steel with a plastic grip, and shall be quickly and easily removable from the table. The grip section of the handle shall be spring loaded so that it will fold out of the way when not in use

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.

3.3.5 Carrying cases. Two carrying cases shall be supplied with the table in which the disassembled components shall be packed for storage and shipment. Foam sections shall have the surfaces to be bonded lightly sanded with 180 grit paper or cloth, and then thoroughly cleaned. The pieces shall be bonded together using contact adhesive. The adhesive manufacturer's directions for use shall be strictly followed.

The thread size shall be size FF for all stitching; except that the handles, reinforcement and zipper shall be stitched with size 3 thread. Ends shall be backstitched at least one inch. All thread ends shall be trimmed. Thread breaks shall be overstitched not less than one inch at each break. The zippers shall be size 7, type MBT (closed, both ends with metal bridge-stop), nickel plated finish, slider No. 2818, non-lock, olive drab 3/4 inch wide tape. The carrying cases shall be in conformance with Drawing 24181. When Government furnished Material is supplied (duck and webbing), the contractor shall visually examine such material for defects in accordance with CCC-D-950 and MIL-W-530, as applicable, before fabricating such material into carrying cases.

3.4 Electrical conductivity. The table shall be electrically conductive throughout in accordance with NFPA bulletin 56A (see 4.4.1).

3.5 Case marking. Carrying cases shall be marked in accordance with Drawing 24181.

3.6 Finish. All exposed metal surfaces shall be finished in a No. 4 commercial finish. Exposed welds shall be ground smooth, and finished to match adjoining surfaces.

3.7 Workmanship. Workmanship shall be first class throughout. The finished table shall be free from defects which detract from its appearance, or may impair its serviceability. There shall be no sharp edges on areas exposed during use.

3.8 Welding. Welding shall be durable, sound in every detail and in accordance with latest practices recommended by the American Welding Society. All welds shall be produced by qualified welding operators with satisfactory welding equipment, using satisfactory filler metal and electrode, where applicable. Surfaces to be welded shall be free from grease, oil, oxides or any other foriegn matter. Any cleaning method which will remove these substances and leave the surface clean is satisfactory. All exposed weld joints shall blend into the adjacent metal in gradual, smooth curves. Beads shall be smooth and free of slag, excessive undercut, or excessive spatter. Sufficient metal shall shall be added to provide a suitable fillet or reinforcement, unless

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.

otherwise specified. Fusion through the full thickness of the base metal, in areas other than those where 100% penetration is a requirement of joint design, are not acceptable when welding materials over 0.050 inch in thickness. Welds shall not be oxidized. All welds shall be continuous with no visible voids, pin holes, or cracks. Unexposed welds shall be cleaned and rough surfaces removed.

3.9 Service data. The contractor shall enclose, with each unit, two copies each of one or more service data booklets. In addition, the contractor shall mail 15 copies of the approved service data to the contracting officer for distribution. The service data shall contain complete instructions for operation and maintenance, and a list of component parts. Each part of the unit other than common fasteners shall be identified in the service data and parts list by the manufacturer's part number. As soon as practicable after award of the contract or purchase order, the contractor shall furnish two preliminary copies of his proposed service data for review and comment by Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) to allow incorporation by the contractor of required revisions and corrections in the final service data, without delaying delivery under the contract or purchase order. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of contract, and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the raw materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component or material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in Table I.

Table I. Sampling for Examination

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	1.0
Minor defects	II	4.0
For dimensional examination	S-3	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in Table II below. The acceptance number shall be zero for all tests specified herein.

Table II. Sampling Criteria for Tests

Component	Characteristic	Reqmt	Test Procedure	Inspection level
Table	Electrical conductivity	3.4	4.4.1	100 percent
	Stability	3.3	4.4.2	S-2
	Load test	3.3	4.4.3	S-2

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.

4.2.3 Raw material. Not less than 4 ozs. of each lot of material to be used in construction of this table shall be properly labeled for identification and if requested forwarded to a Government test facility to insure compliance with specifications. In lieu of test certificates by the manufacturer, certified test certificates submitted by the steel manufacturer shall be acceptable.

4.3 Examination. The tables shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with Table III.

Table III Classification of Defects

Categories and defects*	
Major	
101	Any component part missing, broken, or otherwise unusable.
102	Exposed edges not free of burs or sharpness.
103	Table fails to move and turn easily
104	Elevating mechanism does not operate easily.
105	Table does not sit level with all four casters contacting the floor, when assembled.
106	Welding not in accordance with specification.
107	Zippers on carrying cases do not operate.
108	Table and legs do not pack into carrying cases.
Minor	
201	Identification marking not complete, permanent or legible.
202	Identification marking not securely attached

* Examination shall not be restricted to the classified possible defects listed above.

4.3.2 Dimensional examination. The table shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Electrical conductivity. The assembled stand, complete with rubber casters, shall be tested in accordance with NFPA No. 56A.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.

4.4.2 Stability. The table shall be extended to its full height and the wheels shall be locked to prevent rolling. A force shall be applied to the vertical upright 40 inches above the floor level and in the direction opposite to the pedestal legs. The stand shall not tip until a force greater than 7 pounds has been applied.

4.4.3 Load test. The table shall be capable of supporting an evenly distributed 100 pound static load for 10 minutes without failure or damage to any component. This load test shall be performed with the table at full height, at a height of 50 inches, and at its lowest height. The elevating mechanism shall operate smoothly with the load in place.

4.5 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

Table IV Examination for Preparation for Delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and Packing Materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.

5.1.1 Level A.

5.1.1.1 Unit package. Each table shall be disassembled and the disassembled components positioned in their respective compartments within the carrying cases and the cases secured.

5.1.2 Level C. Packaging shall be accomplished as specified in 5.1.1.1.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Each table packaged as specified in 5.1.1.1 shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-601, overseas type; PPP-B-621, class 2 or PPP-B-640, class 2, style E. Closure and strapping shall be as specified in the appendix of the applicable box specification. Metal strapping is not permitted for PPP-B-640 box.

5.2.1 1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof as specified in the appendix of PPP-B-640.

5.2.2 Level B. Each table packaged as specified in 5.1.1.1 shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-601, domestic type; PPP-B-621, class 1 or PPP-B-640, class 1, style E. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Exterior container. Each exterior container shall be marked in accordance with MIL-STD-129. Unit package marking not required

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Intended use. This surgical instrument table is intended for use in the surgical suite of the Medical Unit Self-contained Transportable (MUST) hospital.

6.2 Ordering data Procurement documents should specify:

(a) Title, number and date of this specification.

(b) the applicable revision date of the referenced drawing

(c) the applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification covers the following item of supply appearing in the Federal Supply Catalog:

NSN	Item Identification
6530-00-442-3181	TABLE, SURGICAL INSTRUMENT AND DRESSING, Corrosion-Resisting Steel, 40 x 20 x 38 inches.

6.4 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Custodians
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6530-1355
1S

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No 22-R255

INSTRUCTIONS The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government Fold on lines on reverse side, staple in corner, and send to preparing activity Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements Attach any pertinent data which may be of use in improving this document If there are additional papers, attach to form and place both in an envelope addressed to preparing activity

DOCUMENT IDENTIFIER AND TITLE MIL-T-37240 TABLE, SURGICAL INSTRUMENT, AND SURGICAL DRESSING

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1 HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A GIVE PARAGRAPH NUMBER AND WORDING

B RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2 COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3 IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4 REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS. DEFENSE PERSONNEL SUPPORT CENTER
ATTN DIRECTORATE OF MEDICAL MATERIEL. CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:16Z
Check the source to verify that this is the current version before use.