A-A-51211
5 December 1985

# COMMERCIAL ITEM DESCRIPTION

## STAND, LABORATORY APPARATUS SUPPORT

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification NNN-S-750.

This commercial item description covers one type of lab apparatus support stand.

Salient characteristics:

Shall consist of a rectangular base and support rod suitable for use in laboratory procedures. Shall be of the following dimensions in TABLE I.

TABLE I

| Size | Base | Rod | | |
|---|---|---|---|---|
| | Dimensions (inches) | Length (inches) | Diameter (inch) | Thread (per inch) UNC |
| 1 | 6 x 9 | 24 | 1/2 | 13 |
| 2 | 6 x 11 | 36 | 1/2 | 13 |
| 3 | 5 x 8 | 20 | 3/8 | 16 |

The base shall be fabricated of a suitable metal of sufficient weight to ensure stability in use. The four corners of the base shall be provided with legs cast internally or securely attached to the base. The top of the base shall be flat and level with a drilled hole, 1/2 inch nominal diameter through the midline of the base at a nominal 1 inch from the edge of the short dimension.

The support rod shall be fabricated of a suitable metal having a nominal dimension as listed in table I. One end of the rod shall be threaded with UNC threads and of a sufficient length to secure the rod to the base. The other end shall be rounded. The rod when attached to the base shall be perpendicular to the flat surface on which it rests.

The laboratory apparatus support stand shall not have any of the following which will be considered defects.

FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:52Z
Check the source to verify that this is the current version before use.

Sharp or rough edges
Rust or corrosion
Rocking when on a flat surface
Rod not perpendicular to the horizontal axis
Scratches, dents
Casting flaws

Workmanship. The stand shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One stand of the size, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspections. Inspections, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Unless otherwise specified, sampling for examination shall be conducted at inspection level S-II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-II, AQL 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:52Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Numbers are covered by this document:

| | |
|---|---|
| 6640-00-440-1200 | STAND, LABORATORY APPARATUS SUPPORT, Size 1 |
| 6640-00-440-1250 | STAND, LABORATORY APPARATUS SUPPORT, Size 2 |
| 6640-00-290-6717 | STAND, LABORATORY APPARATUS SUPPORT, Size 3 |

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:
VA-OSS

Project No. 6640-1344
Location S3-8-3



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:52Z
Check the source to verify that this is the current version before use.