A-A-51214
9 December 1985

COMMERCIAL ITEM DESCRIPTION

DISPENSER, DENTAL ACRYLIC RESIN

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers dental acrylic resin liquid and powder dispensers.

Salient characteristics:

The dispenser shall be type I, dental acrylic resin liquid dispenser, or type II, dental acrylic resin powder dispenser as specified in the procurement document for the applicable National Stock Number and Item Identification.

The flexible plastic dispenser shall be suitable for use in the precise application of methyl methacrylate monomer (type I, liquid dispenser) and polymer (type II, powder dispenser) for repair of denture prosthesis. Both the liquid and the powder dispenser shall consist of a bottle, a screw cap, delivery tubing, and an AISI 304 stainless steel flow tube. The liquid dispenser shall also include an AISI 308L stainless steel probing rod or stylus. Liquid or powder shall be expelled through the delivery tube when the bottle is squeezed. The pick up end of the delivery tube shall be diagonally cut to assist in the transfer of the monomer or polymer to the tube. The tapered end of the delivery tube shall have a 17 gauge stainless steel flow tube. The type I dispenser shall have a securely held stylus (0.035 $\pm$ 0.002 inch in diameter) for use as a stirring rod. The delivery tube shall form an angle of $60^{\circ} \pm 10^{\circ}$ with respect to the central axis of the bottle. The interior of the screw cap shall be threaded to properly mate with the threads on the neck of the bottle. All screw caps and bottles shall be interchangeable with each other. The completely assembled dispenser shall rest evenly on a smooth, level surface. The bottle shall have a 2-ounce capacity and a 1/2 $\pm$ 1/16 inch neck opening. The rod shall provide for touch contact of the monomer at the selected site, and shall serve as a stirring rod during the mixing of the monomer and polymer. The monomer shall be discharged at the end of the flow tube, and shall flow to the end of the probing rod, where it shall be dispensed as individual drops. There shall be no leakage in the entire system.

Material. The bottle, screw cap, and delivery tubing shall be fabricated from flexible polyethylene or polypropylene. The polyethylene or polypropylene shall have a natural color with no color additives. The flow tube and the probing rod shall be fabricated from austenitic, corrosion-resisting steel.

Workmanship. The dispenser shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One dispenser of the type specified, constitutes one unit.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:52Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one dispenser. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:52Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

Stock Number	Item Identification
6520-00-142-9039	DISPENSER, DENTAL ACRYLIC RESIN LIQUID, Plastic, 2 fl. oz. (59 ml).
6520-00-142-9044	DISPENSER, DENTAL, ACRYLIC RESIN POWDER, Plastic, 2 oz. (57 gram).

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:
VA-OSS

05390

Project No. 6520-1876

U.S. GOVERNMENT PRINTING OFFICE: 1984 - 605-036/40456

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:52Z
Check the source to verify that this is the current version before use.