A-A-51215
20 November 1985

COMMERCIAL ITEM DESCRIPTION

BRUSH, DENTURE

The General Services Administration has authorized the use of this commercial item description in preference to Interim Federal Specification, H-B-00185a (DSA-DM), type 1.

This commercial item description covers a brush suitable for use in the dental laboratory for scrub cleaning dental prosthesis during fabrication.

Salient characteristics:

The brush shall consist of a one piece methacrylate block and 61 tufts of 0.012 + 0.001 inch diameter black nylon. The distal end of the block shall have 61 evenly spaced holes arranged in three rows. Rows shall be staggered; middle row shall contain 21 holes, and outside rows shall contain 20 holes. One tuft shall be placed in each hole and the tuft shall be fastened to the block by a metal staple. Metal shall be corrosion resisting or shall be made corrosion resisting by plating. Top of tufts shall be neatly and evenly trimmed to form a bristle area with a straight top. The proximal end of block shall serve as a handle. The handle shall have a hole for hanging purposes. The block shall be smoothly finished. Style and design of brush shall be as shown in Figure 1.

Workmanship. The brush shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One brush packaged in a box or sealed plastic bag, as specified, constitutes one unit.

Quality Assurance Provisions.


Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil - Downloaded: 2025-10-26 08:58Z
Check the source to verify that this is the current version before use.



BRUSH, DENTURE

FIG. 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one brush. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6520-00-502-9050 BRUSH, DENTURE

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

A-A-51215

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING AGENCY:
VA-OSS

47835

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6520-1804

Location S2-12-7

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.